



ACCUTEST®

Laboratories

formerly Evergreen Analytical, Inc.

#1944982

012

August 20, 2009

Peter Gintautas
Colorado Oil & Gas Conservation Comm.
213 Corundum Rd
Trinidad, CO 81082

Lab Work Order: 09-5730
Client Project ID: Complaint 2002 12599

Dear Peter Gintautas:

Enclosed are the analytical results for the samples shown in the Laboratory Work Order Summary.

THE INVOICE WILL BE MAILED FROM OUR NEW JERSEY OFFICE UNDER SEPARATE COVER.

The enclosed data for testing performed at Accutest Laboratory (formerly Evergreen Analytical) have been reviewed for quality assurance. A case narrative is included to describe any anomalies associated with the samples or data.

Accutest will dispose of all samples 44 days from the sample receipt date. If you want samples returned, please advise us by mail or fax as soon as possible.

A copy of this project report and supporting data will be retained for a period of five years unless we are otherwise advised by you. A document retrieval charge will apply.

Thank you for using the services of Accutest Laboratories. If you have any questions concerning the analytical data, please contact me. Please direct other questions to Client Services.

Sincerely,

Joseph J Egry IV/ Tiffany Pham
Quality Assurance

WORK ORDER Summary**Evergreen Analytical, Inc.****09-5730****Rpt To:** Peter Gintautas**Email To:** peter.gintautas@state.co.us

8/3/2009 2:30:22 PM

Colorado Oil & Gas Conservation
Comm.213 Corundum Rd
Trinidad, CO 81082
(719) 846-3091**Client Project ID:** Complaint 2002 12599**QC Level:** Level I+**Comments:**

Sample ID	Client Sample ID	Matrix	Collection Date	Date Received	Test Code	Test Name	Hold	MS	Date Due	Hold Time
09-5730-01A	York WW	Water	3/24/09 0000	8/03/09	6010_WT *	6010: Total Metals	<input type="checkbox"/>	<input type="checkbox"/>	8/17/09	9/20/09
09-5730-01B	York WW	Water	3/24/09 0000	8/03/09	6010_D *	6010: Dissolved Metals	<input type="checkbox"/>	<input type="checkbox"/>	8/17/09	9/20/09

Page _____ of _____

Page ____ of ____

(D-4-1)*

- 2 work

☐ 3 - 5 work days, 50% ☐ 6 - 9 work days, 25%

***Subject to surcharge & exceptions noted in fee schedule.**

For laboratory:	
-----------------	--

Use Only

WO# 09-5730

B.O.F.#

CS(0) /

C/S (I) NA / 11D

Temp. $^{\circ}\text{C}$ 14.2 / Ice no

Seals Present Y / ☒ / NA
Complete Date Y / 12 / 2001

300	Saltpies Pres. Y / N / NA	Headspace Y / N / NA
-----	---------------------------	----------------------

By 7/

Witnessed by _____

[illegible][illegible][illegible]

NO SOUNDS

[illegible]

~~A~~ ~~Section~~ A C/AIB

1/20 dissolved

[illegible]

1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24	25	26	27	28	29	30	31	32	33	34	35	36	37	38	39	40	41	42	43	44	45	46	47	48	49	50	51	52	53	54	55	56	57	58	59	60	61	62	63	64	65	66	67	68	69	70	71	72	73	74	75	76	77	78	79	80	81	82	83	84	85	86	87	88	89	90	91	92	93	94	95	96	97	98	99	100
---	---	---	---	---	---	---	---	---	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	-----

[illegible][illegible]

1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24	25	26	27	28	29	30	31	32	33	34	35	36	37	38	39	40	41	42	43	44	45	46	47	48	49	50	51	52	53	54	55	56	57	58	59	60	61	62	63	64	65	66	67	68	69	70	71	72	73	74	75	76	77	78	79	80	81	82	83	84	85	86	87	88	89	90	91	92	93	94	95	96	97	98	99	100
---	---	---	---	---	---	---	---	---	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	----	-----

--	--

2, O-Phosphate, Sulfate (Circle)

Time	Received by: (Signature)	Date/Time
------	--------------------------	-----------

--	--	--	--

[illegible]

--	--

Evergreen Analytical, Inc.

Date: 20-Aug-09

Lab Order: 09-5730

Client Project ID Complaint 2002 12599

CASE NARRATIVE

SAMPLE RECEIVING

Sample(s) were hand delivered to the laboratory by the client.

Custody seals were not present.

The temperature of the sample(s) upon arrival was 14.2°C.

Sample(s) were received in good condition, in the proper container, and within holding times. JD

QUALITY ASSURANCE (QA)

Analyses performed on samples in this work order by EAL meet the requirements of the EAL Quality Assurance Program unless otherwise explained. Analyses of RCRA samples meet the requirements of NELAC and Utah Rule R444-14 unless otherwise explained. CMS

CLIENT SERVICES

There are no anomalies to report. PM

METALS ANALYSIS

Matrix spike and matrix spike duplicate (MS/MSD) recoveries for silicon are outside the QC limits. This is assumed due to sample matrix interference, since the laboratory control spikes (LCS) recoveries are within QC limits. Silicon is not stable and can easily precipitate out of an acid solution. There are no other anomalies to report. JT

All total basis results were significantly higher than dissolved basis results, except for sodium. The sodium results are within the normal analytical variability, showing an RPD of only 6.9 %. The total basis analysis also showed 19 mg/L aluminum while the dissolved analysis was non-detect for this metal. CMS

Evergreen Analytical, Inc.
4036 Youngfield Street, Wheat Ridge, Colorado 80033-3862
(303) 425-6021

004

Client Sample ID: York WW
Client Project ID: Complaint 2002 12599
Date Collected: 3/24/09
Date Received: 8/3/09

Lab Work Order 09-5730
Lab Sample ID: 09-5730-01
Sample Matrix: Water

DISSOLVED METALS

Method: SW6010B

Prep Method: E200.7/SW3010A

Date Prepared: 8/7/09
Date Analyzed: 8/13/09

Lab File ID: 081309PM
Method Blank: MB-20239

Dilution Factor: 1
Lab Fraction ID: 09-5730-01B

Analytes	CAS Number	Result	LQL	Units
Calcium	7440-70-2	1.6	0.39	mg/L
Iron	7439-89-6	0.44	0.070	mg/L
Magnesium	7439-95-4	0.18	0.15	mg/L
Manganese	7439-96-5	0.031	0.0050	mg/L
Potassium	7440-09-7	0.89	0.34	mg/L
Silicon	7440-21-3	4.6	0.050	mg/L
Sodium	7440-23-5	150	0.40	mg/L

TOTAL METALS

Method: SW6010B

Prep Method: E200.7/SW3010A

Date Prepared: 8/10/09
Date Analyzed: 8/14/09

Lab File ID: 081309PM
Method Blank: MB-20271

Dilution Factor: 1
Lab Fraction ID: 09-5730-01A

Analytes	CAS Number	Result	LQL	Units
Calcium	7440-70-2	3.3	0.39	mg/L
Iron	7439-89-6	8.3	0.070	mg/L
Magnesium	7439-95-4	2.2	0.15	mg/L
Potassium	7440-09-7	3.3	0.34	mg/L
Sodium	7440-23-5	140	0.40	mg/L

Date Prepared: 8/10/09
Date Analyzed: 8/14/09

Lab File ID: 081309PM
Method Blank: MB-20271

Dilution Factor: 1
Lab Fraction ID: 09-5730-01A

Analytes	CAS Number	Result	LQL	Units
Manganese	7439-96-5	0.20	0.0050	mg/L



Analyst



Approved

Qualifiers: B - Analyte detected in the associated Method Blank, value not subtracted from result
E - Extrapolated value. Value exceeds calibration range
H - Sample analysis exceeded analytical holding time
J - Indicates an estimated value when the compound is detected, but is below the LQL
S - Spike Recovery outside accepted limits
U - Compound analyzed for but not detected
X - See case narrative
* - Value exceeded the Maximum Contamination Level (MCL), TCLP limit, or if compound is undetected, LQL exceeds MCL.

Definitions: NA - Not Applicable
LQL - Lower Quantitation Limit
Surr - Surrogate

Print Date: 8/20/2009

Evergreen Analytical, Inc.
4036 Youngfield Street, Wheat Ridge, Colorado 80033-3862
(303) 425-6021

005

Client Sample ID: York WW
Client Project ID: Complaint 2002 12599
Date Collected: 3/24/09
Date Received: 8/3/09

Lab Work Order 09-5730
Lab Sample ID: 09-5730-01
Sample Matrix: Water

TOTAL METALS

Method: SW6010B

Prep Method: E200.7/SW3010A

Date Prepared: 8/14/09
Date Analyzed: 8/19/09

Lab File ID: 081709PM
Method Blank: MB-20344

Dilution Factor: 1
Lab Fraction ID: 09-5730-01A

Analytes	CAS Number	Result	LQL	Units
Silicon	7440-21-3	17	0.050	mg/L



Analyst



Approved

Qualifiers: B - Analyte detected in the associated Method Blank, value not subtracted from result
E - Extrapolated value. Value exceeds calibration range
H - Sample analysis exceeded analytical holding time
J - Indicates an estimated value when the compound is detected, but is below the LQL
S - Spike Recovery outside accepted limits
U - Compound analyzed for but not detected
X - See case narrative
* - Value exceeded the Maximum Contamination Level (MCL), TCLP limit, or if compound is undetected, LQL exceeds MCL.

Definitions: NA - Not Applicable
LQL - Lower Quantitation Limit
Surr - Surrogate

Print Date: 8/20/2009

QUALITY ASSURANCE REPORTS

METHOD BLANKS (MB)

LABORATORY CONTROL SPIKES (LCS)

MATRIX SPIKES (MS/MSD)*

DUPLICATES (DUP)*

• **For Metals or Wet Chemistry analyses: only included if requested.**

Work Order: 09-5730

Client Project ID: Complaint 2002 12599

ANALYTICAL QC SUMMARY REPORT

BatchID: 20239

Sample ID: MB-20239	SampType: MBLK	TestCode: 200.7_D	Run ID: ICP-OPTIMA 5300 DV_090813A	Prep Date: 8/7/2009	Units: mg/L						
	Batch ID: 20239	TestNo: E200.7, Rev.	FileID: 081309PM	Analysis Date: 8/13/2009	SeqNo: 891602						
Analyte	Result	LQL	SPK value	SPK Ref Val	%REC	LowLimit	HighLimit	RPD Ref Val	%RPD	RPDLimit	Qual

Calcium	0	0.387									
Iron	0	0.0700									
Magnesium	0	0.150									
Manganese	0	0.00500									
Potassium	0	0.340									
Silicon	0	0.0500									
Sodium	0	0.400									

Sample ID: LCS-20239	SampType: LCS	TestCode: 200.7_D	Run ID: ICP-OPTIMA 5300 DV_090813A	Prep Date: 8/7/2009	Units: mg/L						
	Batch ID: 20239	TestNo: E200.7, Rev.	FileID: 081309PM	Analysis Date: 8/13/2009	SeqNo: 891603						
Analyte	Result	LQL	SPK value	SPK Ref Val	%REC	LowLimit	HighLimit	RPD Ref Val	%RPD	RPDLimit	Qual

Calcium	9.816	0.387	10	0	98.2	85	115	0	0		
Iron	4.886	0.0700	5	0	97.7	85	115	0	0		
Magnesium	9.665	0.150	10	0	96.7	85	115	0	0		
Manganese	1.864	0.00500	2	0	93.2	85	115	0	0		
Potassium	9.713	0.340	10	0	97.1	85	115	0	0		
Silicon	4.936	0.0500	5	0	98.7	85	115	0	0		
Sodium	9.859	0.400	10	0	98.6	85	115	0	0		

Sample ID: 09-5712-03CMS	SampType: MS	TestCode: 200.7_D	Run ID: ICP-OPTIMA 5300 DV_090813A	Prep Date: 8/7/2009	Units: mg/L						
	Batch ID: 20239	TestNo: E200.7, Rev.	FileID: 081309PM	Analysis Date: 8/13/2009	SeqNo: 891609						
Analyte	Result	LQL	SPK value	SPK Ref Val	%REC	LowLimit	HighLimit	RPD Ref Val	%RPD	RPDLimit	Qual

Calcium	127.4	0.484	12.5	113.1	114	75	125	0	0		
Iron	5.942	0.0875	6.25	0	95.1	75	125	0	0		
Magnesium	25.96	0.188	12.5	14.1	94.9	75	125	0	0		
Manganese	2.382	0.00625	2.5	0.1257	90.2	75	125	0	0		
Potassium	18.82	0.425	12.5	6.775	96.3	75	125	0	0		

Qualifiers:

U - Not detected at or above the Reporting Limit
J - Analyte detected below quantitation limits
S - Spike Recovery outside acceptance limits
E - Extrapolated value, value exceeds calibration range.

R - RPD outside acceptance limits
B - Analyte detected in the associated Method Blank
H - Prep or analytical holding time exceeded
X - See case narrative

Work Order: 09-5730

Client Project ID: Complaint 2002 12599

ANALYTICAL QC SUMMARY REPORT

BatchID: 20239

Sample ID: 09-5712-03CMS	SampType: MS	TestCode: 200.7_D	Run ID: ICP-OPTIMA 5300 DV_090813A					Prep Date: 8/7/2009	Units: mg/L		
	Batch ID: 20239	TestNo: E200.7, Rev.	FileID: 081309PM					Analysis Date: 8/13/2009	SeqNo: 891609		
Analyte	Result	LQL	SPK value	SPK Ref Val	%REC	LowLimit	HighLimit	RPD Ref Val	%RPD	RPDLimit	Qual
Silicon	21.26	0.0625	6.25	14.9	102	75	125	0	0		
Sodium	68.45	0.500	12.5	56	99.6	75	125	0	0		

Sample ID: 09-5712-03CMSD	SampType: MSD	TestCode: 200.7_D	Run ID: ICP-OPTIMA 5300 DV_090813A					Prep Date: 8/7/2009	Units: mg/L		
	Batch ID: 20239	TestNo: E200.7, Rev.	FileID: 081309PM					Analysis Date: 8/13/2009	SeqNo: 891610		
Analyte	Result	LQL	SPK value	SPK Ref Val	%REC	LowLimit	HighLimit	RPD Ref Val	%RPD	RPDLimit	Qual
Calcium	127.2	0.484	12.5	113.1	113	75	125	127.4	0.127	20	
Iron	5.948	0.0875	6.25	0	95.2	75	125	5.942	0.0924	20	
Magnesium	25.94	0.188	12.5	14.1	94.8	75	125	25.96	0.0520	20	
Manganese	2.381	0.00625	2.5	0.1257	90.2	75	125	2.382	0.0471	20	
Potassium	18.81	0.425	12.5	6.775	96.3	75	125	18.82	0.0425	20	
Silicon	21.23	0.0625	6.25	14.9	101	75	125	21.26	0.159	20	
Sodium	68.29	0.500	12.5	56	98.3	75	125	68.45	0.235	20	

Qualifiers:

U - Not detected at or above the Reporting Limit
 J - Analyte detected below quantitation limits
 S - Spike Recovery outside acceptance limits
 E - Extrapolated value, value exceeds calibration range.

R - RPD outside acceptance limits
 B - Analyte detected in the associated Method Blank
 H - Prep or analytical holding time exceeded
 X - See case narrative

Work Order: 09-5730

Client Project ID: Complaint 2002 12599

ANALYTICAL QC SUMMARY REPORT

BatchID: 20271

Sample ID: MB-20271	SampType: MBLK	TestCode: 200.7_T	Run ID: ICP-OPTIMA 5300 DV_090813A	Prep Date: 8/10/2009	Units: mg/L						
	Batch ID: 20271	TestNo: E200.7, Rev.	FileID: 081309PM	Analysis Date: 8/14/2009	SeqNo: 891639						
Analyte	Result	LQL	SPK value	SPK Ref Val	%REC	LowLimit	HighLimit	RPD Ref Val	%RPD	RPDLimit	Qual

Calcium	0	0.387									
Iron	0	0.0700									
Magnesium	0	0.150									
Potassium	U	0.340									
Sodium	0	0.400									

Sample ID: MB-20271	SampType: MBLK	TestCode: 200.7_T	Run ID: ICP-OPTIMA 5300 DV_090813A	Prep Date: 8/10/2009	Units: mg/L						
	Batch ID: 20271	TestNo: E200.7, Rev.	FileID: 081309PM	Analysis Date: 8/14/2009	SeqNo: 892292						
Analyte	Result	LQL	SPK value	SPK Ref Val	%REC	LowLimit	HighLimit	RPD Ref Val	%RPD	RPDLimit	Qual

Manganese	U	0.00500									
-----------	---	---------	--	--	--	--	--	--	--	--	--

Sample ID: LCS-20271	SampType: LCS	TestCode: 200.7_T	Run ID: ICP-OPTIMA 5300 DV_090813A	Prep Date: 8/10/2009	Units: mg/L						
	Batch ID: 20271	TestNo: E200.7, Rev.	FileID: 081309PM	Analysis Date: 8/14/2009	SeqNo: 891640						
Analyte	Result	LQL	SPK value	SPK Ref Val	%REC	LowLimit	HighLimit	RPD Ref Val	%RPD	RPDLimit	Qual

Calcium	8.867	0.387	10	0	88.7	85	115	0	0		
Iron	4.437	0.0700	5	0	88.7	85	115	0	0		
Magnesium	8.706	0.150	10	0	87.1	85	115	0	0		
Potassium	8.872	0.340	10	0.03265	88.7	85	115	0	0		
Sodium	9.065	0.400	10	0	90.6	85	115	0	0		

Sample ID: LCS-20271	SampType: LCS	TestCode: 200.7_T	Run ID: ICP-OPTIMA 5300 DV_090813A	Prep Date: 8/10/2009	Units: mg/L						
	Batch ID: 20271	TestNo: E200.7, Rev.	FileID: 081309PM	Analysis Date: 8/14/2009	SeqNo: 892291						
Analyte	Result	LQL	SPK value	SPK Ref Val	%REC	LowLimit	HighLimit	RPD Ref Val	%RPD	RPDLimit	Qual

Manganese	1.766	0.00500	2	0.001056	88.3	85	115	0	0		
-----------	-------	---------	---	----------	------	----	-----	---	---	--	--

Qualifiers:

U - Not detected at or above the Reporting Limit
J - Analyte detected below quantitation limits
S - Spike Recovery outside acceptance limits
E - Extrapolated value, value exceeds calibration range.

R - RPD outside acceptance limits
B - Analyte detected in the associated Method Blank
H - Prep or analytical holding time exceeded
X - See case narrative

Work Order: 09-5730

Client Project ID: Complaint 2002 12599

ANALYTICAL QC SUMMARY REPORT

BatchID: 20271

Sample ID: 09-5678-01CMS		SampType: MS		TestCode: 6010_D		Run ID: ICP-OPTIMA 5300 DV_090813A		Prep Date: 8/10/2009		Units: mg/L	
		Batch ID: 20271		TestNo: SW6010B		FileID: 081309PM		Analysis Date: 8/14/2009		SeqNo: 891642	
Analyte	Result	LQL	SPK value	SPK Ref Val	%REC	LowLimit	HighLimit	RPD Ref Val	%RPD	RPDLimit	Qual
Calcium	86.89	0.43	11.11	75.51	102	75	125	0	0		
Iron	5.011	0.078	5.556	0.04291	90.2	75	125	0	0		
Magnesium	19.19	0.17	11.11	9.287	89.1	75	125	0	0		
Manganese	1.899	0.0056	2.222	0.01756	84.7	75	125	0	0		
Potassium	16.01	0.38	11.11	5.693	92.9	75	125	0	0		
Silicon	2.248	0.056	5.556	6.979	-85.1	75	125	0	0		BS
Sodium	192.4	0.44	11.11	179.2	119	75	125	0	0		

Sample ID: 09-5678-01CMSD		SampType: MSD		TestCode: 6010_D		Run ID: ICP-OPTIMA 5300 DV_090813A		Prep Date: 8/10/2009		Units: mg/L	
		Batch ID: 20271		TestNo: SW6010B		FileID: 081309PM		Analysis Date: 8/14/2009		SeqNo: 891643	
Analyte	Result	LQL	SPK value	SPK Ref Val	%REC	LowLimit	HighLimit	RPD Ref Val	%RPD	RPDLimit	Qual
Calcium	86.04	0.43	11.11	75.51	94.7	75	125	86.89	0.991	20	
Iron	5.007	0.078	5.556	0.04291	90.1	75	125	5.011	0.0973	20	
Magnesium	19.07	0.17	11.11	9.287	88	75	125	19.19	0.632	20	
Manganese	1.898	0.0056	2.222	0.01756	84.6	75	125	1.899	0.0458	20	
Potassium	16.03	0.38	11.11	5.693	93	75	125	16.01	0.0765	20	
Silicon	1.854	0.056	5.556	6.979	-92.3	75	125	2.248	19.3	20	BS
Sodium	190.9	0.44	11.11	179.2	105	75	125	192.4	0.798	20	

Qualifiers:

U - Not detected at or above the Reporting Limit
 J - Analyte detected below quantitation limits
 S - Spike Recovery outside acceptance limits
 E - Extrapolated value, value exceeds calibration range.

R - RPD outside acceptance limits
 B - Analyte detected in the associated Method Blank
 H - Prep or analytical holding time exceeded
 X - See case narrative

Work Order: 09-5730
 Client Project ID: Complaint 2002 12599

ANALYTICAL QC SUMMARY REPORT

BatchID: 20344

Sample ID: MB-20344	SampType: MBLK	TestCode: 200.7_T	Run ID: ICP-OPTIMA 5300 DV_090818A	Prep Date: 8/14/2009	Units: mg/L						
	Batch ID: 20344	TestNo: E200.7, Rev.	FileID: 081709PM	Analysis Date: 8/19/2009	SeqNo: 894729						
Analyte	Result	LQL	SPK value	SPK Ref Val	%REC	LowLimit	HighLimit	RPD Ref Val	%RPD	RPDLimit	Qual

Silicon U 0.0500

Sample ID: LCS-20344	SampType: LCS	TestCode: 200.7_T	Run ID: ICP-OPTIMA 5300 DV_090818A	Prep Date: 8/14/2009	Units: mg/L						
	Batch ID: 20344	TestNo: E200.7, Rev.	FileID: 081709PM	Analysis Date: 8/19/2009	SeqNo: 894730						
Analyte	Result	LQL	SPK value	SPK Ref Val	%REC	LowLimit	HighLimit	RPD Ref Val	%RPD	RPDLimit	Qual

Silicon 4.411 0.0500 5 0.01261 88.2 85 115 0 0

Sample ID: 09-5730-01AMS	SampType: MS	TestCode: 6010_WT	Run ID: ICP-OPTIMA 5300 DV_090818A	Prep Date: 8/14/2009	Units: mg/L						
Client ID: York WW	Batch ID: 20344	TestNo: SW6010B	FileID: 081709PM	Analysis Date: 8/19/2009	SeqNo: 894732						
Analyte	Result	LQL	SPK value	SPK Ref Val	%REC	LowLimit	HighLimit	RPD Ref Val	%RPD	RPDLimit	Qual

Silicon 4.112 0.056 5.556 17.18 -235 75 125 0 0 S

Sample ID: 09-5730-01AMSD	SampType: MSD	TestCode: 6010_WT	Run ID: ICP-OPTIMA 5300 DV_090818A	Prep Date: 8/14/2009	Units: mg/L						
Client ID: York WW	Batch ID: 20344	TestNo: SW6010B	FileID: 081709PM	Analysis Date: 8/19/2009	SeqNo: 894735						
Analyte	Result	LQL	SPK value	SPK Ref Val	%REC	LowLimit	HighLimit	RPD Ref Val	%RPD	RPDLimit	Qual

Silicon 5.899 0.056 5.556 17.18 -203 75 125 4.112 35.7 20 SR

Qualifiers:

U - Not detected at or above the Reporting Limit
 J - Analyte detected below quantitation limits
 S - Spike Recovery outside acceptance limits
 E - Extrapolated value, value exceeds calibration range

R - RPD outside acceptance limits
 B - Analyte detected in the associated Method Blank
 H - Prep or analytical holding time exceeded
 X - See case narrative