

Calibration Report

Database File: Sample Oil Test.db
Database Path: C:\Program Files (x86)\Fluke\DPC Software\Sample Oil Test.db

Created By: Administrator
Created On: 10/15/2015 10:47:03 AM

General Test Calibration Report

Work Order: Sample Oil Test
Customer: Fluke
Asset: Sample Oil Test
Calibration Date: 10/15/2015 10:47:03 AM
Calibration Type: Full
Procedure Name: Sample Oil Test
Procedure Revision: 1.0
Temperature: 72.0
Humidity: 45.0
Technician: Administrator

Test Points: 5
Passed: 5
Failed: 0
Out Of Tolerance: 0

Sample Oil Test

Test Point Number: 1
Test Type: Pressure
Calibration Pump Operation: Manual
Pressure Module: 700G29

Nominal: 0.000 PSI
Applied: 0.000 PSI
Error: 0.000 PSI
Error %: 0.000 %
Tolerance: 0.000 PSI
Tolerance %: 0.000 %
Result: Pass

As Found: 0.000 PSI
As Left: 0.000 PSI

Test Points

Test Point Number: 2
Test Type: Pressure
Calibration Pump Operation: Manual
Pressure Module: 700G29

Nominal: 50.000 PSI
Applied: 50.000 PSI
Error: 0.000 PSI
Error %: 0.000 %
Tolerance: 0.000 PSI
Tolerance %: 0.000 %
Result: Pass

As Found: 50.000 PSI
As Left: 50.000 PSI

Test Points

Test Point Number: 3
Test Type: Pressure
Calibration Pump Operation: Manual
Pressure Module: 700G29

Nominal: 100.000 PSI
Applied: 100.000 PSI
Error: 0.000 PSI
Error %: 0.000 %
Tolerance: 0.000 PSI
Tolerance %: 0.000 %
Result: Pass

As Found: 100.000 PSI
As Left: 100.000 PSI

Test Points

Test Point Number: 4
Test Type: Pressure
Calibration Pump Operation: Manual
Pressure Module: 700G29

Nominal: 150.000 PSI
Applied: 150.000 PSI
Error: 0.000 PSI
Error %: 0.000 %
Tolerance: 0.000 PSI
Tolerance %: 0.000 %
Result: Pass

As Found: 150.000 PSI
As Left: 150.000 PSI

Test Points

Test Point Number: 5
Test Type: Pressure
Calibration Pump Operation: Manual
Pressure Module: 700G29

Nominal: 200.000 PSI
Applied: 200.000 PSI
Error: 0.000 PSI
Error %: 0.000 %
Tolerance: 0.000 PSI
Tolerance %: 0.000 %
Result: Pass

As Found: 200.000 PSI
As Left: 200.000 PSI

Transaction Information Report

Transaction Type: Create
Transaction Date: 10/15/2015 10:47:03 AM

Reference

Work Order: Sample Oil Test
Asset: Sample Oil Test
Customer: Fluke
Procedure: Sample Oil Test
Procedure Revision: 1.0

Reading

Temperature: 72.0
Humidity: 45.0
Technician: Administrator