

Session Report

1/22/2015

Information Panel

Name Grant Hurt 1G-14H G268
Start Time 1/22/2015 11:27:04 AM
Stop Time 1/22/2015 12:37:14 PM
Device Name BGN010014
Model Type SoundPro DL
Device Firmware Rev R.13F
Comments

Calibration History

Date	Calibration Action	Level	Cal. Model Type	Serial Number	Cert. Due Date
1/22/2015 11:26:48 AM	Calibration	114.0			
1/22/2015 12:37:38 PM	Verification	113.7			

Summary Data Panel

Description	Meter	Value	Description	Meter	Value
Leq	1	46.1 dB			
Exchange Rate	1	3 dB	Weighting	1	A
Response	1	SLOW	Bandwidth	1	OFF
Leq	2	63.2 dB			
Exchange Rate	2	3 dB	Weighting	2	C
Response	2	FAST			

Logged Data Chart

Grant Hurt 1G-14H G268: Logged Data Chart

