

Session Report

2/11/2015

Information Panel

Name Grant Salisbury-Elmquist
Start Time 2/11/2015 6:15:18 AM
Stop Time 2/11/2015 7:08:25 AM
Device Name BGN010014
Model Type SoundPro DL
Device Firmware Rev R.13F
Comments

Calibration History

Date	Calibration Action	Level	Cal. Model Type	Serial Number	Cert. Due Date
2/11/2015 6:14:59 AM	Calibration	114.0			
2/11/2015 7:09:29 AM	Verification	114.1			

Summary Data Panel

Description	Meter	Value	Description	Meter	Value
Leq	1	46.4 dB			
Exchange Rate	1	3 dB	Weighting	1	A
Response	1	SLOW	Bandwidth	1	OFF
Leq	2	61.3 dB			
Exchange Rate	2	3 dB	Weighting	2	C
Response	2	FAST			

Logged Data Chart

Grant Salisbury-Elmquist: Logged Data Chart

