

# Session Report

11/19/2015

## Information Panel


Name Waag111915  
Start Time 11/19/2015 3:45:39 PM  
Stop Time 11/19/2015 4:32:27 PM  
Device Name BGN010010  
Model Type SoundPro DL  
Device Firmware Rev R.13F  
Comments

## Summary Data Panel

Description	Meter	Value	Description	Meter	Value
Leq	1	53.7 dB			
Exchange Rate	1	3 dB	Weighting	1	A
Response	1	SLOW	Bandwidth	1	OFF
Leq	2	66.2 dB			
Exchange Rate	2	3 dB	Weighting	2	C
Response	2	FAST			

## Logged Data Chart

Waag111915: Logged Data Chart



## Calibration History

Date	Calibration Action	Level	Cal. Model Type	Serial Number	Cert. Due Date
11/19/2015 3:45:13 PM	Calibration	114.0			
11/19/2015 4:32:48 PM	Verification	114.0			