

Session Report

8/4/2016

Information Panel

Name	Ruh 8-4-2016
Start Time	8/4/2016 1:12:29 PM
Stop Time	8/4/2016 2:17:11 PM
Device Name	BGN010010
Model Type	SoundPro DL
Device Firmware Rev	R.13F
Comments	Winds out of the NE at approx 3-5 mph/Approx GPS 40.50912 -104.29152

Summary Data Panel

Description	Meter	Value	Description	Meter	Value
Leq	1	49.6 dB			
Exchange Rate	1	3 dB	Weighting	1	A
Response	1	FAST	Bandwidth	1	OFF
Leq	2	70 dB			
Exchange Rate	2	3 dB	Weighting	2	C
Response	2	FAST			

Calibration History

Date	Calibration Action	Level	Cal. Model Type	Serial Number	Cert. Due Date
8/4/2016 1:08:21 PM	Calibration	113.9			
8/4/2016 2:17:37 PM	Verification	113.9			

Logged Data Chart

Ruh 8-4-2016: Logged Data Chart

