

Session Report

6/28/2017

Information Panel


Name S005_BGN010010_28062017_092925
Start Time 6/28/2017 7:21:21 AM
Stop Time 6/28/2017 8:21:32 AM
Device Name BGN010010
Model Type SoundPro DL
Device Firmware Rev R.13F
Comments

Summary Data Panel

Description	Meter	Value	Description	Meter	Value
Leq	1	77.2 dB			
Exchange Rate	1	3 dB	Weighting	1	C
Response	1	SLOW	Bandwidth	1	OFF
Leq	2	77.2 dB			
Exchange Rate	2	3 dB	Weighting	2	C
Response	2	FAST			

Logged Data Chart

S005_BGN010010_28062017_092925: Logged Data Chart



Calibration History

Date	Calibration Action	Level	Cal. Model Type	Serial Number	Cert. Due Date
6/28/2017 7:17:50 AM	Calibration	114.0			
6/28/2017 8:22:49 AM	Verification	113.8			