



facility 439136
project 10243

Radium-224/226

Case Narrative

COGCC


PW NORM 2017 – 10048

Work Order Number: 1705202

1. This report consists of the analytical results and supporting documentation for one water sample received by ALS on 05/09/2017.
2. This sample was prepared according to the current revision of SOP 701. Modifications were made to the method as described on QASS 472112.
3. The sample was analyzed for the presence of ^{224}Ra and ^{226}Ra according to the current revision of SOP 714. The analyses were completed on 05/23/2017.
4. The analysis results for this sample are reported in units of pCi/L. The water sample was filtered prior to analysis.
5. Sample 1705202-1 contained a significant concentration of Ba, which required a reduced aliquant to be analyzed. For further information regarding this occurrence, please refer to QASS 472112 in section 6 of this report.
6. The requested MDC for Ra-224 and Ra-226 was not met for samples 1705202-1 and -1D, due to reduced aliquots of the samples taken for analysis. The samples were counted for a maximum count time of 1000 minutes and results are reported without further qualification. The results are flagged with an "M" and/or "M3" qualifier on the final reports. The reported activity with an "M3" qualifier exceeds the achieved MDC.
7. No further anomalous situations were encountered during the preparation or analysis of this sample. All remaining quality control criteria were met.




The data contained in the following report have been reviewed and approved by the personnel listed below. In addition, ALS certifies that the analyses reported herein are true, complete and correct within the limits of the methods employed.



Pik Yee Yuen
Radiochemistry Primary Data Reviewer

5/25/17
Date



Radiochemistry Final Data Reviewer

5/25/17
Date

Section 1

CHAIN OF CUSTODY

ALS -- Fort Collins

Sample Number(s) Cross-Reference Table

OrderNum: 1705202

Client Name: COGCC

Client Project Name: PW NORM 2017

Client Project Number: 10048

Client PO Number: CT 2017-3066

| Client Sample Number | Lab Sample Number | COC Number | Matrix | Date Collected | Time Collected |
|----------------------|-------------------|------------|--------|----------------|----------------|
| 439136 | 1705202-1 | | WATER | 09-May-17 | 12:50 |
| 439136 | 1705202-2 | | WATER | 09-May-17 | 12:50 |



MTF: (800) 443-1511 PH: (970) 490-1511 FX: (970) 490-1522

Chain-of-Custody

Turnaround time for samples received after 2 p.m. will be calculated beginning from the next business day.

Turnaround time for samples received Saturday will be calculated beginning from the next business day.

ALS WORKORDER #

1705202

[illegible]



Chain-of-Custody

Turnaround time for samples received after 2 p.m. will be calculated beginning from the next business day.

ALS WORKORDER #

1705202

[illegible]



ALS Environmental - Fort Collins
CONDITION OF SAMPLE UPON RECEIPT FORM

Client: COGCC

Workorder No: 1705202

Project Manager: SS

Initials: CAT Date: 5-10-17

| | | | |
|--|-----------------|------------|-----------|
| 1. Does this project require any special handling in addition to standard ALS procedures? | | YES | <u>NO</u> |
| 2. Are custody seals on shipping containers intact? | <u>NONE</u> | YES | NO |
| 3. Are Custody seals on sample containers intact? | <u>NONE</u> | YES | NO |
| 4. Is there a COC (Chain-of-Custody) present or other representative documents? | | <u>YES</u> | NO |
| 5. Are the COC and bottle labels complete and legible? | | <u>YES</u> | NO |
| 6. Is the COC in agreement with samples received? (IDs, dates, times, no. of samples, no. of containers, matrix, requested analyses, etc.) | | <u>YES</u> | NO |
| 7. Were airbills / shipping documents present and/or removable? | <u>DROP OFF</u> | YES | NO |
| 8. Are all aqueous samples requiring preservation preserved correctly? (excluding volatiles) | N/A | YES | <u>NO</u> |
| 9. Are all aqueous non-preserved samples pH 4-9? | N/A | <u>YES</u> | NO |
| 10. Is there sufficient sample for the requested analyses? | | <u>YES</u> | NO |
| 11. Were all samples placed in the proper containers for the requested analyses? | | <u>YES</u> | NO |
| 12. Are all samples within holding times for the requested analyses? | | <u>YES</u> | NO |
| 13. Were all sample containers received intact? (not broken or leaking, etc.) | | <u>YES</u> | NO |
| 14. Are all samples requiring no headspace (VOC, GRO, RSK/MEE, Rx CN/S, radon) headspace free? Size of bubble: ____ < green pea ____ > green pea | N/A | <u>YES</u> | NO |
| 15. Do any water samples contain sediment? Amount Amount of sediment: ____ dusting ____ moderate ____ heavy | N/A | YES | <u>NO</u> |
| 16. Were the samples shipped on ice? | | <u>YES</u> | <u>NO</u> |
| 17. Were cooler temperatures measured at 0.1-6.0°C? IR gun used*: <u>#2</u> #4 | | <u>YES</u> | <u>NO</u> |
| Cooler #: <u>1</u> <u>2</u> | | | |
| Temperature (°C): <u>6.0</u> <u>Amb</u> | | | |
| No. of custody seals on cooler: <u>0</u> <u>0</u> | | | |
| External µR/hr reading: <u>NA</u> <u>NA</u> | | | |
| Background µR/hr reading: <u>NA</u> | | | |
| Were external µR/hr readings ≤ two times background and within DOT acceptance criteria? YES / NO / <u>NA</u> (If no, see Form 008.) | | | |

Additional Information: PROVIDE DETAILS BELOW FOR A NO RESPONSE TO ANY QUESTION ABOVE, EXCEPT #1 AND #16.

All RAD samples in the Amb cooler.

Added 3.5 Ml HNO₃ to EA. RAD and TOTAL metals bottle. Final pH < 2. HNO₃ lot no. 152495.

If applicable, was the client contacted? YES / NO / NA Contact: _____

Date/Time: _____

Project Manager Signature / Date: _____

*IR Gun #2: Oakton, SN 29922500201-0066

*IR Gun #4: Oakton, SN 2372220101-0002

Section 2



SAMPLE RESULTS SUMMARY

Isotopic Radium by Alpha Spectroscopy Sample Results Summary

Client Name: COGCC
Client Project Name: PW NORM 2017
Client Project Number: 10048
Laboratory Name: ALS -- Fort Collins
PAI Work Order: 1705202

Page: 1 of 1
Reported on: Thursday, May 25, 2017
9:33:58 AM

| Lab Sample ID | Client Sample ID | Sample Type | Nuclide | Result +/- 2 s TPU | MDC | DL | Units | Matrix | Prep Batch | Date Analyze | Flags |
|---------------|------------------|-------------|---------|---------------------|----------|----|-------|--------|-------------|--------------|-------|
| 1705202-1 | 439136 | Sample | Ra-224 | 5.1E+00 +/- 6.7E+00 | 2.81E+01 | NA | pCi/l | WATER | RAS170511-1 | 5/12/2017 | U,M |
| 1705202-1 | 439136 | Sample | Ra-226 | 1.25E+02 +/- 1E+01 | 3E+00 | NA | pCi/l | WATER | RAS170511-1 | 5/12/2017 | M3 |

Comments:

Data Package ID: RAS1705202-1

Qualifiers/Flags:

- U - Result is less than the sample specific MDC.
- LT - Result is less than Requested MDC, greater than sample specific MDC.
- Y1 - Chemical Yield is in control at 100-110%. Quantitative Yield is assumed.
- Y2 - Chemical Yield outside default limits.
- M - The requested MDC was not met.
- M3 - The requested MDC was not met, but the reported activity is greater than the reported MDC.

Abbreviations:

- TPU - Total Propagated Uncertainty
- MDC - Sample specific Minimum Detectable Concentration
- BDL - Below Detection Limit

Section 3

QC RESULTS SUMMARY



Isotopic Radium by Alpha Spectroscopy

PAI 714_Ra226 Rev 13

Method Blank Results

Lab Name: ALS -- Fort Collins

Work Order Number: 1705202

Client Name: COGCC

ClientProject ID: PW NORM 2017 10048

Lab ID: RAS170511-1MB

Sample Matrix: WATER

Prep SOP: PAI 701 Rev 1

Date Collected: 11-May-17

Date Prepared: 11-May-17

Date Analyzed: 12-May-17

Prep Batch: RAS170511-1

QCBatchID: RAS170511-1-1

Run ID: RAS170511-1

Count Time: 1000 minutes

Final Aliquot: 500 ml

Result Units: pCi/l

File Name: Spectrum #1

| CASNO | Target Nuclide | Result +/- 2 s TPU | MDC | Requested MDC | DL | Lab Qualifier |
|------------|----------------|--------------------|----------|---------------|----|---------------|
| 13233-32-4 | Ra-224 | -2E-03 +/- 4.2E-02 | 1.85E-01 | 1E+00 | NA | U |
| 13982-63-3 | Ra-226 | -7E-03 +/- 1.2E-02 | 5.2E-02 | 1E+00 | NA | U |

Chemical Yield Summary

| Carrier/Tracer | Amount Added | Result | Units | Yield | Control Limits | Flag |
|----------------|--------------|-----------|-------|-------|----------------|------|
| At-217 | 1.080E+01 | 9.650E+00 | pCi/l | 89.4 | 60 - 100 % | |

Comments:

Qualifiers/Flags:

U - Result is less than the sample specific MDC.

Y1 - Chemical Yield is in control at 100-110%. Quantitative Yield is assumed.

Y2 - Chemical Yield outside default limits.

LT - Result is less than Requested MDC, greater than sample specific MDC.

Abbreviations:

TPU - Total Propagated Uncertainty

MDC - Sample specific Minimum Detectable Concentration

BDL - Below Detection Limit

M - Requested MDC not met.

B - Analyte concentration greater than MDC.

B3 - Analyte concentration greater than MDC but less than Requested MDC.

DL - Decision Level

Data Package ID: RAS1705202-1

Isotopic Radium by Alpha Spectroscopy

PAI 714_Ra226 Rev 13
Laboratory Control Sample(s)

Lab Name: ALS -- Fort Collins
Work Order Number: 1705202
Client Name: COGCC
ClientProject ID: PW NORM 2017 10048

Lab ID: RAS170511-1LCS

Sample Matrix: WATER
Prep SOP: PAI 701 Rev 1
Date Collected: 11-May-17
Date Prepared: 11-May-17
Date Analyzed: 12-May-17

Prep Batch: RAS170511-1
QCBatchID: RAS170511-1-1
Run ID: RAS170511-1
Count Time: 1000 minutes

Final Aliquot: 500 ml
Result Units: pCi/l
File Name: Spectrum #1

| CASNO | Target Nuclide | Results +/- 2s TPU | MDC | Spike Added | % Rec | Control Limits | Lab Qualifier |
|------------|----------------|----------------------|-------|-------------|-------|----------------|---------------|
| 13233-32-4 | Ra-224 | 9.33E+00 +/- 6.4E-01 | 2E-01 | 8.950E+00 | 104 | 75 - 125 | P |
| 13982-63-3 | Ra-226 | 9.5E+00 +/- 6.5E-01 | 6E-02 | 9.210E+00 | 103 | 75 - 125 | P |

Chemical Yield Summary

| Carrier/Tracer | Amount Added | Result | Units | Yield | Control Limits | Flag |
|----------------|--------------|-----------|-------|-------|----------------|------|
| At-217 | 1.080E+01 | 9.550E+00 | pCi/l | 88.5 | 60 - 100 % | |

Comments:

Qualifiers/Flags:

U - Result is less than the sample specific MDC.
LT - Result is less than Requested MDC, greater than sample specific MDC.
Y1 - Chemical Yield is in control at 100-110%. Quantitative Yield is assumed.
Y2 - Chemical Yield outside default limits.
L - LCS Recovery below lower control limit.
H - LCS Recovery above upper control limit.
P - LCS Recovery within control limits.
M - The requested MDC was not met.
M3 - The requested MDC was not met, but thereported activity is greater than the reported MDC.

Abbreviations:

TPU - Total Propagated Uncertainty
MDC - Minimum Detectable Concentration

Data Package ID: RAS1705202-1

Isotopic Radium by Alpha Spectroscopy

PAI 714_Ra226 Rev 13

Duplicate Sample Results (DER)

Lab Name: ALS -- Fort Collins

Work Order Number: 1705202

Client Name: COGCC

ClientProject ID: PW NORM 2017 10048

Field ID: 439136

Lab ID: 1705202-1DUP

Sample Matrix: WATER

Prep SOP: PAI 701 Rev 1

Date Collected: 09-May-17

Date Prepared: 11-May-17

Date Analyzed: 12-May-17

Prep Batch: RAS170511-1

QCBatchID: RAS170511-1-1

Run ID: RAS170511-1

Count Time: 1000 minutes

Report Basis: Filtered

Final Aliquot: 6.00 ml

Prep Basis: Filtered

Moisture(%): NA

Result Units: pCi/l

File Name: Spectrum #1

| CASNO | Analyte | Sample | | | | Duplicate | | | DER | DER Lim |
|------------|---------|---------------------|---------|----------|-------|----------------------|---------|---------|--------|---------|
| | | Result +/- | 2 s TPU | MDC | Flags | Result +/- | 2 s TPU | MDC | | |
| 13233-32-4 | Ra-224 | 5.1E+00 +/- 6.7E+00 | | 2.81E+01 | U,M | -4E+00 +/- 6.4E+00 | | 2.9E+01 | 0.9907 | 2 |
| 13982-63-3 | Ra-226 | 1.25E+02 +/- 1E+01 | | 3E+00 | M3 | 1.12E+02 +/- 9.4E+00 | | 3.1E+00 | 0.9313 | 2 |

Comments: This sample was filtered prior to analysis.

Duplicate Qualifiers/Flags:

U - Result is less than the sample specific MDC.

Y1 - Chemical Yield is in control at 100-110%. Quantitative yield is assumed.

Y2 - Chemical Yield outside default limits.

W - DER is greater than Warning Limit of 1.42

D - DER is greater than Control Limit of 2

LT - Result is less than Request MDC, greater than sample specific MDC

M - Requested MDC not met.

M3 - The requested MDC was not met, but the reported activity is greater than the reported MDC.

L - LCS Recovery below lower control limit.

H - LCS Recovery above upper control limit.

P - LCS, Matrix Spike Recovery within control limits.

N - Matrix Spike Recovery outside control limits

Abbreviations:

TPU - Total Propagated Uncertainty

DER - Duplicate Error Ratio

BDL - Below Detection Limit

NR - Not Reported

Data Package ID: RAS1705202-1

Date Printed: Thursday, May 25, 2017

ALS -- Fort Collins

LIMS Version: 6.842

Page 1 of 1

Section 4

INDIVIDUAL SAMPLE RESULTS

4

Isotopic Radium by Alpha Spectroscopy

PAI 714_Ra226 Rev 13

Sample Results

Lab Name: ALS -- Fort Collins

Work Order Number: 1705202

Client Name: COGCC

ClientProject ID: PW NORM 2017 10048

Field ID: 439136

Lab ID: 1705202-1

Sample Matrix: WATER

Prep SOP: PAI 701 Rev 1

Date Collected: 09-May-17

Date Prepared: 11-May-17

Date Analyzed: 12-May-17

Prep Batch: RAS170511-1

QCBatchID: RAS170511-1-1

Run ID: RAS170511-1

Count Time: 1000 minutes

Report Basis: Filtered

Final Aliquot: 6.00 ml

Prep Basis: Filtered

Moisture(%): NA

Result Units: pCi/l

File Name: Spectrum #1

| CASNO | Target Nuclide | Result +/- 2 s TPU | MDC | Requested MDC | DL | Lab Qualifier |
|------------|----------------|---------------------|----------|---------------|----|---------------|
| 13233-32-4 | Ra-224 | 5.1E+00 +/- 6.7E+00 | 2.81E+01 | 1E+00 | NA | U,M |
| 13982-63-3 | Ra-226 | 1.25E+02 +/- 1E+01 | 3E+00 | 1E+00 | NA | M3 |

Chemical Yield Summary

| Carrier/Tracer | Amount Added | Result | Units | Yield | Control Limits | Flag |
|----------------|--------------|-----------|-------|-------|----------------|------|
| At-217 | 8.990E+02 | 7.760E+02 | pCi/l | 92.4 | 60 - 100 % | |

Comments: This sample was filtered prior to analysis.

Qualifiers/Flags:

U - Result is less than the sample specific MDC.

Y1 - Chemical Yield is in control at 100-110%. Quantitative Yield is assumed.

Y2 - Chemical Yield outside default limits.

LT - Result is less than Requested MDC, greater than sample specific MDC.

M3 - The requested MDC was not met, but the reported activity is greater than the reported MDC.

M - The requested MDC was not met.

Abbreviations:

TPU - Total Propagated Uncertainty

MDC - Sample specific Minimum Detectable Concentration

BDL - Below Detection Limit

DL - Decision Level

Data Package ID: RAS1705202-1

Isotopic Radium by Alpha Spectroscopy

PAI 714_Ra226 Rev 13 Sample Duplicate Results

Lab Name: ALS -- Fort Collins
Work Order Number: 1705202
Client Name: COGCC
ClientProject ID: PW NORM 2017 10048

| | |
|-----------|--------------|
| Field ID: | 439136 |
| Lab ID: | 1705202-1DUP |

Sample Matrix: WATER
Prep SOP: PAI 701 Rev 1
Date Collected: 09-May-17
Date Prepared: 11-May-17
Date Analyzed: 12-May-17

Prep Batch: RAS170511-1
QCBatchID: RAS170511-1-1
Run ID: RAS170511-1
Count Time: 1000 minutes
Report Basis: Filtered

Final Aliquot: 6.00 ml
Prep Basis: Filtered
Moisture(%): NA
Result Units: pCi/l
File Name: Spectrum #1

| CASNO | Target Nuclide | Result +/- 2 s TPU | MDC | Requested MDC | DL | Lab Qualifier |
|------------|----------------|----------------------|---------|---------------|----|---------------|
| 13233-32-4 | Ra-224 | -4E+00 +/- 6.4E+00 | 2.9E+01 | 1E+00 | NA | U,M |
| 13982-63-3 | Ra-226 | 1.12E+02 +/- 9.4E+00 | 3.1E+00 | 1E+00 | NA | M3 |

Chemical Yield Summary

| Carrier/Tracer | Amount Added | Result | Units | Yield | Control Limits | Flag |
|----------------|--------------|-----------|-------|-------|----------------|------|
| At-217 | 8.990E+02 | 8.310E+02 | pCi/l | 86.3 | 60 - 100 % | |

Comments: This sample was filtered prior to analysis.

Qualifiers/Flags:

U - Result is less than the sample specific MDC.
Y1 - Chemical Yield is in control at 100-110%. Quantitative yield is assumed.
Y2 - Chemical Yield outside default limits.
LT - Result is less than Requested MDC, greater than sample specific MDC.
M - The requested MDC was not met.
M3 - The requested MDC was not met, but thereported activity is greater than the reported MDC.
W - DER is greater than Warning Limit of 1.42
D - DER is greater than Control Limit of 2

Abbreviations:

TPU - Total Propagated Uncertainty
MDC - Sample specific Minimum Detectable Concentration
BDL - Below Detection Limit
DL - Decision Level

Data Package ID: RAS1705202-1

Date Printed: Thursday, May 25, 2017

ALS -- Fort Collins
LIMS Version: 6.842

Page 1 of 1

Section 5

RAW DATA

5

Isotopic Radium by Alpha Spectroscopy Raw Data Report

Laboratory Name: ALS -- Fort Collins

Prep SOP: PAI 701

Reported on: Wednesday, May 24, 2017

PAI Work Order: 1705202

Analytical SOP: PAI 714_Ra226

1:14:14 PM

| Sample ID QC Type | Nuclide Type | Sample Date/Time | Prep Batch QC BatchID | Ingrowth Date /Time | Decay Date/Time | Matrix %Moist. | Samp Aliq Analy Aliq | Inst ID Det ID | AnRunID File Name | Count Date/Time | Net Cnts Bkg Cnts | BaseEff Bkg(min) | CndDur(min) Yield | Activity +/- 2 s TPU | MDC DeclEv | ReportUnits ReportBasis | DER RPD | %Spk. Recov Flags |
|----------------------|------------------------|--------------------------|------------------------------|------------------------|--------------------|-------------------|-------------------------|-------------------|----------------------------|----------------------|----------------------|---------------------|----------------------|-------------------------|---------------|----------------------------|------------|----------------------|
| 1705202-1 | Ra-224 Trg. Analyte | 5/9/2017 12:50:00 PM | RAS170511-1 RAS170511-1-1 | NA NA | NA NA | WATER NA | 6 ml 6 ml | AlphaSpec2 88 | RAS170511-1 Spectrum #1 | 5/12/2017 1:44 PM | 64,000 36,000 | 28.84% 1000 | 1000 92.4% | 5.1E+00 6.7E+00 | 2.81E+01 | pCi/l Filtered | NA NA | NA U,M |
| 1705202-1 | Ra-226 Trg. Analyte | 5/9/2017 12:50:00 PM | RAS170511-1 RAS170511-1-1 | NA NA | NA NA | WATER NA | 6 ml 6 ml | AlphaSpec2 88 | RAS170511-1 Spectrum #1 | 5/12/2017 1:44 PM | 420,000 2,000 | 28.84% 1000 | 1000 92.4% | 1.25E+02 1E+01 | 3E+00 | pCi/l Filtered | NA NA | NA M3 |
| 1705202-1 | Ra-224 Trg. Analyte | 5/9/2017 12:50:00 PM | RAS170511-1 RAS170511-1-1 | NA NA | NA NA | WATER NA | 6 ml 6 ml | AlphaSpec2 89 | RAS170511-1 Spectrum #1 | 5/12/2017 1:44 PM | 49,000 32,000 | 29.52% 1000 | 1000 86.3% | -4E+00 6.4E+00 | 2.9E+01 | pCi/l Filtered | 0.99 NA | NA U,M |
| 1705202-1 | Ra-226 Trg. Analyte | 5/9/2017 12:50:00 PM | RAS170511-1 RAS170511-1-1 | NA NA | NA NA | WATER NA | 6 ml 6 ml | AlphaSpec2 89 | RAS170511-1 Spectrum #1 | 5/12/2017 1:44 PM | 360,000 2,000 | 29.52% 1000 | 1000 86.3% | 1.12E+02 9.4E+00 | 3.1E+00 | pCi/l Filtered | 0.93 NA | NA M3 |
| RAS170511-1 | Ra-224 Trg. Analyte | 5/11/2017 12:00:00 PM | RAS170511-1 RAS170511-1-1 | NA NA | NA NA | WATER NA | 500 ml 500 ml | AlphaSpec2 81 | RAS170511-1 Spectrum #1 | 5/12/2017 1:44 PM | 46,000 28,000 | 28.35% 1000 | 1000 89.4% | -2E-03 4.2E-02 | 1.85E-01 | pCi/l As Received | NA NA | NA U |
| RAS170511-1 | Ra-226 Trg. Analyte | 5/11/2017 12:00:00 PM | RAS170511-1 RAS170511-1-1 | NA NA | NA NA | WATER NA | 500 ml 500 ml | AlphaSpec2 81 | RAS170511-1 Spectrum #1 | 5/12/2017 1:44 PM | -2,000 5,000 | 28.35% 1000 | 1000 89.4% | -7E-03 1.2E-02 | 5.2E-02 | pCi/l As Received | NA NA | NA U |
| RAS170511-1 | Ra-224 Trg. Analyte | 5/11/2017 12:00:00 PM | RAS170511-1 RAS170511-1-1 | NA NA | NA NA | WATER NA | 500 ml 500 ml | AlphaSpec2 82 | RAS170511-1 Spectrum #1 | 5/12/2017 1:44 PM | 2382,000 21,000 | 28.92% 1000 | 1000 88.5% | 9.33E+00 6.4E-01 | 2E-01 | pCi/l As Received | NA NA | 104 P |
| RAS170511-1 | Ra-226 Trg. Analyte | 5/11/2017 12:00:00 PM | RAS170511-1 RAS170511-1-1 | NA NA | NA NA | WATER NA | 500 ml 500 ml | AlphaSpec2 82 | RAS170511-1 Spectrum #1 | 5/12/2017 1:44 PM | 2565,000 9,000 | 28.92% 1000 | 1000 88.5% | 9.5E+00 6.5E-01 | 6E-02 | pCi/l As Received | NA NA | 103 P |

Comments:

Data Package ID: RAS1705202-1

Qualifiers/Flags:

- U - Result is less than the sample specific MDC.
- Y1 - Chemical Yield is in control at 100-110%. Quantitative yield is assumed.
- Y2 - Chemical Yield outside default limits.
- W - DER is greater than Warning Limit of 1.42
- D - DER is greater than Control Limit of 2
- + - Duplicate RPD not within limits.
- LT - Result is less than Request MDC, greater than sample specific MDC
- * - Aliquot Basis is 'As Received' while the Report Basis is 'Dry Weight'
- # - Aliquot Basis is 'Dry Weight' while the Report Basis is 'As Received'

M - Requested MDC not met.

M3 - The requested MDC was not met, but the reported activity is greater than the reported MDC.

L - LCS Recovery below lower control limit.

H - LCS Recovery above upper control limit.

P - LCS, Matrix Spike Recovery within control limits.

N - Matrix Spike Recovery outside control limits

NC - Not Calculated for duplicate results less than 5 times MDC

B - Analyte concentration greater than MDC.

B3 - Analyte concentration greater than MDC but less than Requested MDC.

Notes:

- 1) The Tracer results are not yield corrected (i.e. activity measured not activity added).
- 2) Where sample time is not available, 12:00 PM (Mountain) is used for decay correction.

Abbreviations:

TR- Tracer TA - Target Analyte
TPU - Total Propagated Uncertainty
MDC - Minimum Detectable Concentration
DER - Duplicate Error Ratio
BDL - Below Detection Limit

ALS Laboratory Group - Fort Collins

Alpha-Spectroscopy Analysis Report

Sample: 1705202-1
Spectrum #1 Analysis #1

Sample Size : 0.01

Detector: 88
Batch Name: RAS170511-1_A
Nuclide Library: Radium
Analysis Method: ROI Analysis, Set Name = Ra224/Ra226
ROI Set: Ra224/Ra226

Acquisition

Acquisition Start Date: 5/12/2017 1:44:55PM
Live Time: 1,000.00 min.
Real Time: 1,000.09 min.
Dead Time: 0.01 %

Calibration

Bkgd Info: Sample: B17051088; Det: 88; Spectrum #1; 5/10/2017 10:45:37 AM

Calibration Date: 5/10/2017 10:12:28AM

Efficiency Calibration: C17051088

Efficiency: 28.84% +/- 0.20% TPU(2 sigma)

Energy Calibration: C17051088

Energy Cal: Gain = 9.9003 keV / Ch

Offset = 3,021.28 keV

Quadratic = 0.0000 keV / Ch²


Tracer

Tracer Name: At-217

Tracer Activity: 20.00 DPM/mL x (Vol.)1.00 mL = 20.00 DPM

Tracer Nuclide: At-217

Tracer Recovery: 3.89%



Nuclide Summary (ROI)

| Nuclide | Peak Energy keV | ROI Start keV | ROI End keV | FWHM keV | B.R. % | Gross Counts | Bkgd Counts | Net Counts | Activity pCi/L | 1.00Sigma TPU pCi/L | Critical Level pCi/L | MDA pCi/L |
|---------|-----------------|---------------|-------------|----------|--------|--------------|-------------|------------|----------------|---------------------|----------------------|-----------|
| Ra-226 | 4783.5 | 4506.3 | 4833.0 | 28.0 | 100.1 | 422.00 | 2.00 | 420.00 | 2.8E+003 | 3.0E+002 | 2.2E+001 | 6.2E+001 |
| Ra-224 | 5684.5 | 5545.9 | 5743.9 | 23.2 | 95.1 | 100.00 | 36.00 | 64.00 | 4.5E+002 | 9.2E+001 | 9.8E+001 | 2.2E+002 |
| Ac-225 | 5823.1 | 5753.8 | 5912.2 | 61.4 | 78.3 | 262.00 | 48.00 | 214.00 | 1.8E+003 | 2.3E+002 | 1.4E+002 | 3.0E+002 |
| Fr-221 | 6328.0 | 6050.8 | 6387.4 | 47.4 | 100.0 | 361.00 | 63.00 | 298.00 | 1.3E+004 | 1.6E+003 | 8.1E+002 | 1.7E+003 |
| Po-216 | 6763.6 | 6634.9 | 6852.7 | 81.7 | 100.0 | 33.00 | 25.00 | 8.00 | 5.4E+001 | 5.1E+001 | 7.8E+001 | 1.7E+002 |
| At-217 | 7080.4 | 6862.6 | 7159.6 | 49.4 | 99.9 | 265.00 | 41.00 | 224.00 | 5.8E+001 | 4.6E+000 | 1.0E+002 | 2.2E+002 |

JA

DP

ALS Laboratory Group - Fort Collins

Alpha-Spectroscopy Analysis Report

Sample: 1705202-1
Spectrum #1 Analysis #1

Sample Size : 0.01

Detector: 88
Batch Name: RAS170511-1_B
Nuclide Library: Radium
Analysis Method: ROI Analysis, Set Name = Ra224/Ra226
ROI Set: Ra224/Ra226

Acquisition Start Date: 5/23/2017 6:44:23AM
Live Time: 300.00 min.
Real Time: 300.17 min.
Dead Time: 0.06 %

Bkgd Info: Sample: B17051088; Det: 88; Spectrum #1; 5/10/2017 10:45:37 AM

Calibration Date: 5/10/2017 10:12:28AM

Efficiency Calibration: C17051088

Efficiency: 28.84% +/- 0.20% TPU(2 sigma)

Energy Calibration: C17051088


Energy Cal: Gain = 9.9003 keV / Ch
Offset = 3,021.28 keV
Quadratic = 0.0000 keV / Ch²

Tracer Name: At-217

Tracer Activity: 20.00 DPM/mL x (Vol.)1.00 mL = 20.00 DPM

Tracer Nuclide: At-217

Tracer Recovery: 55.40%



| Nuclide Summary (ROI) | | | | | | | | | | | | |
|-----------------------|-----------------|---------------|-------------|----------|--------|--------------|-------------|------------|----------------|---------------------|----------------------|-----------|
| Nuclide | Peak Energy keV | ROI Start keV | ROI End keV | FWHM keV | B.R. % | Gross Counts | Bkgd Counts | Net Counts | Activity pCi/L | 1.00Sigma TPU pCi/L | Critical Level pCi/L | MDA pCi/L |
| Ra-226 | 4783.5 | 4506.3 | 4833.0 | 26.0 | 100.1 | 109.00 | 0.60 | 108.40 | 1.7E+002 | 1.9E+001 | 2.3E+000 | 8.8E+000 |
| Ra-224 | 5684.5 | 5545.9 | 5743.9 | 12.0 | 95.1 | 225.00 | 10.80 | 214.20 | 3.5E+002 | 3.3E+001 | 1.0E+001 | 2.5E+001 |
| Ac-225 | 5823.1 | 5753.8 | 5912.2 | 74.9 | 78.3 | 848.00 | 14.40 | 833.60 | 1.7E+003 | 1.2E+002 | 1.4E+001 | 3.4E+001 |
| Fr-221 | 6328.0 | 6050.8 | 6387.4 | 46.3 | 100.0 | 1,036.00 | 18.90 | 1,017.10 | 3.1E+003 | 2.2E+002 | 2.5E+001 | 5.9E+001 |
| Po-216 | 6763.6 | 6634.9 | 6852.7 | 24.1 | 100.0 | 13.00 | 7.50 | 5.50 | 8.6E+000 | 7.1E+000 | 8.0E+000 | 2.0E+001 |
| At-217 | 7080.4 | 6862.6 | 7159.6 | 48.8 | 99.9 | 970.00 | 12.30 | 957.70 | 8.3E+002 | 2.7E+001 | 1.0E+001 | 2.5E+001 |

JA

JM

ALS Laboratory Group - Fort Collins


Alpha-Spectroscopy Analysis Report

Sample: 1705202-1D Sample Size : 0.01
Spectrum #1 Analysis #1

Detector: 89 Acquisition Start Date: 5/12/2017 1:44:55PM
Batch Name: RAS170511-1_A Live Time: 1,000.00 min.
Nuclide Library: Radium Real Time: 1,000.01 min.
Analysis Method: ROI Analysis, Set Name = Ra224/Ra226 Dead Time: 0.00 %
ROI Set: Ra224/Ra226

Calibration
Bkgd Info: Sample: B17051089; Det: 89; Spectrum #1; 5/10/2017 10:46:22 AM
Calibration Date: 5/10/2017 10:12:37AM Energy Calibration: C17051089
Efficiency Calibration: C17051089 Energy Cal: Gain = 9.9003 keV / Ch
Efficiency: 29.52% +/- 0.18% TPU(2 sigma) Offset = 3,021.28 keV
Quadratic = 0.0000 keV / Ch²

Tracer Name: At-217 Tracer Nuclide: At-217
Tracer Activity: 20.00 DPM/mL x (Vol.)1.00 mL = 20.00 DPM Tracer Recovery: 4.22%



Nuclide Summary (ROI)

| Nuclide | Peak Energy keV | ROI Start keV | ROI End keV | FWHM keV | B.R. % | Gross Counts | Bkgd Counts | Net Counts | Activity pCi/L | 1.00Sigma TPU pCi/L | Critical Level pCi/L | MDA pCi/L |
|---------|-----------------|---------------|-------------|----------|--------|--------------|-------------|------------|----------------|---------------------|----------------------|-----------|
| Ra-226 | 4783.5 | 4506.3 | 4833.0 | 26.9 | 100.1 | 362.00 | 2.00 | 360.00 | 2.2E+003 | 2.2E+002 | 2.0E+001 | 5.6E+001 |
| Ra-224 | 5684.5 | 5545.9 | 5743.9 | 45.0 | 95.1 | 81.00 | 32.00 | 49.00 | 3.1E+002 | 7.3E+001 | 8.4E+001 | 1.8E+002 |
| Ac-225 | 5823.1 | 5753.8 | 5912.2 | 71.0 | 78.3 | 255.00 | 26.00 | 229.00 | 1.8E+003 | 2.0E+002 | 9.1E+001 | 2.0E+002 |
| Fr-221 | 6328.0 | 6050.8 | 6387.4 | 47.8 | 100.0 | 313.00 | 54.00 | 259.00 | 1.0E+004 | 1.2E+003 | 6.7E+002 | 1.5E+003 |
| Po-216 | 6763.6 | 6634.9 | 6852.7 | 23.8 | 100.0 | 40.00 | 22.00 | 18.00 | 1.1E+002 | 4.8E+001 | 6.6E+001 | 1.5E+002 |
| At-217 | 7080.4 | 6862.6 | 7159.6 | 45.3 | 99.9 | 273.00 | 24.00 | 249.00 | 6.3E+001 | 4.4E+000 | 6.9E+001 | 1.5E+002 |

JA

JP


ALS Laboratory Group - Fort Collins

Alpha-Spectroscopy Analysis Report

| | |
|---|--|
| Sample: 1705202-1D Spectrum #1 Analysis #1 | Sample Size : 0.01 |
| Detector: 89 Batch Name: RAS170511-1_B Nuclide Library: Radium Analysis Method: ROI Analysis, Set Name = Ra224/Ra226 ROI Set: Ra224/Ra226 | Acquisition Start Date: 5/23/2017 6:44:24AM Live Time: 300.00 min. Real Time: 300.10 min. Dead Time: 0.03 % |

| | |
|---|---|
| Bkgd Info: Sample: B17051089; Det: 89; Spectrum #1; 5/10/2017 10:46:22 AM Calibration Date: 5/10/2017 10:12:37AM Efficiency Calibration: C17051089 Efficiency: 29.52% +/- 0.18% TPU(2 sigma) | Calibration Energy Calibration: C17051089 Energy Cal: Gain = 9.9003 keV / Ch Offset = 3,021.28 keV Quadratic = 0.0000 keV / Ch ² |
|---|---|

| | |
|--|---|
| Tracer Name: At-217 Tracer Activity: 20.00 DPM/mL x (Vol.)1.00 mL = 20.00 DPM | Tracer Tracer Nuclide: At-217 Tracer Recovery: 51.76% |
|--|---|



| Nuclide Summary (ROI) | | | | | | | | | | | | |
|-----------------------|-----------------|---------------|-------------|----------|--------|--------------|-------------|------------|----------------|---------------------|----------------------|-----------|
| Nuclide | Peak Energy keV | ROI Start keV | ROI End keV | FWHM keV | B.R. % | Gross Counts | Bkgd Counts | Net Counts | Activity pCi/L | 1.00Sigma TPU pCi/L | Critical Level pCi/L | MDA pCi/L |
| Ra-226 | 4783.5 | 4506.3 | 4833.0 | 24.4 | 100.1 | 103.00 | 0.60 | 102.40 | 1.7E+002 | 1.9E+001 | 2.4E+000 | 9.2E+000 |
| Ra-224 | 5684.5 | 5545.9 | 5743.9 | 13.0 | 95.1 | 200.00 | 9.60 | 190.40 | 3.3E+002 | 3.2E+001 | 1.0E+001 | 2.5E+001 |
| Ac-225 | 5823.1 | 5753.8 | 5912.2 | 61.5 | 78.3 | 789.00 | 7.80 | 781.20 | 1.6E+003 | 1.1E+002 | 1.1E+001 | 2.8E+001 |
| Fr-221 | 6328.0 | 6050.8 | 6387.4 | 46.4 | 100.0 | 976.00 | 16.20 | 959.80 | 3.1E+003 | 2.2E+002 | 2.4E+001 | 5.8E+001 |
| Po-216 | 6763.6 | 6634.9 | 6852.7 | 18.4 | 100.0 | 15.00 | 6.60 | 8.40 | 1.4E+001 | 7.7E+000 | 7.9E+000 | 2.0E+001 |
| At-217 | 7080.4 | 6862.6 | 7159.6 | 46.6 | 99.9 | 923.00 | 7.20 | 915.80 | 7.8E+002 | 2.6E+001 | 8.3E+000 | 2.1E+001 |

ALS Laboratory Group - Fort Collins

Alpha-Spectroscopy Analysis Report

Sample: RAS170511-1MB
Spectrum #1 Analysis #1

Sample Size : 0.50

Detector: 81
Batch Name: RAS170511-1_A
Nuclide Library: Radium
Analysis Method: ROI Analysis, Set Name = Ra224/Ra226
ROI Set: Ra224/Ra226

Acquisition

Acquisition Start Date: 5/12/2017 1:44:56PM
Live Time: 1,000.00 min.
Real Time: 1,000.01 min.
Dead Time: 0.00 %

Calibration

Bkgd Info: Sample: B17051081; Det: 81; Spectrum #1; 5/10/2017 10:45:36 AM

Calibration Date: 5/10/2017 10:11:19AM

Efficiency Calibration: C17051081

Efficiency: 28.35% +/- 0.16% TPU(2 sigma)

Energy Calibration: C17051081

Energy Cal: Gain = 9.7851 keV / Ch

Offset = 3,042.97 keV

Quadratic = 0.0000 keV / Ch²


Tracer

Tracer Name: At-217

Tracer Activity: 20.00 DPM/mL x (Vol.)1.00 mL = 20.00 DPM

Tracer Nuclide: At-217

Tracer Recovery: 4.10%



Nuclide Summary (ROI)

| Nuclide | Peak Energy keV | ROI Start keV | ROI End keV | FWHM keV | B.R. % | Gross Counts | Bkgd Counts | Net Counts | Activity pCi/L | 1.00Sigma TPU pCi/L | Critical Level pCi/L | MDA pCi/L |
|---------|-----------------|---------------|-------------|----------|--------|--------------|-------------|------------|----------------|---------------------|----------------------|-----------|
| Ra-226 | 4784.7 | 4510.7 | 4833.6 | 13.5 | 100.1 | 3.00 | 5.00 | -2.00 | -1.6E-001 | 2.2E-001 | 4.0E-001 | 1.0E+000 |
| Ra-224 | 5675.2 | 5538.2 | 5733.9 | .0 | 95.1 | 74.00 | 28.00 | 46.00 | 3.8E+000 | 8.9E-001 | 1.0E+000 | 2.2E+000 |
| Ac-225 | 5812.2 | 5743.7 | 5900.2 | 54.3 | 78.3 | 193.00 | 19.00 | 174.00 | 1.7E+001 | 2.1E+000 | 1.0E+000 | 2.3E+000 |
| Fr-221 | 6311.2 | 6037.2 | 6369.9 | 43.9 | 100.0 | 275.00 | 37.00 | 238.00 | 1.2E+002 | 1.4E+001 | 7.2E+000 | 1.6E+001 |
| Po-216 | 6741.7 | 6614.5 | 6829.8 | 17.2 | 100.0 | 22.00 | 17.00 | 5.00 | 3.9E-001 | 4.9E-001 | 7.4E-001 | 1.7E+000 |
| At-217 | 7054.9 | 6839.6 | 7133.2 | 41.0 | 99.9 | 243.00 | 11.00 | 232.00 | 7.4E-001 | 5.1E-002 | 6.0E-001 | 1.4E+000 |

JA

JP

ALS Laboratory Group - Fort Collins

Alpha-Spectroscopy Analysis Report

Sample: RAS170511-1MB
Spectrum #1 Analysis #1

Sample Size : 0.50

Detector: 81
Batch Name: RAS170511-1_B
Nuclide Library: Radium
Analysis Method: ROI Analysis, Set Name = Ra224/Ra226
ROI Set: Ra224/Ra226

Acquisition

Acquisition Start Date: 5/23/2017 6:44:22AM
Live Time: 300.00 min.
Real Time: 300.00 min.
Dead Time: 0.00 %

Calibration

Bkgd Info: Sample: B17051081; Det: 81; Spectrum #1; 5/10/2017 10:45:36 AM

Calibration Date: 5/10/2017 10:11:19AM

Efficiency Calibration: C17051081

Efficiency: 28.35% +/- 0.16% TPU(2 sigma)

Energy Calibration: C17051081

Energy Cal: Gain = 9.7851 keV / Ch

Offset = 3,042.97 keV

Quadratic = 0.0000 keV / Ch²


Tracer

Tracer Name: At-217

Tracer Activity: 20.00 DPM/mL x (Vol.)1.00 mL = 20.00 DPM

Tracer Nuclide: At-217

Tracer Recovery: 53.60%



Nuclide Summary (ROI)

| Nuclide | Peak Energy keV | ROI Start keV | ROI End keV | FWHM keV | B.R. % | Gross Counts | Bkgd Counts | Net Counts | Activity pCi/L | 1.00Sigma TPU pCi/L | Critical Level pCi/L | MDA pCi/L |
|---------|-----------------|---------------|-------------|----------|--------|--------------|-------------|------------|----------------|---------------------|----------------------|-----------|
| At-217 | 7054.9 | 6839.6 | 7133.2 | 44.8 | 99.9 | 914.00 | 3.30 | 910.70 | 9.7E+000 | 3.2E-001 | 6.7E-002 | 1.9E-001 |
| Ra-226 | 4784.7 | 4510.7 | 4833.6 | 12.8 | 100.1 | 4.00 | 1.50 | 2.50 | 4.9E-002 | 4.6E-002 | 4.5E-002 | 1.4E-001 |
| Ra-224 | 5675.2 | 5538.2 | 5733.9 | 13.5 | 95.1 | 213.00 | 8.40 | 204.60 | 4.3E+000 | 4.0E-001 | 1.1E-001 | 2.8E-001 |
| Ac-225 | 5812.2 | 5743.7 | 5900.2 | 71.1 | 78.3 | 775.00 | 5.70 | 769.30 | 1.9E+001 | 1.4E+000 | 1.1E-001 | 2.9E-001 |
| Fr-221 | 6311.2 | 6037.2 | 6369.9 | 40.1 | 100.0 | 986.00 | 11.10 | 974.90 | 3.8E+001 | 2.7E+000 | 2.4E-001 | 5.9E-001 |
| Po-216 | 6741.7 | 6614.5 | 6829.8 | .6 | 100.0 | 17.00 | 5.10 | 11.90 | 2.4E-001 | 9.4E-002 | 8.4E-002 | 2.2E-001 |

JA

DP

ALS Laboratory Group - Fort Collins


Alpha-Spectroscopy Analysis Report

Sample: RAS170511-1LCS Sample Size : 0.50
Spectrum #1 Analysis #1

Detector: 82 Acquisition Start Date: 5/12/2017 1:44:56PM
Batch Name: RAS170511-1_A Live Time: 1,000.00 min.
Nuclide Library: Radium Real Time: 1,000.01 min.
Analysis Method: ROI Analysis, Set Name = Ra224/Ra226 Dead Time: 0.00 %
ROI Set: Ra224/Ra226

Bkgd Info: Sample: B17051082; Det: 82; Spectrum #1; 5/10/2017 10:45:36 AM
Calibration Date: 5/10/2017 10:11:28AM Energy Calibration: C17051082
Efficiency Calibration: C17051082 Energy Cal: Gain = 9.7851 keV / Ch
Efficiency: 28.92% +/- 0.20% TPU(2 sigma) Offset = 3,042.97 keV
Quadratic = 0.0000 keV / Ch²

Tracer Name: At-217 Tracer Nuclide: At-217
Tracer Activity: 20.00 DPM/mL x (Vol.)1.00 mL = 20.00 DPM Tracer Recovery: 3.67%



| Nuclide Summary (ROI) | | | | | | | | | | | | |
|-----------------------|-----------------|---------------|-------------|----------|--------|--------------|-------------|------------|----------------|---------------------|----------------------|-----------|
| Nuclide | Peak Energy keV | ROI Start keV | ROI End keV | FWHM keV | B.R. % | Gross Counts | Bkgd Counts | Net Counts | Activity pCi/L | 1.00Sigma TPU pCi/L | Critical Level pCi/L | MDA pCi/L |
| Fr-221 | 6311.2 | 6037.2 | 6369.9 | 72.4 | 100.0 | 2,573.00 | 43.00 | 2,530.00 | 1.4E+003 | 1.4E+002 | 8.5E+000 | 1.8E+001 |
| Po-216 | 6741.7 | 6614.5 | 6829.8 | 46.0 | 100.0 | 1,990.00 | 15.00 | 1,975.00 | 1.7E+002 | 1.5E+001 | 7.7E-001 | 1.8E+000 |
| At-217 | 7054.9 | 6839.6 | 7133.2 | 44.7 | 99.9 | 226.00 | 14.00 | 212.00 | 6.6E-001 | 4.8E-002 | 7.4E-001 | 1.7E+000 |
| Ra-226 | 4784.7 | 4510.7 | 4833.6 | 25.5 | 100.1 | 2,574.00 | 9.00 | 2,565.00 | 2.2E+002 | 2.0E+001 | 5.9E-001 | 1.4E+000 |
| Ra-224 | 5675.2 | 5538.2 | 5733.9 | 45.5 | 95.1 | 2,403.00 | 21.00 | 2,382.00 | 2.1E+002 | 1.9E+001 | 9.5E-001 | 2.1E+000 |
| Ac-225 | 5812.2 | 5743.7 | 5900.2 | 77.9 | 78.3 | 199.00 | 15.00 | 184.00 | 2.0E+001 | 2.4E+000 | 9.8E-001 | 2.2E+000 |

JA

[Signature]


ALS Laboratory Group - Fort Collins

Alpha-Spectroscopy Analysis Report

| | |
|---|--|
| Sample: RAS170511-1LCS Spectrum #1 Analysis #1 | Sample Size : 0.50 |
| Detector: 82 Batch Name: RAS170511-1_B Nuclide Library: Radium Analysis Method: ROI Analysis, Set Name = Ra224/Ra226 ROI Set: Ra224/Ra226 | Acquisition Start Date: 5/23/2017 6:44:22AM Live Time: 300.00 min. Real Time: 300.00 min. Dead Time: 0.00 % |

| | |
|---|---|
| Bkgd Info: Sample: B17051082; Det: 82; Spectrum #1; 5/10/2017 10:45:36 AM Calibration Date: 5/10/2017 10:11:28AM Efficiency Calibration: C17051082 Efficiency: 28.92% +/- 0.20% TPU(2 sigma) | Calibration Energy Calibration: C17051082 Energy Cal: Gain = 9.7851 keV / Ch Offset = 3,042.97 keV Quadratic = 0.0000 keV / Ch ² |
|---|---|

| | |
|--|---|
| Tracer Name: At-217 Tracer Activity: 20.00 DPM/mL x (Vol.)1.00 mL = 20.00 DPM | Tracer Tracer Nuclide: At-217 Tracer Recovery: 53.07% |
|--|---|



| Nuclide Summary (ROI) | | | | | | | | | | | | |
|-----------------------|-----------------|---------------|-------------|----------|--------|--------------|-------------|------------|----------------|---------------------|----------------------|-----------|
| Nuclide | Peak Energy keV | ROI Start keV | ROI End keV | FWHM keV | B.R. % | Gross Counts | Bkgd Counts | Net Counts | Activity pCi/L | 1.00Sigma TPU pCi/L | Critical Level pCi/L | MDA pCi/L |
| Ra-226 | 4784.7 | 4510.7 | 4833.6 | 26.3 | 100.1 | 755.00 | 2.70 | 752.30 | 1.5E+001 | 1.0E+000 | 6.0E-002 | 1.7E-001 |
| Ra-224 | 5675.2 | 5538.2 | 5733.9 | 130.4 | 95.1 | 318.00 | 6.30 | 311.70 | 6.4E+000 | 5.4E-001 | 9.7E-002 | 2.5E-001 |
| Ac-225 | 5812.2 | 5743.7 | 5900.2 | 71.2 | 78.3 | 827.00 | 4.50 | 822.50 | 2.1E+001 | 1.4E+000 | 9.9E-002 | 2.7E-001 |
| Fr-221 | 6311.2 | 6037.2 | 6369.9 | 46.1 | 100.0 | 1,138.00 | 12.90 | 1,125.10 | 4.3E+001 | 3.0E+000 | 2.6E-001 | 6.2E-001 |
| Po-216 | 6741.7 | 6614.5 | 6829.8 | 39.8 | 100.0 | 93.00 | 4.50 | 88.50 | 1.7E+000 | 2.2E-001 | 7.8E-002 | 2.1E-001 |
| At-217 | 7054.9 | 6839.6 | 7133.2 | 46.4 | 99.9 | 924.00 | 4.20 | 919.80 | 9.6E+000 | 3.2E-001 | 7.5E-002 | 2.0E-001 |

JA

JP

ALS

Alpha Spectrometer Instrument Run Log

Date: 5/11/17 5/12/17

| Detector | Batch ID | Sample ID | Iso/Matrix | Duration | Initial |
|----------|--------------|----------------|------------|----------|---------|
| 105 | PAS170506-1A | 1704493-12 | Pu/F | 1000 | JA |
| 106 | | -13 | | | |
| 107 | | -14 | | | |
| 108 | | -15 | | | |
| 109 | | -16 | | | |
| 110 | | -17 | | | |
| 111 | | -18 | | | |
| 112 | | -19 | | | |
| 113 | | -20 | | | |
| 114 | | AS170506-1MB | | | |
| 115 | | PMB | | | |
| 116 | | LC5 | | | |
| 128 | UAS170506-4A | 1704504-1 | U/F | 420 | JA |
| 129 | | -2 | | | |
| 130 | | -3 | | | |
| 131 | | 1705057-1 | | | |
| 132 | | -2 | | | |
| 165 | PAS170506-1B | AS170506-1LC50 | Pu/F | 1000 | JA |
| 117 | UAS170508-2B | 1704505-19 | U/F | 360 | JA |
| 118 | | -20 | | | |
| 119 | | -200 | | | |
| 120 | | -21 | | | |
| 121 | | -22 | | | |
| 122 | | -23 | | | |

5/11/17

5/12/17

| Detector | Batch ID | Sample ID | Iso/Matrix | Duration | Initial |
|----------|--------------|---------------|------------|----------|---------|
| 123 | UAS170508-2B | 1704505-24 | U/F | 360 | JA |
| 124 | | 1704605-3 | | | |
| 125 | | -4 | | | |
| 126 | | AS170508-2MB | | | |
| 127 | | -LC5 | | | |
| 81 | PAS170511-1A | PAS170511-1MB | Pu/F | 1000 | JA |
| 82 | | -105 | | | |
| 83 | | 1705158-1 | | | |
| 84 | | 1705177-1 | | | |
| 85 | | -2 | | | |
| 86 | | -3 | | | |
| 87 | | 1705181-1 | | | |
| 88 | | 1705202-1 | | | |
| 89 | | -10 | | | |
| 90 | | 1705203-1 | | | |
| 91 | | 1705214-1 | | | |
| 92 | | -2 | | | |
| 93 | | -20 | | | |
| 94 | | -4 | | | |
| 95 | | 1705237-1 | | | |
| 115 | UAS170506-2A | 1704493-21 | U/F | 1000 | JA |
| 116 | | -22 | AM/F | | |
| | | JA | 5/10/17 | | |
| | | | | | |
| | | | | | |
| | | | | | |

JMA 5/12/17

Notes:

Reviewed by: JMA

Date: 5/12/17

471137

ALS

Alpha Spectrometer Instrument Run Log

5/22/17
JMT5/24/17
Date: 5/22/17 5/23/17

| Detector | Batch ID | Sample ID | Iso/Matrix | Duration | Initial |
|----------|---------------|---------------|------------|----------|---------|
| 81 | RAS170519-1-A | RAS170519-1-A | Ra/W | 1000 | JA |
| 82 | | ↓ -110S | | | |
| 83 | | 1705415-3 | | | |
| 84 | | ↓ -4 | | | |
| 85 | | ↓ -5 | | | |
| 86 | | ↓ -6 | | | |
| 87 | | ↓ -7 | | | |
| 88 | | ↓ -70 | | | ↓ |
| 81 | RAS170511-1-B | RAS170511-1-B | Ra/W | 300 | JA |
| 82 | | ↓ -110S | | | |
| 83 | | 1705158-1 | | | |
| 84 | | 1705177-1 | | | |
| 85 | | ↓ -2 | | | |
| 86 | | ↓ -3 | | | |
| 87 | | 1705181-1 | | | |
| 88 | | 1705202-1 | | | |
| 89 | | ↓ -10 | | | |
| 90 | | 1705203-1 | | | |
| 91 | | 1705214-1 | | | |
| 92 | | ↓ -2 | | | |
| 93 | | ↓ -20 | | | |
| 94 | | ↓ -4 | | | |
| 95 | | 1705237-1 | | | ↓ |

5/22/17

5/23/17

JMT 5/23/17

| Detector | Batch ID | Sample ID | Iso/Matrix | Duration | Initial |
|----------|---------------|--------------|------------|----------|---------|
| 81 | UAS170522-1-A | 1705137 | Ur/W | 360 | JA |
| 82 | | ↓ -2 | | | |
| 83 | | ↓ -3 | | | |
| 84 | | ↓ -4 | | | |
| 85 | | ↓ -40 | | | |
| 86 | | ↓ -5 | | | |
| 87 | | ↓ -6 | | | |
| 88 | | ↓ -7 | | | |
| 89 | | ↓ -8 | | | |
| 90 | | ↓ -11 | | | |
| 91 | | 1705261-1 | | | |
| 92 | | ↓ -2 | | | |
| 93 | | ↓ -3 | | | |
| 94 | | ↓ -4 | | | |
| 95 | | ↓ -5 | | | ↓ |
| 117 | UAS170522-1-B | 1705261-1 | Ur/W | 360 | JA |
| 118 | | ↓ -7 | | | |
| 119 | | ↓ -8 | | | |
| 120 | | ↓ -11 | | | |
| 121 | | 1705389-1 | | | |
| 122 | | AS170522-1MB | | | |
| 123 | | ↓ -110S | | | ↓ |

JMT 5/24/17

Notes:

Reviewed by: JMT
Date: 5/24/17

471148

Section 6

QUALITY ASSURANCE SUMMARY REPORTS

6

QUALITY ASSURANCE SUMMARY SHEET

ALS W.O. #/ BATCH 1705181
1705158 1705203
1705177 1705214
1705202 1705237 / RAS 1705111TEST Ra224/226METHOD 2 SPECSOP/REV (PREP) 701SOP/REV (ANAL) 714

Briefly document any QA or other problems or deviations associated with the analysis of samples. Problems could result from: log-in, color, odor, dilution, consistency, scheduling, equipment, or instrumentation, or may include documentation of minor deviations necessary due to unique DQO's or sample characteristics.

The samples were analyzed following SOP 701r1 with the following modifications:

1. The manganese dioxide pre-concentration step was performed by direct precipitation of manganese dioxide in the sample aliquant rather than passing the sample aliquant through manganese dioxide resin.
2. The sample source was counted twice in order to allow a more accurate measurement of Ra-224. The 1st count was made as soon as possible to minimize Ra-224 decay and also to minimize interference from Ac-225 (progeny of the Ra-225 tracer).
3. Net Ra-224 counts were determined by subtraction of the instrument background counts and Ac-225 counts estimated to be in the Ra-224 ROI from the Ra-224 gross counts. The Ac-225 contribution to the Ra-224 ROI was calculated by multiplying the Ac-225 counts found during the 1st count by the ratio of Ra-224 net counts to Ac-225 counts found in the 2nd count. This approach gives a sample specific correction for Ac-225 counts in the Ra-224 ROI. The magnitude of the correction is small due to making the 1st count quickly to minimize Ac-225 ingrowth

Samples 1705158-1, 1705202-1, and 1705203-1 contained significant concentrations of Ba. Aliquant volumes were chosen to provide an optimal amount of Ba for making the alpha counting source. Counting sources are made by making Ra/BaSO₄ micro-precipitates. If the micro-precipitate mass is too large (containing more than ~100µg of Ba), alpha particles are attenuated resulting in degraded spectral quality. The Ba concentrations for these samples are given in the raw data section of this report. The aliquant volumes used for these samples (6 to 15 ml) were much smaller than the typically used volume of 500ml. The small sample volumes resulted in much higher ssMDCs for these samples.

The duplicate error ratio (DER) for sample 1705214-2 and -2Dup (DER=2.6) was higher than acceptable. The spectral quality for the sample aliquant was not as good as the spectral quality for the duplicate. It appears that scattered alpha counts raised the background in the Ra-226 ROI for the sample causing an overestimation of Ra-226 activity. Ra-226 activity in both aliquants was found to be less than 1pCi/L.

TECHNICIAN/ANALYST

DATE

DEPARTMENT MANAGER

DATE

| Sample Preparation | RAS170511-1 | Ra-224 | Ra-224 | Ra-224 | Ra-224 | Ra-224 | Ra-224 | RAS170511-1 |
|--|------------------|------------------|-----------------|----------------|-----------------|-----------------|--------|-------------|
| Analyte | RAS170511-1MB | RAS170511-ILCS | 1705168-1 | 1705177-1 | 1705177-2 | 1705177-3 | | |
| Sampling Date: | 5/12/17 11:25 AM | 5/12/17 11:25 AM | 5/8/17 10:46 AM | 5/8/17 9:45 AM | 5/8/17 10:42 AM | 5/8/17 10:47 AM | | |
| Sample Size (L): | 0.5 | 0.5 | 0.01 | 0.5 | 0.5 | 0.5 | | |
| Spike Added (mL): | 0 | 1 | 0 | 0 | 0 | 0 | | |
| Tracer Added (mL): | 0.4 | 0.4 | 0.4 | 0.4 | 0.4 | 0.4 | | |
| Separation Date: | 5/12/17 | 5/12/17 | 5/12/17 | 5/12/17 | 5/12/17 | 5/12/17 | | |
| Separation Time: | 11:25 AM | 11:25 AM | 11:25 AM | 11:25 AM | 11:25 AM | 11:25 AM | | |
| Sample Analysis | | | | | | | | |
| 1st count Detector ID: | 81 | 82 | 83 | 84 | 85 | 86 | | |
| 2nd count Detector ID: | 81 | 82 | 83 | 84 | 85 | 86 | | |
| 1st Count Start Date: | 5/12/17 | 5/12/17 | 5/12/17 | 5/12/17 | 5/12/17 | 5/12/17 | | |
| 1st Count Start Time: | 1:45 PM | 1:45 PM | 1:45 PM | 1:45 PM | 1:45 PM | 1:45 PM | | |
| 1st Count Duration(min): | 1000 | 1000 | 1000 | 1000 | 1000 | 1000 | | |
| 2nd Count Start Date: | 5/23/17 | 5/23/17 | 5/23/17 | 5/23/17 | 5/23/17 | 5/23/17 | | |
| 2nd Count Start Time: | 6:44 AM | 6:44 AM | 6:44 AM | 6:44 AM | 6:44 AM | 6:44 AM | | |
| 2nd Count Duration(min): | 300 | 300 | 300 | 300 | 300 | 300 | | |
| 1st count Detector Efficiency: | 0.2835 | 0.2892 | 0.3046 | 0.3044 | 0.2983 | 0.2867 | | |
| 2nd count Detector Efficiency: | 0.2835 | 0.2892 | 0.3046 | 0.3044 | 0.2983 | 0.2867 | | |
| Bkg. Cal. Ct. Dur. (min): | 1000 | 1000 | 1000 | 1000 | 1000 | 1000 | | |
| 1st count Ra-224 Gross Cts: | 74 | 2403 | 89 | 3974 | 4697 | 4658 | | |
| 1st count Ac-225 Bkg Cts (per sample ct duration): | 28 | 21 | 28 | 34 | 31 | 16 | | |
| 1st count Ac-225 Gross Cts: | 193 | 199 | 213 | 1301 | 1454 | 758 | | |
| 2nd count At-217 Gross Cts: | 914 | 924 | 969 | 759 | 891 | 793 | | |
| 2nd count At-217 Bkg Cts (per sample ct duration): | 3.3 | 4.2 | 3.3 | 6.6 | 3.3 | 6.3 | | |
| 2nd count Ra-224 net counts | 205 | 312 | 241 | 732 | 826 | 517 | | |
| 2nd count Ac-225 net counts | 769 | 823 | 811 | 935 | 1202 | 844 | | |
| Calculations | | | | | | | | |
| Tracer Added (dpm@sep date): | 29.8 | 29.8 | 29.8 | 29.8 | 29.8 | 29.8 | | |
| 1st count Elapsed Time, sep. to ct.midpt., (hrs): | 10.67 | 10.67 | 10.67 | 10.67 | 10.67 | 10.67 | | |
| 2nd count Elapsed Time, sep. to ct.midpt., (hrs): | 261.82 | 261.82 | 261.82 | 261.82 | 261.82 | 261.82 | | |
| 1st count At-217 Ingrowth factor: | 0.030022 | 0.030022 | 0.030022 | 0.030022 | 0.030022 | 0.030022 | | |
| 2nd count At-217 Ingrowth factor: | 0.402307 | 0.402307 | 0.402307 | 0.402307 | 0.402307 | 0.402307 | | |
| Expected At-217, at mid pt of 1st count (dpm): | 0.8940 | 0.8940 | 0.8940 | 0.8940 | 0.8940 | 0.8940 | | |
| Expected At-217, at mid pt of 2nd count (dpm): | 11.9795 | 11.9795 | 11.9795 | 11.9795 | 11.9795 | 11.9795 | | |
| Chemical Yield (2nd count): | 0.89384 | 0.88498 | 0.88217 | 0.68777 | 0.82804 | 0.76352 | | |
| Ra-224 decay factor, sampling to 1st ct.midpt.: | 0.9187 | 0.9187 | 0.4260 | 0.4225 | 0.4258 | 0.4260 | | |
| Standard Denominator: | 0.2455 | 0.2479 | 0.0024 | 0.0933 | 0.1109 | 0.0983 | | |
| Sample Activity Conc. (pCi/L): | -0.00157 | 9.32591 | -2.46349 | 31.51903 | 33.17127 | 42.60556 | | |
| Std. Uncert., Ra-224 (pCi/L): | 0.0415 | 0.1987 | 4.6866 | 0.6788 | 0.6202 | 0.6953 | | |
| Std. Uncert., At-217 (rel): | 0.0332 | 0.0331 | 0.0323 | 0.0367 | 0.0337 | 0.0359 | | |
| Std. Uncert., Ra-225-229 act. (rel): | 0.0561 | 0.0561 | 0.0561 | 0.0561 | 0.0561 | 0.0561 | | |
| Std. Uncert., Sample Volume (rel): | 0.0060 | 0.0060 | 0.0060 | 0.0060 | 0.0060 | 0.0060 | | |
| Combined Std Uncertainty (pCi/L): | 0.04154 | 0.64184 | 4.68935 | 2.22829 | 2.26705 | 2.93344 | | |
| Critical Level (pCi/L): | 0.08758 | 0.09517 | 9.98383 | 0.81769 | 0.68284 | 0.52654 | | |
| MDC (pCi/L): | 0.18541 | 0.20049 | 21.01032 | 1.86068 | 1.38699 | 1.07799 | | |
| LCS Recovery: | | 103.5% | | | | | | |
| | #DIV/0! | | #DIV/0! | #DIV/0! | #DIV/0! | | | |

Section 7

LABORATORY BENCH SHEETS



Radiochemistry Instrument Worksheet

ALS -- Fort Collins

Prep Batch: RAS170511-1

Prep Procedure: RalSO

Analytical QASS / NCR? Y / N

472112

| Prep Num | LabID | QC Type | Init Aliq | Fin Aliq | Units | Report Units | Cnt 1 File | Cnt 1 Inst/Det | Cnt 1 Pos Chk By | Cnt 2 File | Cnt 2 Inst/Det | Cnt 2 Pos Chk By | Cnt 3 File | Cnt 3 Inst/Det | Cnt 3 Pos Chk By | Notes |
|----------|-------------|---------|-----------|----------|-------|--------------|------------|----------------|------------------|------------|----------------|------------------|------------|----------------|------------------|-------|
| 1 | 1705158-1 | SMP | 10 | 10 | ml | pCi/l | _51581 | 83 JPEJA | 83 JPEJA | _51581 | 83 JPEJA | 83 JPEJA | _51581 | | | AB |
| 1 | 1705177-1 | SMP | 500 | 500 | ml | pCi/l | _51771 | 84 | 84 | _51771 | 84 | 84 | _51771 | | | |
| 1 | 1705177-2 | SMP | 500 | 500 | ml | pCi/l | _51772 | 85 | 85 | _51772 | 85 | 85 | _51772 | | | |
| 1 | 1705177-3 | SMP | 500 | 500 | ml | pCi/l | _51773 | 86 | 86 | _51773 | 86 | 86 | _51773 | | | |
| 1 | 1705181-1 | SMP | 500 | 500 | ml | pCi/l | _51811 | 87 | 87 | _51811 | 87 | 87 | _51811 | | | |
| 1 | 1705202-1 | SMP | 6 | 6 | ml | pCi/l | _52021 | 88 | 88 | _52021 | 88 | 88 | _52021 | | | |
| 1 | 1705202-1 | DUP | 6 | 6 | ml | pCi/l | _52021D | 89 | 89 | _52021D | 89 | 89 | _52021D | | | |
| 1 | 1705203-1 | SMP | 15 | 15 | ml | pCi/l | _52031 | 90 | 90 | _52031 | 90 | 90 | _52031 | | | |
| 1 | 1705214-1 | SMP | 500 | 500 | ml | pCi/l | _52141 | 91 | 91 | _52141 | 91 | 91 | _52141 | | | |
| 1 | 1705214-2 | SMP | 300 | 300 | ml | pCi/l | _52142 | 92 | 92 | _52142 | 92 | 92 | _52142 | | | |
| 1 | 1705214-2 | DUP | 300 | 300 | ml | pCi/l | _52142D | 93 | 93 | _52142D | 93 | 93 | _52142D | | | |
| 1 | 1705214-4 | SMP | 500 | 500 | ml | pCi/l | _52144 | 94 | 94 | _52144 | 94 | 94 | _52144 | | | |
| 1 | 1705237-1 | SMP | 500 | 500 | ml | pCi/l | _52371 | 95 | 95 | _52371 | 95 | 95 | _52371 | | | |
| 1 | RAS170511-1 | MB | 500 | 500 | ml | pCi/l | _S17051B | 81 | 81 | _S17051B | 81 | 81 | _S17051B | | | |
| 1 | RAS170511-1 | LCS | 500 | 500 | ml | pCi/l | _S17051L | 82 | 82 | _S17051L | 82 | 82 | _S17051L | | | |

Handwritten signature/initials

Spike Solution Information

| Soln # | Nuclide | SolnID | Exp Date | Prep Conc | Units | Prep Date | Aliquot | Units | Pipet ID |
|--------|---------|--------------|----------|-----------|--------|-----------|---------|-------|----------|
| S1 | Ra-224 | 1026.4095.84 | | 9.930 | DPW/ml | 05/11/17 | 1 | ml | RS-005 |
| S1 | Ra-226 | 1026.4095.84 | | 10.219 | DPW/ml | 05/11/17 | 1 | ml | RS-005 |
| S1 | Rn-222 | 1026.4095.84 | | 10.219 | DPW/ml | 05/11/17 | 1 | ml | RS-005 |
| S1 | Th-232 | 1026.4095.84 | | 9.930 | DPW/ml | 05/11/17 | 1 | ml | RS-005 |

Sample Barcodes

| | | | | | |
|--------------------------------|--|---------------------------------|--|------------------------------|--|
| 1705158-1 RAS170511-1PS1 | | 1705177-1 RAS170511-1PS2 | | 1705177-2 RAS170511-1PS3 | |
| 1705177-3 RAS170511-1PS4 | | 1705181-1 RAS170511-1PS5 | | 1705202-1 RAS170511-1PS7 | |
| 1705202-1DUP RAS170511-1PS6 | | 1705203-1 RAS170511-1PS8 | | 1705214-1 RAS170511-1PS9 | |
| 1705214-2 RAS170511-1PS10 | | 1705214-2DUP RAS170511-1PS11 | | 1705214-4 RAS170511-1PS12 | |

Handwritten signature/initials

Radiochemistry Instrument Worksheet

ALS -- Fort Collins

Prep Batch: RAS170511-1

Prep Procedure: RaISO

Analytical QASS / NCR? Y / N 472112

| | | | | | | | | | | | | | | | | |
|-----------|-----------------|---------|-----------|----------|-------|--------------|------------|---------------|------------------|------------|---------------|------------------|------------|---------------|------------------|-------|
| Prep Num | LabID | QC Type | Init Aliq | Fin Aliq | Units | Report Units | Cnt 1 File | Cnt 1 Ins/Det | Cnt 1 Pos Chk By | Cnt 2 File | Cnt 2 Ins/Det | Cnt 2 Pos Chk By | Cnt 3 File | Cnt 3 Ins/Det | Cnt 3 Pos Chk By | Notes |
| 1705237-1 | RAS170511-1PS13 | | | | | | | | | | | | | | | |

Reporting Units

| LabID: | TstGrpName: | RptUnits: |
|-----------|-----------------|-----------|
| 1705237-1 | Ra226_224_USGS | pCi/l |
| 1705214-1 | Ra226_224_USGS | pCi/l |
| 1705203-1 | Ra226_224_COGCC | pCi/l |
| 1705202-1 | Ra226_224_COGCC | pCi/l |
| 1705181-1 | Ra226_224_USGS | pCi/l |
| 1705177-1 | Ra226_224_COGCC | pCi/l |
| 1705158-1 | Ra226_224_COGCC | pCi/l |
| 1705214-2 | Ra226_224_USGS | pCi/l |
| 1705177-2 | Ra226_224_COGCC | pCi/l |
| 1705177-3 | Ra226_224_COGCC | pCi/l |
| 1705214-4 | Ra226_224_USGS | pCi/l |

Radiochemistry Prep Worksheet

ALS -- Fort Collins

Prep Batch: RAS170511-1

Prep Procedure: RALSO

Reviewed By: sdw

Review Date: 5/11/2017

Non-Routine Pre-Treatment? Y / N

Batch: NA

Re-Prep? Y / N

Batch: NA

Prep QASS / NCR? Y / N

Signature

Prep SOP: PAI 701 Rev: 1

Prep SOP: NONE

Matrix Class: liquid

Prep Analyst: Steven D. White

Prep Date: 5/11/2017

Prep Dept: RS

Balance: na

Balance: na

| Samp Num | Prep Num | LabID | QC Type | Dish No. | Init Aliq ml | Fin Aliq ml | Prep Basis | Ingrowth Date/Time | Micro Init | Micro Date | Standards | Prep Notes |
|----------|----------|-------------|---------|----------|--------------|-------------|-------------|--------------------|------------|------------|-----------|------------|
| 1 | 1 | 1705158-1 | SMP | | 10 | 10 | As Received | | SW | 5/12/17 | | |
| 2 | 1 | 1705177-1 | SMP | | 500 | 500 | Filtered | | | | | |
| 3 | 1 | 1705177-2 | SMP | | 500 | 500 | Filtered | | | | | |
| 4 | 1 | 1705177-3 | SMP | | 500 | 500 | Filtered | | | | | |
| 5 | 1 | 1705181-1 | SMP | | 500 | 500 | As Received | | | | | |
| 6 | 1 | 1705202-1 | SMP | | 6 | 6 | Filtered | | | | | |
| 7 | 1 | 1705202-1 | DUP | | 6 | 6 | Filtered | | | | | |
| 8 | 1 | 1705203-1 | SMP | | 15 | 15 | Filtered | | | | | |
| 9 | 1 | 1705214-1 | SMP | | 500 | 500 | As Received | | | | | |
| 10 | 1 | 1705214-2 | SMP | | 300 | 300 | As Received | | | | | |
| 11 | 1 | 1705214-2 | DUP | | 300 | 300 | As Received | | | | | |
| 12 | 1 | 1705214-4 | SMP | | 500 | 500 | As Received | | | | | |
| 13 | 1 | 1705237-1 | SMP | | 500 | 500 | As Received | | | | | |
| 14 | 1 | RAS170511-1 | MB | | 500 | 500 | As Received | | | | | |
| 15 | 1 | RAS170511-1 | LCS | | 500 | 500 | As Received | | | | | |

Signature

S1

Comments

Spiked By: Steven D. White Date: 5/11/2017

Witnessed By: Steven D. White Date: 5/11/2017

| Spike Solution Information | | | | | | | | | |
|----------------------------|---------|--------------|----------|-----------|--------|-----------|---------|-------|----------|
| Soln # | Nuclide | SolnID | Exp Date | Prep Conc | Units | Prep Date | Aliquot | Units | Pipet ID |
| S1 | Ra-224 | 1026.4095.84 | | 9.930 | DPM/ml | 05/11/17 | 1 | ml | RS-005 |
| S1 | Ra-226 | 1026.4095.84 | | 10.219 | DPM/ml | 05/11/17 | 1 | ml | RS-005 |
| S1 | Rn-222 | 1026.4095.84 | | 10.219 | DPM/ml | 05/11/17 | 1 | ml | RS-005 |
| S1 | Th-232 | 1026.4095.84 | | 9.930 | DPM/ml | 05/11/17 | 1 | ml | RS-005 |

Signature

Ra-224/Ra-226 by Alpha Spectrometry (Aqueous Matrix)

Prep Batch ID: RA5170511-1

Analyst: Steen D. White

Diphonix Resin lot #: 033111DPM

Ra-225 Tracer: C 4-14-17

Ra-224 QC Spike: SW 5-27-2016

Ra-226 QC Spike: 916,4095.68

For LIMS purposes ONLY: QC Spike: 1026.4095.84 (1.0ml)
1000ppm Ba ID:

1st Count Date: 5-12-17
Duration (minutes): 1,000

2nd Count Date: 5-23-17
Duration (minutes): 300

Diphonix Column Separation & Collection Date/Time: 5-12-17 @ 11:25

Refrigeration or Sonication Date/Time Start: 1 @ 12:07
End: 1 @ 12:07

| Collection Date | Time | pH | Filtered By Client/ALS | ALS Sample ID | Cup ID | Volume (ml) | Tracer Vol (ml) | QC Spike Vol (ml) | 1000ppm Ba (ml) | 1st Count Detector | 2nd Count Detector | Notes/Comments: |
|-----------------|------|----|------------------------|---------------|--------|-------------|-----------------|-------------------|-----------------|--------------------|--------------------|-----------------|
| na | na | na | na | MB | 1 | 500 | 0.140 | --na-- | 0.09 | 81 JA | 81 JA | |
| na | na | na | na | LCS | 2 | 500 | | 0.10 | 1 | 82 | 82 | |
| 5-8-17 | 1046 | 22 | Client | 1705158-1 | 3 | 10 | | --na-- | 0 | 83 | 83 | |
| 5-8-17 | 0945 | | ALS | 1705177-1 | 4 | 500 | | --na-- | 0.09 | 84 | 84 | |
| | 1042 | | | -2 | 5 | | | --na-- | 1 | 85 | 85 | |
| | 1047 | | | -3 | 6 | | | --na-- | 0.06 | 86 | 86 | |
| 5-8-17 | 1315 | | Client | 1705181-1 | 7 | | | --na-- | 0.09 | 87 | 87 | |
| 5-9-17 | 1250 | | ALS | 1705202-1 | 8 | 6 | | --na-- | 0 | 88 | 88 | |
| | 1 | | | -1 Dup | 9 | | | --na-- | 1 | 89 | 89 | |
| 5-9-17 | 0930 | | | 1705203-1 | 10 | 15 | | --na-- | 1 | 90 | 90 | |
| 5-9-17 | 1100 | | Client | 1705214-1 | 11 | 500 | | --na-- | 0.08 | 91 | 91 | |
| | 1415 | | | -2 | 12 | 300 | | --na-- | 0 | 92 | 92 | |
| | 1 | | | -2 Dup | 13 | | | --na-- | 1 | 93 | 93 | |
| | 1100 | | | -4 | 14 | 500 | | --na-- | 0.05 | 94 | 94 | |
| 5-10-17 | 1100 | | | 1705237-1 | 15 | | | --na-- | 0.09 | 95 | 95 | |
| | | | | | 16 | | | --na-- | | | | |
| | | | | | 17 | | | --na-- | | | | |
| | | | | | 18 | | | --na-- | | | | |
| | | | | | 19 | | | --na-- | | | | |
| | | | | | 20 | | | --na-- | | | | |
| | | | | | 21 | | | --na-- | | | | |
| | | | | | 22 | | | --na-- | | | | |
| | | | | | 23 | | | --na-- | | | | |

filtered due to oil layer
filtered due to oil layer

Screens

| Sample Id1 | Ca | Fe | K | Mg | Na | Sr | Mn | S | Al | Ba | Pb | Ni |
|------------|---------|---------|---------|---------|---------|---------|---------|--------|---------|--------|---------|---------|
| CCV | 52.7205 | 20.6546 | 50.1670 | 51.7807 | 50.6940 | 0.5043 | 1.0243 | 5.1356 | 50.7704 | 1.0175 | 1.0391 | 1.0504 |
| CCB | 0.0423 | -0.0064 | 0.1311 | 0.0405 | 0.1289 | -0.0007 | -0.0001 | 0.0082 | 0.0158 | 0.0018 | 0.0123 | -0.0003 |
| 1705181-1 | 11.5155 | 0.4934 | 0.6060 | 4.6103 | 6.3420 | 0.0230 | 0.0052 | 1.7790 | -0.0665 | 0.0240 | -0.0007 | -0.0012 |
| 1705214-1 | 18.8821 | -0.0355 | 0.3036 | 10.0319 | 28.1694 | 0.0230 | -0.0005 | 4.9137 | -0.0551 | 0.0395 | -0.0006 | 0.0012 |
| 1705214-2 | 23.4482 | 0.4378 | 1.7482 | 9.8005 | 7.6707 | 0.1172 | 0.0768 | 1.7790 | -0.0709 | 0.3105 | -0.0017 | 0.0000 |
| 1705214-4 | 34.7681 | -0.0464 | 0.5358 | 6.2576 | 8.1792 | 0.0826 | -0.0012 | 2.3747 | -0.0652 | 0.0834 | 0.0008 | 0.0000 |
| 1705237-1 | 9.2568 | -0.0475 | 1.1080 | 2.0864 | 14.4965 | 0.0736 | 0.0067 | 2.8143 | -0.0519 | 0.0274 | -0.0019 | -0.0003 |
| CCV | 52.3190 | 20.5589 | 50.1920 | 51.5398 | 50.4979 | 0.5063 | 1.0193 | 5.1684 | 50.7141 | 1.0226 | 1.0265 | 1.0471 |
| CCB | 0.0513 | -0.0040 | 0.1235 | 0.0417 | 0.1406 | -0.0005 | 0.0003 | 0.0041 | 0.0234 | 0.0020 | 0.0155 | -0.0006 |

Sample vol (ml) ml 1000g in Ba

181-1 500 0.09

214-1 500 0.08

-2 300 0

-4 500 0.05

237-1 500 0.09

1st count 5/12 1000min (due to reduced vol.)

2nd count ~ 5/23 300min

tracer: 0.4ml

Screens

| Sample Id1 | Ca | Fe | K | Mg | Na | Sr | Mn | S | Al | Ba | Pb | Ni |
|----------------|----------|---------|---------|---------|----------|---------|---------|---------|---------|---------|---------|---------|
| CCV | 51.7633 | 20.3585 | 49.9543 | 51.5826 | 50.2073 | 0.5023 | 1.0101 | 5.0575 | 50.7350 | 1.0156 | 1.0270 | 1.0174 |
| CCB | -0.0253 | -0.0389 | 0.0368 | -0.0236 | 0.0377 | -0.0012 | -0.0009 | -0.0041 | -0.0507 | -0.0008 | 0.0020 | 0.0006 |
| 100X 1705095-1 | 180.8864 | 0.5234 | 16.2659 | 22.8274 | 206.1871 | 5.9668 | 0.0216 | 1.2777 | -0.0715 | 0.0724 | -0.0018 | -0.0013 |
| 100X 1705158-1 | 0.0456 | -0.0521 | 0.0884 | -0.0501 | 8.5924 | 0.0119 | -0.0013 | 0.0452 | -0.0785 | 0.0926 | 0.0005 | -0.0010 |
| 100X 1705177-1 | 8.7786 | -0.0517 | 1.7423 | 2.8498 | 91.9193 | 0.2324 | -0.0013 | 35.2547 | -0.0829 | -0.0013 | 0.0017 | 0.0003 |
| 100X 1705177-2 | 8.8470 | -0.0437 | 1.7491 | 2.8515 | 90.9329 | 0.2315 | 0.0001 | 35.7025 | -0.0715 | -0.0012 | -0.0008 | 0.0014 |
| 100X 1705177-3 | 9.6090 | 0.0342 | 1.1591 | 1.4981 | 48.4507 | 0.2672 | 0.0003 | 6.0682 | -0.0766 | -0.0005 | -0.0015 | -0.0004 |
| 100X 1705202-1 | 1.0769 | 0.4599 | 0.3328 | 0.2168 | 50.1114 | 0.1023 | 0.0056 | 0.0575 | -0.0823 | 0.1577 | -0.0025 | 0.0024 |
| 100X 1705203-1 | 0.0436 | 0.1595 | 0.7088 | 0.0586 | 43.1160 | 0.0488 | 0.0022 | 0.0657 | -0.0811 | 0.0645 | -0.0027 | 0.0001 |
| CCV | 52.1421 | 20.3762 | 49.6325 | 51.6225 | 49.8078 | 0.5006 | 1.0114 | 5.0863 | 50.3481 | 1.0049 | 1.0345 | 1.0217 |
| CCB | -0.0297 | -0.0394 | 0.0694 | -0.0265 | 0.0445 | -0.0012 | -0.0008 | 0.0205 | -0.0513 | -0.0007 | -0.0011 | 0.0000 |

Sample vol. (ml) est Ba to add (ml 1000 ppm)

158-1 ~ 9.3 ug/ml Ba

10

0

177-1 ~ 900mg/L Ca low Ba

250 500

0.09

-2

-3

202-1 ~ 16 ug/ml Ba

6

0

203-1 ~ 6.5 ug/ml Ba

15

0

| Sample Id1 | Ca | Fe | K | Mg | Na | Sr | Mn | S | Al | Ba | Pb | Ni |
|------------|----------|----------|----------|----------|----------|---------|--------|-----------|---------|---------|---------|---------|
| CCV | 52.3190 | 20.5589 | 50.1920 | 51.5398 | 50.4979 | 0.5063 | 1.0193 | 5.1684 | 50.7141 | 1.0226 | 1.0265 | 1.0471 |
| CCB | 0.0513 | -0.0040 | 0.1235 | 0.0417 | 0.1406 | -0.0005 | 0.0003 | 0.0041 | 0.0234 | 0.0020 | 0.0155 | -0.0006 |
| 1705177-1 | 525.9361 | 0.0811 | 215.9052 | 183.3913 | 176.9344 | 13.7658 | 0.0219 | 1219.4988 | 0.2965 | 0.0099 | -0.0018 | 0.0282 |
| 1705177-2 | 508.7043 | 0.3012 | 218.3355 | 180.2003 | 172.8211 | 13.7265 | 0.1088 | 1179.5974 | 0.8728 | 0.0117 | -0.0009 | 0.0624 |
| 1705177-3 | 625.8498 | 4.3457 | 178.0348 | 113.7966 | 214.7379 | 16.9502 | 0.1122 | 578.2457 | -0.0437 | 0.0859 | 0.0192 | 0.0196 |
| CCV | 52.2157 | 20.6384 | 51.3507 | 51.9980 | 39.1053 | 0.5153 | 1.0251 | 40.4930 | 51.7387 | 1.0401 | 1.0384 | 1.0365 |
| CCB | 0.0916 | 0.0026 | 0.1252 | 0.0602 | 1.0741 | 0.0003 | 0.0003 | 21.5078 | 0.0285 | 0.0022 | 0.0127 | 0.0020 |
| 1705158-1 | 6.1817 | 0.0310 | 10.9358 | 1.3940 | 270.7578 | 0.9008 | 0.0281 | 6.7707 | 0.0070 | 8.4460 | -0.0011 | -0.0018 |
| 1705202-1 | 76.7521 | 41.7307 | 62.9640 | 19.0554 | 226.3697 | 7.4105 | 0.6951 | 9.4084 | 0.1024 | 12.3435 | 0.0009 | 0.3839 |
| 1705203-1 | 8.1529 | 160.7713 | 117.7244 | 7.8548 | 231.2136 | 3.6469 | 1.4557 | 1121.2366 | -0.0222 | 5.3439 | 0.0091 | 0.0191 |

Section 8

STANDARDS TRACEABILITY DOCUMENTS



Project 1026.4095.84 Ra224/226 working standard
Continued from Page _____

1E 3/12/15

10dpm/ml QC Spike Solution

2.0ml of 102.5dpm/ml Ra-226 (916.4095.68) and 0.082ml of 2442dpm/ml Ra-224 (ST100301-30) were diluted to 20ml final volume. The Ra-224 stock solution (ST100301-30) is a 10mg/ml Th-232 primary ICP standard in secular equilibrium with Th-228 (and also in secular equilibrium with Ra-224).

The QC spike solution contains:

10.25dpm/ml Ra-226 and 10.0dpm/ml Ra-224

1E 3/12/15

JP 6/23/15

Continued on Page _____

7 ECU

Signed

3/12/15

Date

Read and Understood By

[Signature]

Signed

03/25/15

Date

Project

Continued from Page

916.4095.68

Ra-226 working std

Prepare a working dilution of 916.3610.76

1. Density of 0.1M HCl, lot # 0000092116
 Mass of 100mL vol. flask: 68.3000g Balance # 12
 Mass of flask & 100mL acid: 168.0372g Balance# 12
 Net Mass: 99.7372g
 Density: 0.9974 g/mL
2. Mass of 916.3610.76 transferred:
 Mass of open empty nalgene: 74.5396g Balance# 12
 Mass of nalgene & standard: 77.9525g Balance# 12
 Net mass of standard transferred: 3.4129g Balance#
3. Dilute to final volume:
 Mass of nalgene, standard, & diluent: 1076.1g Balance# 26
 Mass of empty nalgene (from above): 74.5396g Balance# 12
 Net mass of new dilution: 1001.5604g Balance# NA

4. Final activity calculation:

$$30.156 \text{ dpm/g} (0.9974 \text{ g/mL}) \left(\frac{3.4129 \text{ g}}{1001.5604 \text{ g}} \right) = 102.49 \text{ dpm/mL}$$

Std ID: 916.4095.68

Description: Ra-226

Expiration: 1/6/2016

Activity: 102.49 dpm/mL

2s Uncertainty: 5.02 dpm/mL

Ref. Date: 7/1/2010

Ref Time: N/A

Prep Date: 12/8/2014 Prep by: TE

Matrix/Comp. 0.1M HCl

Half Life (y): 1.60E+03

Reverification Log

| Analysis Date | Initials | Expiration Date |
|---------------|----------|-----------------|
| 12/29/16 | JP | 12/29/2017 |
| | | |
| | | |
| | | |

Continued on Page

Signed

Date

Read and Understood By

Signed

Date

Prepare a intermediate dilution of 916.3610.76

0.1 M HCl dilutant lot # H45A12

Density of diluent

Mass of 100 ml Vol Flask

68.7982 g

12

Flask + Acid

168.0565 g

2

Net

99.7883 g

$$\rho = 0.9976 \text{ g/ml}$$

Mass of parent transference

Mass of Open Full Ampule + Receptor

38.2022 g

12

Mass of Open Empty Ampule + Receptor

33.2251 g

4

Net

4.9831 g

Dilute to Final Vol./Mass

Mass of Open Empty 40 ml Vol. / 12.1

7.397 g

12

Mass of Vial, Std. + diluent

57.7016 g

1

Net

36.5045 g

Activity, C.L.

100 9/27/10

1860 D. R₀ (10 dpm) 4.9831 g
1 R₀ 5.65176 g

36.5045 g

30.156 g

30.156 g

Jan 9/27/10


Continued on Page _____

Read and Understood By


Signed

9/27/10

Date


Signed

9/27/10

Date



Eckert & Ziegler
Analytics

REC
7/6/10
R50#
9/6

1380 Seaboard Industrial Blvd.
Atlanta, Georgia 30318
Tel 404-352-8677
Fax 404-352-2837
www.analyticinc.com

CERTIFICATE OF CALIBRATION Standard Radionuclide Source

82863-307

Ra-226 5 mL Liquid in Flame Sealed Vial

Customer: ALS Laboratory Group / Fort Collins
P.O. No.: 73828 08-10-10, Item 1

This standard radionuclide source was prepared gravimetrically from a master solution calibrated by Eckert & Ziegler Analytics, using a germanium gamma spectrometer system. Radionuclide purity and calibration were checked with a germanium gamma spectrometer system. The nuclear decay rate and reference date for this source are given below. Eckert & Ziegler Analytics (EZA) maintains traceability to the National Institute of Standards and Technology through a Measurements Assurance Program as described in USNRC Regulatory Guide 4.18, Revision 1, February, 1979, and compliance with ANSI N42.22-1998, "Traceability of Radioactive Sources to NIST." EZA is accredited by the Health Physics Society (HPS) for the production of NIST-traceable sources, and this source was produced in accordance with the HPS accreditation requirements. Customers may report any concerns with the accreditation program to the HPS Secretariat, 1313 Dolley Madison Blvd., Ste. 402, McLean, VA 22101.

| Isotope | Half-Life, Days | Activity (Bq) | Uncertainty*, % | | | Reference Date (12:00 PM EST) |
|---------|--------------------|------------------|-----------------|----------------|-----|----------------------------------|
| | | | u ₁ | u ₂ | U | |
| Ra-226 | 5.344E+08 | 1.860E+04 | 0.8 | 2.4 | 4.9 | 07/01/2010 |

*Uncertainty: U - Relative expanded uncertainty, k = 2. See NIST Technical Note 1287, "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Results."

Comments:

Impurities: γ -impurities <0.1%. 5.08176 g 0.1M HCl solution with approximately 30 μ g/g Ba carrier.

Source Prepared by:

W. Mao, Radiochemist

QA Approved:

J. D. McCorvey, QA Manager Alternate

Date:

6/30/10

ATA Form 005 Rev. 10

Single Isotope Certificate, Rev 1 9/28/2008

Corporate Office

24937 Avenue Tibbitts Valencia, California 91355



Laboratory

1380 Seaboard Industrial Blvd. Atlanta, Georgia, 30318

Section 9

ADDITIONAL SUPPORTING DOCUMENTATION

Alpha Spectroscopy

Quality Control Data

Weekly Background, Energy, and Efficiency Calibrations

Calibration Data Summary

Laboratory Name: **ALS -- Fort Collins**
PAI Work Order: 1705202

Prep SOP: PAI 701
Analytical SOP: PAI 714_Ra226

Reported on: Wednesday, May 24, 2017
1:14:14 PM

| Lab Sample ID Spectrum Analysis Date | QC Type | Batch ID Analysis Run | Test Name | Detector Id | Eff Spectrum Bkg Spectrum Egy Spectrum | Eff Date Bkg Date Egy Date | RESULTS %Efficiency Bkg CPM Energy keV | FLAGS Efficiency Background Energy | LCL %Efficiency Bkg CPM Energy keV | LWL %Efficiency Bkg CPM Energy keV | UWL %Efficiency Bkg CPM Energy keV | UCL %Efficiency Bkg CPM Energy keV |
|--|------------|----------------------------|-----------|----------------|--|-------------------------------------|---|---|---|---|---|---|
| 1705202-1 Spectrum #1 5/12/2017 | SMP | RAS170511-1 RAS170511-1 | RaISO | 88 | C17051088 B17051088 C17051088 | 5/10/2017 5/10/2017 5/10/2017 | 28.84 0.3010 5555.8 | Warning Pass Pass | 28.16 0.0000 5505.8 | 29.07 0.0500 5515.8 | 31.08 0.5000 5595.8 | 31.86 0.7500 5605.8 |
| 1705202-1 Spectrum #1 5/12/2017 | DUP | RAS170511-1 RAS170511-1 | RaISO | 89 | C17051089 B17051089 C17051089 | 5/10/2017 5/10/2017 5/10/2017 | 29.52 0.2310 5555.8 | Pass Pass Pass | 28.26 0.0000 5505.8 | 28.77 0.0500 5515.8 | 30.73 0.5000 5595.8 | 31.24 0.7500 5605.8 |
| RAS170511-1 Spectrum #1 5/12/2017 | MB | RAS170511-1 RAS170511-1 | RaISO | 81 | C17051081 B17051081 C17051081 | 5/10/2017 5/10/2017 5/10/2017 | 28.35 0.1830 5548.0 | Warning Pass Pass | 28.30 0.0000 5505.8 | 28.86 0.0500 5515.8 | 30.89 0.5000 5595.8 | 31.42 0.7500 5605.8 |
| RAS170511-1 Spectrum #1 5/12/2017 | LCS | RAS170511-1 RAS170511-1 | RaISO | 82 | C17051082 B17051082 C17051082 | 5/10/2017 5/10/2017 5/10/2017 | 28.92 0.2730 5548.0 | Pass Pass Pass | 27.98 0.0000 5496.0 | 28.49 0.0500 5506.0 | 30.51 0.5000 5586.0 | 31.02 0.7500 5596.0 |

Data Package ID: RAS1705202-1

| | | | | |
|----------------|------------------|-------------------------|---------------------------|---|
| Abbreviations: | Eff - Efficiency | Bkg - Background | LCL - Lower Control Limit | UWL - Upper Warning Limit |
| | Egy - Energy | CPM - Counts per Minute | LWL - Lower Warning Limit | UCL - Upper Control Limit |
| | | | | CI - The Analysis Date exceeds the Calibration Date by more than 14 days. |

Date Printed: Thursday, May 25, 2017

ALS -- Fort Collins

LIMS Version: 6.842

Page 1 of 1

Alpha Spec Calibration Source Re-Certification

Recalibration performed by Isotope Products Laboratories

Primary Certified Source

Source PA ID: 180
 Planchet Label: 9
 Recalibrated on: 10/4/2016
 Received by ALS on: 10/19/2013

| Values from certificate | |
|----------------------------|--|
| Source ID: 92MX223027 | |
| Total Activity: 3745.2 dpm | |
| Ref. Date: 10/15/2013 | |

| Nuclide | Act (Bq) | Act (dpm) | Half-life (yrs) | Decay Corrected |
|---------|----------|-----------|-----------------|-----------------|
| U-234: | 40.54 | 2972.4 | 2.48E+05 | 2972.38 dpm |
| U-235: | 1.09 | 65.58 | 7.04E+08 | 65.58 dpm |
| Am-241: | 11.79 | 707.4 | 432.17 | 704.04 dpm |
| TOTAL | | | | 3741.99 dpm |

Efficiency Determination for Detector:

129

| Source Serial# | PA ID | Sequential # | Count Date | Am-241 net cts | U-234 net cts | U-235 net cts | count dur (s) | Total cpm | Known dpm | Detector efficiency |
|----------------|-------|--------------|------------|----------------|---------------|---------------|---------------|-----------|-----------|---------------------|
| 92MX223027 | 180 | 97-18-103-09 | 10/4/16 | 7502 | 32112 | 1070 | 2100 | 1162.40 | 3741.99 | 31.06% |

Sources 1 through 8 activity determination

| Source Serial# | PA ID | Sequential # | Count Date | Am-241 net cts | U-234 net cts | U-235 net cts | count dur (s) | Detector Efficiency | Am-241 dpm | U-234 dpm | U-235 dpm | Combined dpm |
|----------------|-------|--------------|------------|----------------|---------------|---------------|---------------|---------------------|------------|-----------|-----------|--------------|
| 92MX2230326 | 182 | 97-18-103-01 | 10/4/16 | 12981 | 79837 | 2777 | 2100 | 31.06% | 1193.95 | 7343.17 | 255.42 | 8792.54 |
| 92MX2230328 | 183 | 97-18-103-02 | 10/4/16 | 15085 | 148128 | 3863 | 2100 | 31.06% | 1387.47 | 13824.37 | 355.31 | 15367.15 |
| 92MX2230324 | 184 | 97-18-103-03 | 10/4/16 | 67474 | 70483 | 2608 | 2100 | 31.06% | 6206.06 | 6482.82 | 239.88 | 12928.75 |
| 92MX2230321 | 185 | 97-18-103-04 | 10/4/16 | 21961 | 60440 | 2557 | 2100 | 31.06% | 2019.91 | 5559.09 | 235.19 | 7814.18 |
| 92MX2230325 | 186 | 97-18-103-05 | 10/4/16 | 97983 | 114458 | 3780 | 2100 | 31.06% | 9002.99 | 10527.51 | 347.67 | 19878.16 |
| 92MX2230322 | 187 | 97-18-103-06 | 10/4/16 | 72777 | 78983 | 2564 | 2100 | 31.06% | 6893.81 | 7347.40 | 237.67 | 14278.88 |
| 92MX2230323 | 188 | 97-18-103-07 | 10/4/16 | 43617 | 68953 | 2043 | 2100 | 31.06% | 4011.76 | 8342.09 | 187.81 | 10541.76 |
| 92MX2230329 | 189 | 97-18-103-08 | 10/4/16 | 33968 | 214074 | 7185 | 2100 | 31.06% | 3124.09 | 19889.89 | 661.77 | 23475.75 |

Efficiency Verification

| Source Serial# | PA ID | Sequential # | Count Date | Am-241 net cts | U-234 net cts | U-235 net cts | Count dur (s) | Total cpm | Known dpm | Detector efficiency | RPD | FLAG |
|----------------|-------|--------------|------------|----------------|---------------|---------------|---------------|-----------|-----------|---------------------|--------|------|
| 92MX223027 | 180 | 97-18-103-09 | 10/5/16 | 7807 | 32611 | 1278 | 2100 | 1181.31 | 3741.99 | 31.84% | -2.46% | PASS |

Sources 1 through 8 activity re-verification

| Source Serial# | PA ID | Sequential # | Combined Observed dpm | Combined Certified dpm* | Percent Difference % | Within 5% of Certified value? |
|----------------|-------|--------------|-----------------------|-------------------------|----------------------|-------------------------------|
| 92MX2230326 | 182 | 97-18-103-01 | 8792.54 | 8849.86 | -0.65% | Yes |
| 92MX2230328 | 183 | 97-18-103-02 | 15387.15 | 15982.35 | -3.91% | Yes |
| 92MX2230324 | 184 | 97-18-103-03 | 12928.75 | 13503.77 | -4.26% | Yes |
| 92MX2230321 | 185 | 97-18-103-04 | 7814.18 | 8161.24 | -4.25% | Yes |
| 92MX2230325 | 186 | 97-18-103-05 | 19878.16 | 20979.95 | -5.25% | No |
| 92MX2230322 | 187 | 97-18-103-06 | 14278.88 | 15285.63 | -6.56% | No |
| 92MX2230323 | 188 | 97-18-103-07 | 10541.76 | 10723.95 | -1.70% | Yes |
| 92MX2230329 | 189 | 97-18-103-08 | 23475.75 | 23583.84 | -0.50% | Yes |

* Certificate values decay corrected to the count date

Data from certificates

| Reference Date | U-234 (Bq) | U-234 (dpm) | U-235 (Bq) | U-235 (dpm) | Am-241 (Bq) | Am-241 (dpm) |
|----------------|------------|-------------|------------|-------------|-------------|--------------|
| 5/1/2003 | 124.10 | 7449.00 | 2.43 | 145.74 | 21.43 | 1285.80 |
| 5/1/2003 | 236.30 | 14358.00 | 4.20 | 252.00 | 23.55 | 1413.00 |
| 5/1/2003 | 119.40 | 7164.00 | 1.93 | 115.56 | 106.00 | 6360.00 |
| 4/1/2003 | 101.00 | 6060.00 | 1.26 | 75.84 | 34.50 | 2070.00 |
| 4/1/2003 | 203.00 | 12180.00 | 3.41 | 204.72 | 146.40 | 8784.00 |
| 4/1/2003 | 132.90 | 7974.00 | 3.17 | 189.96 | 121.30 | 7278.00 |
| 4/1/2003 | 107.10 | 6426.00 | 0.83 | 55.54 | 72.26 | 4335.80 |
| 5/1/2003 | 334.80 | 20088.00 | 6.55 | 393.18 | 53.02 | 3181.20 |

OK JP
 10/5/16
 Expires
 10/4/2017

Analyst: ORTEC

Detector: 129

9:14:40AM 10/5/2016

Energy Calibration: SOURCE190_10.04.16 (#9)
Description:

Calibration

Analysis Date: 10/4/2016 12:09:56PM
Calibration Type: Energy And Efficiency

Certificate ID: A9 RSO#190
Prepared by: Isotope Product Laboratories
Description:

Source Info

Certification Date: 10/15/2013 10:44:40AM

Acquisition

Detector: 129, SN:5505430, ID: 129
Acquisition Start Date: 10/4/2016 11:26:06AM

Energy Calibration Equation:

Gain = 9.9003 keV / Ch


Offset = 3,021.28 keV

Quadratic = 0.0000 keV / Ch²

Live Time: 35.00 min.
Real Time: 35.01 min.

Efficiency Calibration Name: SOURCE190_10.04.16 (#9)

Efficiency: 33.86% +/- 1.39% TPU(2 sigma)

**General Analysis**

Method: Manual (ROI)

Initial Calibration: Yes

Algorithm: Linear

Shelf: 0

Nuclide Activity Summary

| Nuclide | Peak Channel | Peak Energy keV | ROI Start Channel | ROI End Channel | Peak FWHM keV | Gross Counts | Net Count Rate (cpm) |
|---------|--------------|-----------------|-------------------|-----------------|---------------|--------------|----------------------|
| U-235 | 139 | 4,396.00 | 100 | 152 | 0.00 | 1,070.00 | 33.71 |
| U-234 | 177 | 4,775.80 | 153 | 190 | 68.12 | 32,112.00 | 953.94 |
| Am-241 | 249 | 5,485.70 | 210 | 270 | 71.83 | 7,502.00 | 221.80 |

JP 10/4/16

Analyst: ORTEC

Detector: 129

12:49:08PM 10/4/2016

Energy Calibration: SOURCE182_10.04.16 (#1)

Description:

CalibrationAnalysis Date: 10/4/2016 12:47:23PM
Calibration Type: Energy And Efficiency

Certificate ID: A1 RSO#182

Prepared by: Isotope Product Laboratories

Description:

Source Info

Certification Date: 5/1/2003 10:27:02AM

Acquisition

Detector: 129, SN:5505430, ID: 129

Acquisition Start Date: 10/4/2016 12:11:42PM

Live Time: 35.00 min.

Real Time: 35.02 min.

Energy Calibration Equation:


Gain = 9.9003 keV / Ch

Offset = 3,021.28 keV

Quadratic = 0.0000 keV / Ch²

Efficiency Calibration Name: SOURCE182_10.04.16 (#1)

Efficiency: 33.04% +/- 2.03% TPU(2 sigma)

**General Analysis**

Method: Manual (ROI)

Algorithm: Linear

Initial Calibration: Yes

Shelf: 0

Nuclide Activity Summary

| Nuclide | Peak Channel | Peak Energy keV | ROI Start Channel | ROI End Channel | Peak FWHM keV | Gross Counts | Net Count Rate (cpm) |
|---------|--------------|-----------------|-------------------|-----------------|---------------|--------------|----------------------|
| U-235 | 139 | 4,396.00 | 100 | 152 | 84.43 | 2,777.00 | 79.34 |
| U-234 | 177 | 4,775.80 | 153 | 190 | 71.29 | 79,837.00 | 2,281.06 |
| Am-241 | 249 | 5,485.70 | 210 | 270 | 72.86 | 12,981.00 | 370.89 |

JP 10/4/16

Analyst: ORTEC

Detector: 129

1:25:35PM 10/4/2016

Energy Calibration: SOURCE183_10.04.16 (#2)

Description:

CalibrationAnalysis Date: 10/4/2016 1:25:29PM
Calibration Type: Energy And Efficiency

Certificate ID: A2 RSO#183

Prepared by: Isotope Product Laboratories

Description:

Source Info

Certification Date: 5/1/2003 10:33:40AM

Acquisition

Detector: 129, SN:5505430, ID: 129

Acquisition Start Date: 10/4/2016 12:49:04PM

Live Time: 35.00 min.

Real Time: 35.03 min.

Energy Calibration Equation:


Gain = 9.9003 keV / Ch

Offset = 3,021.28 keV

Quadratic = 0.0000 keV / Ch²

Efficiency Calibration Name: SOURCE183_10.04.16 (#2)

Efficiency: 32.74% +/- 1.25% TPU(2 sigma)

**General Analysis**

Method: Manual (ROI)

Algorithm: Linear

Initial Calibration: Yes

Shelf: 0

Nuclide Activity Summary

| Nuclide | Peak Channel | Peak Energy keV | ROI Start Channel | ROI End Channel | Peak FWHM keV | Gross Counts | Net Count Rate (cpm) |
|---------|--------------|-----------------|-------------------|-----------------|---------------|--------------|----------------------|
| U-235 | 139 | 4,396.00 | 100 | 152 | 75.93 | 3,863.00 | 110.37 |
| U-234 | 177 | 4,775.80 | 153 | 190 | 69.70 | 148,128.00 | 4,232.23 |
| Am-241 | 249 | 5,485.70 | 210 | 270 | 73.20 | 15,085.00 | 431.00 |

JP 10/4/16

Analyst: ORTEC

Detector: 129

2:17:25PM 10/4/2016

Energy Calibration: SOURCE184_10.04.16 (#3)

Description:

CalibrationAnalysis Date: 10/4/2016 2:16:56PM
Calibration Type: Energy And Efficiency

Certificate ID: A3 RSO#184

Prepared by: Isotope Product Laboratories

Description:

Source Info

Certification Date: 5/1/2003 10:36:52AM

Acquisition

Detector: 129, SN:5505430, ID: 129

Acquisition Start Date: 10/4/2016 1:26:53PM

Live Time: 35.00 min.

Real Time: 35.03 min.

Energy Calibration Equation:


Gain = 9.9003 keV / Ch

Offset = 3,021.28 keV

Quadratic = 0.0000 keV / Ch²

Efficiency Calibration Name: SOURCE184_10.04.16 (#3)

Efficiency: 31.83% +/- 1.26% TPU(2 sigma)

**General Analysis**

Method: Manual (ROI)

Algorithm: Linear

Initial Calibration: Yes

Shelf: 0

Nuclide Activity Summary

| Nuclide | Peak Channel | Peak Energy keV | ROI Start Channel | ROI End Channel | Peak FWHM keV | Gross Counts | Net Count Rate (cpm) |
|---------|--------------|-----------------|-------------------|-----------------|---------------|--------------|----------------------|
| U-235 | 139 | 4,396.00 | 100 | 152 | 0.00 | 2,608.00 | 74.51 |
| U-234 | 177 | 4,775.80 | 153 | 190 | 74.20 | 70,483.00 | 2,013.80 |
| Am-241 | 249 | 5,485.70 | 210 | 270 | 74.76 | 67,474.00 | 1,927.83 |

JP 10/4/16

Analyst: ORTEC

Detector: 129

9:00:58AM 10/5/2016

Energy Calibration: SOURCE185_10.04.16 (#4)

Description:

Calibration

Analysis Date: 10/4/2016 2:53:33PM
Calibration Type: Energy And Efficiency

Certificate ID: A4 RSO#185

Prepared by: Isotope Product Laboratories

Description:

Source Info

Certification Date: 4/1/2003 10:38:09AM

Acquisition

Detector: 129, SN:5505430, ID: 129

Acquisition Start Date: 10/4/2016 2:18:14PM

Live Time: 35.00 min.

Real Time: 35.02 min.

Energy Calibration Equation:


Gain = 9.9003 keV / Ch

Offset = 3,021.28 keV

Quadratic = 0.0000 keV / Ch²

Efficiency Calibration Name: SOURCE185_10.04.16 (#4)

Efficiency: 31.73% +/- 1.30% TPU(2 sigma)



General Analysis

Method: Manual (ROI)

Algorithm: Linear

Initial Calibration: Yes

Shelf: 0

Nuclide Activity Summary

| Nuclide | Peak Channel | Peak Energy keV | ROI Start Channel | ROI End Channel | Peak FWHM keV | Gross Counts | Net Count Rate (cpm) |
|---------|--------------|-----------------|-------------------|-----------------|---------------|--------------|----------------------|
| U-235 | 139 | 4,396.00 | 100 | 152 | 79.20 | 2,557.00 | 73.06 |
| U-234 | 177 | 4,775.80 | 153 | 190 | 80.97 | 60,440.00 | 1,726.86 |
| Am-241 | 249 | 5,485.70 | 210 | 270 | 77.99 | 21,961.00 | 627.46 |

Analyst: ORTEC

Detector: 129

7:04:12AM 10/5/2016

Energy Calibration: SOURCE188_10.04.16 (#7)

Description:

CalibrationAnalysis Date: 10/5/2016 7:03:07AM
Calibration Type: Energy And Efficiency

Certificate ID: A7 RSO#188

Prepared by: Isotope Product Laboratories

Description:

Source Info

Certification Date: 4/1/2003 10:42:01AM

Acquisition

Detector: 129, SN:5505430, ID: 129

Acquisition Start Date: 10/5/2016 6:13:44AM

Live Time: 35.00 min.

Real Time: 35.02 min.

Energy Calibration Equation:


Gain = 9.9003 keV / Ch

Offset = 3,021.28 keV

Quadratic = 0.0000 keV / Ch²

Efficiency Calibration Name: SOURCE188_10.04.16 (#7)

Efficiency: 31.88% +/- 1.31% TPU(2 sigma)

**General Analysis**

Method: Manual (ROI)

Algorithm: Linear

Initial Calibration: Yes

Shelf: 0

Nuclide Activity Summary

| Nuclide | Peak Channel | Peak Energy keV | ROI Start Channel | ROI End Channel | Peak FWHM keV | Gross Counts | Net Count Rate (cpm) |
|---------|--------------|-----------------|-------------------|-----------------|---------------|--------------|----------------------|
| U-235 | 139 | 4,396.00 | 100 | 152 | 77.73 | 2,043.00 | 58.37 |
| U-234 | 177 | 4,775.80 | 153 | 190 | 69.01 | 68,953.00 | 1,970.09 |
| Am-241 | 249 | 5,485.70 | 210 | 270 | 71.83 | 43,617.00 | 1,246.20 |

Analyst: ORTEC

Detector: 129

7:44:05AM

10/5/2016

Energy Calibration: SOURCE189_10.04.16 (#8)

Description:

Calibration

Analysis Date: 10/5/2016 7:43:56AM

Calibration Type: Energy And Efficiency

Certificate ID: A8 RSO#189

Prepared by: Isotope Product Laboratories

Description:

Source Info

Certification Date: 5/1/2003 10:43:18AM

Acquisition

Detector: 129, SN:5505430, ID: 129

Acquisition Start Date: 10/5/2016 7:04:08AM

Live Time: 35.00 min.

Real Time: 35.05 min.

Energy Calibration Equation:


Gain = 9.9003 keV / Ch

Offset = 3,021.28 keV

Quadratic = 0.0000 keV / Ch²

Efficiency Calibration Name: SOURCE189_10.04.16 (#8)

Efficiency: 33.82% +/- 1.28% TPU(2 sigma)

**General Analysis**

Method: Manual (ROI)

Algorithm: Linear

Initial Calibration: Yes

Shelf: 0

Nuclide Activity Summary

| Nuclide | Peak Channel | Peak Energy keV | ROI Start Channel | ROI End Channel | Peak FWHM keV | Gross Counts | Net Count Rate (cpm) |
|---------|--------------|-----------------|-------------------|-----------------|---------------|--------------|----------------------|
| U-235 | 139 | 4,396.00 | 100 | 152 | 87.84 | 7,195.00 | 205.57 |
| U-234 | 177 | 4,775.80 | 153 | 190 | 74.72 | 214,074.00 | 6,116.40 |
| Am-241 | 249 | 5,485.70 | 210 | 270 | 74.87 | 33,966.00 | 970.46 |

Analyst: ORTEC

Detector: 129

8:36:03AM 10/5/2016

Energy Calibration: SOURCE190A_10.04.16 (#9)

Description:

Calibration

Analysis Date: 10/5/2016 8:35:09AM

Calibration Type: Energy And Efficiency

Certificate ID: A9 RSO#190

Prepared by: Isotope Product Laboratories

Description:

Source Info

Certification Date: 10/15/2013 10:44:40AM

Acquisition

Detector: 129, SN:5505430, ID: 129

Acquisition Start Date: 10/5/2016 7:45:09AM

Live Time: 35.00 min.

Real Time: 35.01 min.

Energy Calibration Equation:


Gain = 9.9003 keV / Ch

Offset = 3,021.28 keV

Quadratic = 0.0000 keV / Ch²

Efficiency Calibration Name: SOURCE190A_10.04.16 (#9)

Efficiency: 33.67% +/- 1.38% TPU(2 sigma)

**General Analysis**

Method: Manual (ROI)

Algorithm: Linear

Initial Calibration: Yes

Shelf: 0

Nuclide Activity Summary

| Nuclide | Peak Channel | Peak Energy keV | ROI Start Channel | ROI End Channel | Peak FWHM keV | Gross Counts | Net Count Rate (cpm) |
|---------|--------------|-----------------|-------------------|-----------------|---------------|--------------|----------------------|
| U-235 | 139 | 4,396.00 | 100 | 152 | 0.00 | 1,278.00 | 36.51 |
| U-234 | 177 | 4,775.80 | 153 | 190 | 68.47 | 32,611.00 | 931.74 |
| Am-241 | 249 | 5,485.70 | 210 | 270 | 71.17 | 7,807.00 | 223.06 |



Eckert & Ziegler

Isotope Products

24937 Avenue Tibbitts
Valencia, California 91355

Tel 661-309-1010

Fax 661-257-8303

#190
Received 10/18/13

CERTIFICATE OF CALIBRATION MIXED ALPHA STANDARD SOURCE

| | | | |
|---------------------|---------------------------|-----------------|---------------------|
| Radionuclide: | U-234 | Customer: | ALS LABORATORY |
| Radionuclide: | U-235 | P.O. No.: | FC 3595 / R5576 |
| Radionuclide: | Am-241 | Catalog No.: | *SOURCE-RECAL-STD |
| Half-life (U-234): | (2.454 ± 0.006)E+05 years | Reference Date: | 15-Oct-13 12:00 PST |
| Half-life (U-235): | (7.037 ± 0.011)E+08 years | Source No.: | 92MIX223027 |
| Half-life (Am-241): | 432.17 ± 0.66 years | | |

Contained Radioactivity:

| | | | | | | | | | |
|--------|---------|------|-------|----|-----------------|--------|------|-------|----|
| U-234: | 1.339 | nCi, | 49.54 | Bq | Am-241: | 0.3187 | nCi, | 11.79 | Bq |
| U-235: | 0.02954 | nCi, | 1.093 | Bq | Total Activity: | 1.687 | nCi, | 62.42 | Bq |

Physical Description:

| | |
|------------------------------|---|
| A. Capsule type: | Disk (22 mm OD x 0.79 mm THK) |
| B. Nature of active deposit: | Electrodeposited and diffusion bonded oxide |
| C. Active diameter/volume: | 19 mm |
| D. Backing: | Stainless steel |
| E. Cover: | None |

Radioimpurities: Not determined

Method of Calibration:

This source was assayed using a windowless internal gas flow proportional counter for total alpha activity. Individual nuclide ratios were taken from those determined in May 2001.

Uncertainty of Measurement:

| | |
|---|---------|
| A. Type A (random) uncertainty: | ± 0.5 % |
| B. Type B (systematic) uncertainty: | ± 3.0 % |
| C. Uncertainty in aliquot weighing: | ± 0.0 % |
| D. Total uncertainty at the 99% confidence level: | ± 3.0 % |

Notes:

- See reverse side for leak test(s) performed on this source.
- IPL participates in a NIST measurement assurance program to establish and maintain implicit traceability for a number of nuclides, based on the blind assay (and later NIST certification) of Standard Reference Materials (as in NRC Regulatory Guide 4.15).
- Nuclear data was taken from "Table of Radioactive Isotopes", edited by Virginia Shirley, 1986.
- This source has a working life of 2 years.
- This source had a total alpha surface emission rate of 1893 α/min in 2π on 20-Sep-13.


Quality Control

2-OCT-13
Date

IPL Ref. No.: 987-28

ISO 9001 CERTIFIED

Medical Imaging Laboratory

24937 Avenue Tibbitts Valencia, California 91355

Industrial Gauging Laboratory

1800 North Keystone Street Burbank, California 91504

61 of 67



**Isotope Products
Laboratories**

An Eckert & Ziegler Company

24937 Avenue Tibbitts
Valencia, California 91355

Tel 661•309•1010
Fax 661•257•8303

Re Calibrated 10/4/16

New Exp Date 10/4/2017

PAI 1875
recalibrated 4-15-03

CERTIFICATE OF CALIBRATION MIXED ALPHA STANDARD SOURCE

Radionuclide A: U-234
Radionuclide B: U-235
Radionuclide C: Am-241
Half Life (U-234): $(2.454 \pm 0.006)E+05$ years
Half Life (U-235): $(7.037 \pm 0.011)E+08$ years
Half Life (Am-241): 432.17 ± 0.66 years

Customer: PARAGON ANALYTICS, INC.
P.O. No.: EW040203/R2193
Catalog No.: MISC-STD
Reference Date: 1-May-03 12:00 PST
Source No.: 92MIX2203026

Contained Radioactivity:

| | | | |
|--------|------------------------|------------------------|-----------------------|
| U-234: | 3.354 nCi (124.1 Bq) | Am-241: | 0.5793 nCi (21.43 Bq) |
| U-235: | 0.06566 nCi (2.429 Bq) | Total Activity: | 3.999 nCi (148.0 Bq) |

Physical description:

| | |
|------------------------------|--|
| A. Capsule type: | Disk (22 mm OD X 0.79 mm THK) |
| B. Nature of active deposit: | Electrodeposited and diffusion bonded oxides |
| C. Active Diameter: | 19 mm |
| D. Backing: | Stainless steel |
| E. Cover: | None |

Radioimpurities:

Not determined

Method of Calibration:

This source was assayed using a windowless internal gas flow proportional counter for total alpha activity. Individual nuclide ratios were taken from those determined in Mar 1998.

Uncertainty of Measurement:

| | |
|---|-------------|
| A. Type A (random) uncertainty: | $\pm 0.7\%$ |
| B. Type B (systematic) uncertainty: | $\pm 3.0\%$ |
| C. Uncertainty in aliquot weighing: | $\pm 0.0\%$ |
| D. Total uncertainty at the 99% confidence level: | $\pm 3.1\%$ |

Notes:

- See reverse side for leak test(s) performed on this source.
- IPL participates in a NIST measurement assurance program to establish and maintain implicit traceability for a number of nuclides, based on the blind assay (and later NIST certification) of Standard Reference Materials (As in NRC Regulatory Guide 4.15).
- Nuclear data was taken from "Table of Radioactive Isotopes", edited by Virginia Shirley, 1986.
- This source has a working life of 2 years.
- This source had a total alpha surface emission rate of 4483 α /min in 2 π on 11 Apr 03.

Daniel James Van Dalsem
Quality Control

15-Apr-03
Date Signed

IPL Ref. No.: 987-7

ISO 9001 CERTIFIED

Medical Imaging Laboratory
24937 Avenue Tibbitts Valencia, California 91355

Industrial Gauging Laboratory
1800 North Keystone Street Burbank, California 91504



**Isotope Products
Laboratories**

An Eckert & Ziegler Company

24937 Avenue Tibbitts
Valencia, California 91355

Tel 661-309-1010
Fax 661-257-8303

Re-Calibrated 10/4/16
New Exp Date 10/4/2017

PAI 183
Recalibrated 4-15-03

CERTIFICATE OF CALIBRATION MIXED ALPHA STANDARD SOURCE

Radionuclide A: U-234
Radionuclide B: U-235
Radionuclide C: Am-241
Half Life (U-234): $(2.454 \pm 0.006)E+05$ years
Half Life (U-235): $(7.037 \pm 0.011)E+08$ years
Half Life (Am-241): 432.17 ± 0.66 years

Customer: PARAGON ANALYTICS, INC.
P.O. No.: EW040203/R2193
Catalog No.: MISC-STD
Reference Date: 1-May-03 12:00 PST
Source No.: 92MIX2203028

Contained Radioactivity:

U-234: 6.467 nCi (239.3 Bq)
U-235: 0.1135 nCi (4.200 Bq)

Am-241: 0.6366 nCi (23.55 Bq)
Total Activity: 7.217 nCi (267.1 Bq)

Physical description:

| | |
|------------------------------|--|
| A. Capsule type: | Disk (22 mm OD X 0.79 mm THK) |
| B. Nature of active deposit: | Electrodeposited and diffusion bonded oxides |
| C. Active Diameter: | 19 mm |
| D. Backing: | Stainless steel |
| E. Cover: | None |

Radiopurities:

Not determined

Method of Calibration:

This source was assayed using a windowless internal gas flow proportional counter for total alpha activity. Individual nuclide ratios were taken from those determined in Aug 1992.

Uncertainty of Measurement:

| | |
|---|-------------|
| A. Type A (random) uncertainty: | $\pm 0.7\%$ |
| B. Type B (systematic) uncertainty: | $\pm 3.0\%$ |
| C. Uncertainty in aliquot weighing: | $\pm 0.0\%$ |
| D. Total uncertainty at the 99% confidence level: | $\pm 3.1\%$ |

Notes:

- See reverse side for leak test(s) performed on this source.
- IPL participates in a NIST measurement assurance program to establish and maintain implicit traceability for a number of nuclides, based on the blind assay (and later NIST certification) of Standard Reference Materials (As in NRC Regulatory Guide 4.15).
- Nuclear data was taken from "Table of Radioactive Isotopes", edited by Virginia Shirley, 1986.
- This source has a working life of 2 years.
- This source had a total alpha surface emission rate of 8091 α /min in 2π on 11 Apr 03.

Daniel James Van Dalsen
Quality Control

15-Apr-03
Date Signed

IPL Ref. No.: 987-7

ISO 9001 CERTIFIED

Medical Imaging Laboratory
24937 Avenue Tibbitts Valencia, California 91355

Industrial Gauging Laboratory
1800 North Keystone Street Burbank, California 91504



**Isotope Products
Laboratories**

An Eckert & Ziegler Company

24937 Avenue Tibbitts
Valencia, California 91355

Tel 661-309-1010
Fax 661-257-8303

Re-Calibrated 10/4/16
New Exp Date 10/4/2017
PAT I.D. 184
recalibrated 4-15-03

CERTIFICATE OF CALIBRATION MIXED ALPHA STANDARD SOURCE

Radionuclide A: U-234
Radionuclide B: U-235
Radionuclide C: Am-241
Half Life (U-234): $(2.454 \pm 0.006)E+05$ years
Half Life (U-235): $(7.037 \pm 0.011)E+08$ years
Half Life (Am-241): 432.17 ± 0.66 years

Customer: PARAGON ANALYTICS, INC.
P.O. No.: EW040203/R2193
Catalog No.: MISC-STD
Reference Date: 1-May-03 12:00 PST
Source No.: 92MIX2203024

Contained Radioactivity:

U-234: 3.227 nCi (119.4 Bq)
U-235: 0.05205 nCi (1.926 Bq)

Am-241: 2.866 nCi (106.0 Bq)
Total Activity: 6.145 nCi (227.3 Bq)

Physical description:

A. Capsule type: Disk (22 mm OD X 0.79 mm THK)
B. Nature of active deposit: Electrodeposited and diffusion bonded oxides
C. Active Diameter: 19 mm
D. Backing: Stainless steel
E. Cover: None

Radioimpurities:

Not determined

Method of Calibration:

This source was assayed using a windowless internal gas flow proportional counter for total alpha activity. Individual nuclide ratios were taken from those determined in Aug 1992.

Uncertainty of Measurement:

A. Type A (random) uncertainty: $\pm 0.6\%$
B. Type B (systematic) uncertainty: $\pm 3.0\%$
C. Uncertainty in aliquot weighing: $\pm 0.0\%$
D. Total uncertainty at the 99% confidence level: $\pm 3.1\%$

Notes:

- See reverse side for leak test(s) performed on this source.
- IPL participates in a NIST measurement assurance program to establish and maintain implicit traceability for a number of nuclides, based on the blind assay (and later NIST certification) of Standard Reference Materials (As in NRC Regulatory Guide 4.15).
- Nuclear data was taken from "Table of Radioactive Isotopes", edited by Virginia Shirley, 1986.
- This source has a working life of 2 years.
- This source had a total alpha surface emission rate of 6889 α/min in 2π on 11 Apr 03.

Daniel James Van Dalsem
Quality Control

15-Apr-03
Date Signed

IPL Ref. No.: 987-7

ISO 9001 CERTIFIED

Medical Imaging Laboratory
24937 Avenue Tibbitts Valencia, California 91355

Industrial Gauging Laboratory
1800 North Keystone Street Burbank, California 91504



**Isotope Products
Laboratories**

An Eckert & Ziegler Company

24937 Avenue Tibbitts
Valencia, California 91355

Tel 661-309-1010
Fax 661-257-8303

Recalibrated 10/4/16
New Exp Date 10/4/2017
JP 10/5/16

PAI ID 00185
rec'd from recalibrator
3-28-03

CERTIFICATE OF CALIBRATION ALPHA STANDARD SOURCE

Radionuclide A: U-234
Radionuclide B: U-235
Radionuclide C: Am-241
Half Life (U-234): $(2.454 \pm 0.006)E+05$ years
Half Life (U-235): $(7.037 \pm 0.011)E+08$ years
Half Life (Am-241): 432.17 ± 0.66 years

Customer: PARAGON ANALYTICS, INC.
P.O. No.: EW030603/R2155
Catalog No.: MISC-STD
Reference Date: 1-Apr-03 12:00 PST
Source No.: 92MIX2203021

Contained Radioactivity:

U-234: 2.731 nCi (101.0 Bq)
U-235: 0.03416 nCi (1.264 Bq)

Am-241: 0.9325 nCi (34.50 Bq)
Total Activity: 3.698 nCi (136.8 Bq)

Physical description:

| | |
|------------------------------|--|
| A. Capsule type: | Disk (22 mm OD X 0.79 mm THK) |
| B. Nature of active deposit: | Electrodeposited and diffusion bonded oxides |
| C. Active Diameter: | 19 mm |
| D. Backing: | Stainless steel |
| E. Cover: | None |

Radioimpurities:

Not determined

Method of Calibration:

This source was assayed using a windowless internal gas flow proportional counter for total alpha activity. Individual nuclide ratios were taken from those determined in Aug 1992.

Uncertainty of Measurement:

| | |
|---|-------------|
| A. Type A (random) uncertainty: | $\pm 0.8\%$ |
| B. Type B (systematic) uncertainty: | $\pm 3.1\%$ |
| C. Uncertainty in aliquot weighing: | $\pm 0.0\%$ |
| D. Total uncertainty at the 99% confidence level: | $\pm 3.2\%$ |

Notes:

- See reverse side for leak test(s) performed on this source.
- IPL participates in a NIST measurement assurance program to establish and maintain implicit traceability for a number of nuclides, based on the blind assay (and later NIST certification) of Standard Reference Materials (As in NRC Regulatory Guide 4.15).
- Nuclear data was taken from "Table of Radioactive Isotopes", edited by Virginia Shirley, 1986.
- This source has a working life of 2 years.
- This source had a total alpha surface emission rate of 4145 α /min in 2 π on 18 Mar 03.

Daniel James Van Dalsem
Quality Control

19-Mar-03
Date Signed

IPL Ref. No.: 987-2

ISO 9001 CERTIFIED

Medical Imaging Laboratory
24937 Avenue Tibbitts Valencia, California 91355

Industrial Gauging Laboratory
1800 North Keystone Street Burbank, California 91504



**Isotope Products
Laboratories**

An Eckert & Ziegler Company

24937 Avenue Tibbitts
Valencia, California 91355

Tel 661•309•1010

Fax 661•257•8303

Re Calibrated 10/4/16
New Exp Date 10/4/2017
JP10516
PAID 188
rec'd for recalibration
3-28-03

CERTIFICATE OF CALIBRATION ALPHA STANDARD SOURCE

Radionuclide A: U-234
Radionuclide B: U-235
Radionuclide C: Am-241
Half Life (U-234): $(2.454 \pm 0.006)E+05$ years
Half Life (U-235): $(7.037 \pm 0.011)E+08$ years
Half Life (Am-241): 432.17 ± 0.66 years

Customer: PARAGON ANALYTICS, INC.
P.O. No.: EW030603/R2155
Catalog No.: MISC-STD
Reference Date: 1-Apr-03 12:00 PST
Source No.: 92MIX2203023

Contained Radioactivity:

| | | | |
|--------|-------------------------|-----------------|----------------------|
| U-234: | 2.895 nCi (107.1 Bq) | Am-241: | 1.953 nCi (72.26 Bq) |
| U-235: | 0.02502 nCi (0.9257 Bq) | Total Activity: | 4.873 nCi (180.3 Bq) |

Physical description:

| | |
|------------------------------|--|
| A. Capsule type: | Disk (22 mm OD X 0.79 mm THK) |
| B. Nature of active deposit: | Electrodeposited and diffusion bonded oxides |
| C. Active Diameter: | 19 mm |
| D. Backing: | Stainless steel |
| E. Cover: | None |

Radioimpurities:

Not determined

Method of Calibration:

This source was assayed using a windowless internal gas flow proportional counter for total alpha activity. Individual nuclide ratios were taken from those determined in Aug 1992.

Uncertainty of Measurement:

| | |
|---|-------------|
| A. Type A (random) uncertainty: | $\pm 0.8\%$ |
| B. Type B (systematic) uncertainty: | $\pm 3.1\%$ |
| C. Uncertainty in aliquot weighing: | $\pm 0.0\%$ |
| D. Total uncertainty at the 99% confidence level: | $\pm 3.2\%$ |

Notes:

- See reverse side for leak test(s) performed on this source.
- IPL participates in a NIST measurement assurance program to establish and maintain implicit traceability for a number of nuclides, based on the blind assay (and later NIST certification) of Standard Reference Materials (As in NRC Regulatory Guide 4.15).
- Nuclear data was taken from "Table of Radioactive Isotopes", edited by Virginia Shirley, 1986.
- This source has a working life of 2 years.
- This source had a total alpha surface emission rate of 5463 α /min in 2π on 18 Mar 03.

Daniel James Van Dalsem
Quality Control

19-Mar-03
Date Signed

IPL Ref. No.: 987-2

ISO 9001 CERTIFIED

Medical Imaging Laboratory
24937 Avenue Tibbitts Valencia, California 91355

Industrial Gauging Laboratory
1800 North Keystone Street Burbank, California 91504



**Isotope Products
Laboratories**

An Eckert & Ziegler Company

24937 Avenue Tibbitts
Valencia, California 91355

Tel 661-309-1010

Fax 661-257-8303

ReCalibrated 10/4/16
New Exp. Date 10/4/2017
JP 10/5/16

PAI ID 189
rec'd 4-21-03
recalibrated 4-15-03

CERTIFICATE OF CALIBRATION MIXED ALPHA STANDARD SOURCE

Radionuclide A: U-234
Radionuclide B: U-235
Radionuclide C: Am-241
Half Life (U-234): $(2.454 \pm 0.006)E+05$ years
Half Life (U-235): $(7.037 \pm 0.011)E+08$ years
Half Life (Am-241): 432.17 ± 0.66 years

Customer: PARAGON ANALYTICS, INC.
P.O. No.: EW040203/R2193
Catalog No.: MISC-STD
Reference Date: 1-May-03 12:00 PST
Source No.: 92MIX2203029

Contained Radioactivity:

U-234: 9.048 nCi (334.8 Bq)
U-235: 0.1771 nCi (6.553 Bq)

Am-241: 1.433 nCi (53.02 Bq)
Total Activity: 10.66 nCi (394.4 Bq)

Physical description:

| | |
|------------------------------|--|
| A. Capsule type: | Disk (22 mm OD X 0.79 mm THK) |
| B. Nature of active deposit: | Electrodeposited and diffusion bonded oxides |
| C. Active Diameter: | 19 mm |
| D. Backing: | Stainless steel |
| E. Cover: | None |

Radioimpurities:

Not determined

Method of Calibration:

This source was assayed using a windowless internal gas flow proportional counter for total alpha activity. Individual nuclide ratios were taken from those determined in Mar 1998.

Uncertainty of Measurement:

| | |
|---|-------------|
| A. Type A (random) uncertainty: | $\pm 0.5\%$ |
| B. Type B (systematic) uncertainty: | $\pm 3.0\%$ |
| C. Uncertainty in aliquot weighing: | $\pm 0.0\%$ |
| D. Total uncertainty at the 99% confidence level: | $\pm 3.0\%$ |

Notes:

- See reverse side for leak test(s) performed on this source.
- IPL participates in a NIST measurement assurance program to establish and maintain implicit traceability for a number of nuclides, based on the blind assay (and later NIST certification) of Standard Reference Materials (As in NRC Regulatory Guide 4.15).
- Nuclear data was taken from "Table of Radioactive Isotopes", edited by Virginia Shirley, 1986.
- This source has a working life of 2 years.
- This source had a total alpha surface emission rate of $11950 \alpha/\text{min}$ in 2π on 11 Apr 03.

Daniel James Van Dalsem
Quality Control

15-Apr-03
Date Signed

IPL Ref. No.: 987-7

Medical Imaging Laboratory
24937 Avenue Tibbitts Valencia, California 91355

ISO 9001 CERTIFIED

Industrial Gauging Laboratory
1800 North Keystone Street Burbank, California 91504