

# Session Report

1/9/2018

## Information Panel

Name S109\_BGN010013\_08012018\_134010  
Start Time 1/5/2018 12:07:02 PM  
Stop Time 1/8/2018 1:50:29 PM  
Device Name BGN010013  
Model Type SoundPro DL  
Device Firmware Rev R.13F  
Comments

## Summary Data Panel

Description	Meter	Value	Description	Meter	Value
Leq	1	46.6 dB			
Exchange Rate	1	3 dB	Weighting	1	A
Response	1	FAST	Bandwidth	1	OFF
Leq	2	63.7 dB			
Exchange Rate	2	3 dB	Weighting	2	C
Response	2	SLOW			

## Calibration History

Date	Calibration Action	Level	Cal. Model Type	Serial Number	Cert. Due Date
1/5/2018 12:06:45 PM	Calibration	114.0			
1/8/2018 1:52:10 PM	Verification	114.1			

## Logged Data Chart

S109\_BGN010013\_08012018\_134010: Logged Data Chart

