

# Session Report

2/8/2018

## Information Panel


Name S007\_BGR100002\_08022018\_082450  
Start Time 2/7/2018 5:08:35 PM  
Stop Time 2/7/2018 6:20:00 PM  
Device Name BGR100002  
Model Type SoundPro DL  
Device Firmware Rev R.13H  
Comments

## Summary Data Panel

Description	Meter	Value	Description	Meter	Value
Leq	1	59.9 dB			
Exchange Rate	1	3 dB	Weighting	1	A
Response	1	SLOW	Bandwidth	1	OFF
Leq	2	77.8 dB			
Exchange Rate	2	3 dB	Weighting	2	C
Response	2	FAST			

## Logged Data Chart

S007\_BGR100002\_08022018\_082450: Logged Data Chart



## Calibration History

Date	Calibration Action	Level	Cal. Model Type	Serial Number	Cert. Due Date
2/7/2018 5:07:28 PM	Calibration	114.0			
2/7/2018 6:20:56 PM	Verification	114.1			