

Analytical Report

Report Summary

Client: Alamosa Drilling
Chain Of Custody Number:
Samples Received: 9/14/2018 8:44:00AM
Job Number: 18076-0001
Work Order: P809029
Project Name/Location: Chavez #1 Spill

Report Reviewed By:



Date: 9/19/18

Walter Hinchman, Laboratory Director



Date: 9/19/18

Tim Cain, Project Manager



Envirotech Inc. certifies the test results meet all requirements of TNI unless footnoted otherwise.
Statement of Data Authenticity: Envirotech, Inc, attests the data reported has not been altered in any way.
Partial or incomplete reproduction of this report is prohibited, unless approved by Envirotech, Inc.
Envirotech, Inc, currently holds the appropriate and available Utah TNI certification NM009792018-1 for the data reported.

Alamosa Drilling
 8150 N. Central Expway, Suite 750
 Dallas TX, 75206

Project Name: Chavez #1 Spill
 Project Number: 18076-0001
 Project Manager: Felipe Aragon

Reported:
 09/19/18 15:48

Analytical Report for Samples

| Client Sample ID | Lab Sample ID | Matrix | Sampled | Received | Container |
|------------------|---------------|---------|----------|----------|---------------------|
| Excavation | P809029-01A | Aqueous | 09/13/18 | 09/14/18 | VOA Vial, 40mL; HCl |
| | P809029-01B | Aqueous | 09/13/18 | 09/14/18 | VOA Vial, 40mL; HCl |
| | P809029-01C | Aqueous | 09/13/18 | 09/14/18 | VOA Vial, 40mL; HCl |
| | P809029-01D | Aqueous | 09/13/18 | 09/14/18 | Poly 500mL |
| Water Well | P809029-02A | Aqueous | 09/13/18 | 09/14/18 | VOA Vial, 40mL; HCl |
| | P809029-02B | Aqueous | 09/13/18 | 09/14/18 | VOA Vial, 40mL; HCl |
| | P809029-02C | Aqueous | 09/13/18 | 09/14/18 | VOA Vial, 40mL; HCl |
| | P809029-02D | Aqueous | 09/13/18 | 09/14/18 | Poly 500mL |
| Navajo River | P809029-03A | Aqueous | 09/13/18 | 09/14/18 | VOA Vial, 40mL; HCl |
| | P809029-03B | Aqueous | 09/13/18 | 09/14/18 | VOA Vial, 40mL; HCl |
| | P809029-03C | Aqueous | 09/13/18 | 09/14/18 | VOA Vial, 40mL; HCl |
| | P809029-03D | Aqueous | 09/13/18 | 09/14/18 | Poly 500mL |
| Trip Blank | P809029-04A | Aqueous | 09/13/18 | 09/14/18 | VOA Vial, 40mL; HCl |
| | P809029-04B | Aqueous | 09/13/18 | 09/14/18 | VOA Vial, 40mL; HCl |

Partial or incomplete reproduction of this report is prohibited, unless approved by Envirotech, Inc.

Alamosa Drilling
 8150 N. Central Expway, Suite 750
 Dallas TX, 75206

Project Name: Chavez #1 Spill
 Project Number: 18076-0001
 Project Manager: Felipe Aragon

Reported:
 09/19/18 15:48

**Excavation
 P809029-01 (Water)**

| Reporting | | | | | | | | | |
|-----------|--------|-------|-------|----------|-------|----------|----------|--------|-------|
| Analyte | Result | Limit | Units | Dilution | Batch | Prepared | Analyzed | Method | Notes |

Wet Chem/Gravimetric

| | | | | | | | | | |
|------------------------|------------|------|------|---|---------|----------|----------|---------|--|
| Total Dissolved Solids | 240 | 10.0 | mg/L | 1 | 1837024 | 09/17/18 | 09/19/18 | SM2540C | |
|------------------------|------------|------|------|---|---------|----------|----------|---------|--|

Volatile Organics by EPA 8021

| | | | | | | | | | |
|---------------|-------------|------|------|---|---------|----------|----------|-----------|--|
| Benzene | 1.55 | 1.00 | ug/L | 1 | 1838008 | 09/19/18 | 09/19/18 | EPA 8021B | |
| Toluene | 1.99 | 1.00 | ug/L | 1 | 1838008 | 09/19/18 | 09/19/18 | EPA 8021B | |
| Ethylbenzene | 10.5 | 1.00 | ug/L | 1 | 1838008 | 09/19/18 | 09/19/18 | EPA 8021B | |
| p,m-Xylene | 90.5 | 2.00 | ug/L | 1 | 1838008 | 09/19/18 | 09/19/18 | EPA 8021B | |
| o-Xylene | ND | 1.00 | ug/L | 1 | 1838008 | 09/19/18 | 09/19/18 | EPA 8021B | |
| Total Xylenes | 90.5 | 1.00 | ug/L | 1 | 1838008 | 09/19/18 | 09/19/18 | EPA 8021B | |
| Total BTEX | 105 | 1.00 | ug/L | 1 | 1838008 | 09/19/18 | 09/19/18 | EPA 8021B | |

Surrogate: 4-Bromochlorobenzene-PID 97.2 % 50-150 1838008 09/19/18 09/19/18 EPA 8021B

Anions by 300.0/9056A

| | | | | | | | | | |
|----------|-------------|------|------|---|---------|----------|----------|--------------------|--|
| Chloride | 2.78 | 2.00 | mg/L | 1 | 1838004 | 09/17/18 | 09/17/18 | EPA 300.0/9056A | |
| Sulfate | 44.4 | 2.00 | mg/L | 1 | 1838004 | 09/17/18 | 09/17/18 | EPA 300.0/9056A | |

Partial or incomplete reproduction of this report is prohibited, unless approved by Envirotech, Inc.

Alamosa Drilling
8150 N. Central Expway, Suite 750
Dallas TX, 75206

Project Name: Chavez #1 Spill
Project Number: 18076-0001
Project Manager: Felipe Aragon

Reported:
09/19/18 15:48

Water Well
P809029-02 (Water)

| Reporting | | | | | | | | | |
|-----------|--------|-------|-------|----------|-------|----------|----------|--------|-------|
| Analyte | Result | Limit | Units | Dilution | Batch | Prepared | Analyzed | Method | Notes |

Wet Chem/Gravimetric

| | | | | | | | | | |
|------------------------|------------|------|------|---|---------|----------|----------|---------|--|
| Total Dissolved Solids | 239 | 10.0 | mg/L | 1 | 1837024 | 09/17/18 | 09/19/18 | SM2540C | |
|------------------------|------------|------|------|---|---------|----------|----------|---------|--|

Volatile Organics by EPA 8021

| | | | | | | | | | |
|---------------|----|------|------|---|---------|----------|----------|-----------|--|
| Benzene | ND | 1.00 | ug/L | 1 | 1838008 | 09/18/18 | 09/18/18 | EPA 8021B | |
| Toluene | ND | 1.00 | ug/L | 1 | 1838008 | 09/18/18 | 09/18/18 | EPA 8021B | |
| Ethylbenzene | ND | 1.00 | ug/L | 1 | 1838008 | 09/18/18 | 09/18/18 | EPA 8021B | |
| p,m-Xylene | ND | 2.00 | ug/L | 1 | 1838008 | 09/18/18 | 09/18/18 | EPA 8021B | |
| o-Xylene | ND | 1.00 | ug/L | 1 | 1838008 | 09/18/18 | 09/18/18 | EPA 8021B | |
| Total Xylenes | ND | 1.00 | ug/L | 1 | 1838008 | 09/18/18 | 09/18/18 | EPA 8021B | |
| Total BTEX | ND | 1.00 | ug/L | 1 | 1838008 | 09/18/18 | 09/18/18 | EPA 8021B | |

Surrogate: 4-Bromochlorobenzene-PID 99.6 % 50-150 1838008 09/18/18 09/18/18 EPA 8021B

Anions by 300.0/9056A

| | | | | | | | | | |
|----------|-------------|------|------|---|---------|----------|----------|--------------------|--|
| Chloride | ND | 2.00 | mg/L | 1 | 1838004 | 09/17/18 | 09/17/18 | EPA 300.0/9056A | |
| Sulfate | 38.5 | 2.00 | mg/L | 1 | 1838004 | 09/17/18 | 09/17/18 | EPA 300.0/9056A | |

Partial or incomplete reproduction of this report is prohibited, unless approved by Envirotech, Inc.

Alamosa Drilling
 8150 N. Central Expway, Suite 750
 Dallas TX, 75206

 Project Name: Chavez #1 Spill
 Project Number: 18076-0001
 Project Manager: Felipe Aragon

Reported:
 09/19/18 15:48

Navajo River
P809029-03 (Water)

| Reporting | | | | | | | | | |
|-----------|--------|-------|-------|----------|-------|----------|----------|--------|-------|
| Analyte | Result | Limit | Units | Dilution | Batch | Prepared | Analyzed | Method | Notes |

Wet Chem/Gravimetric

| | | | | | | | | | |
|------------------------|------------|------|------|---|---------|----------|----------|---------|--|
| Total Dissolved Solids | 224 | 10.0 | mg/L | 1 | 1837024 | 09/17/18 | 09/19/18 | SM2540C | |
|------------------------|------------|------|------|---|---------|----------|----------|---------|--|

Volatile Organics by EPA 8021

| | | | | | | | | | |
|---------------|----|------|------|---|---------|----------|----------|-----------|--|
| Benzene | ND | 1.00 | ug/L | 1 | 1838008 | 09/18/18 | 09/18/18 | EPA 8021B | |
| Toluene | ND | 1.00 | ug/L | 1 | 1838008 | 09/18/18 | 09/18/18 | EPA 8021B | |
| Ethylbenzene | ND | 1.00 | ug/L | 1 | 1838008 | 09/18/18 | 09/18/18 | EPA 8021B | |
| p,m-Xylene | ND | 2.00 | ug/L | 1 | 1838008 | 09/18/18 | 09/18/18 | EPA 8021B | |
| o-Xylene | ND | 1.00 | ug/L | 1 | 1838008 | 09/18/18 | 09/18/18 | EPA 8021B | |
| Total Xylenes | ND | 1.00 | ug/L | 1 | 1838008 | 09/18/18 | 09/18/18 | EPA 8021B | |
| Total BTEX | ND | 1.00 | ug/L | 1 | 1838008 | 09/18/18 | 09/18/18 | EPA 8021B | |

| | | | | | | |
|-------------------------------------|--------|--------|---------|----------|----------|-----------|
| Surrogate: 4-Bromochlorobenzene-PID | 99.9 % | 50-150 | 1838008 | 09/18/18 | 09/18/18 | EPA 8021B |
|-------------------------------------|--------|--------|---------|----------|----------|-----------|

Anions by 300.0/9056A

| | | | | | | | | | |
|----------|-------------|------|------|---|---------|----------|----------|--------------------|--|
| Chloride | ND | 2.00 | mg/L | 1 | 1838004 | 09/17/18 | 09/17/18 | EPA 300.0/9056A | |
| Sulfate | 48.9 | 2.00 | mg/L | 1 | 1838004 | 09/17/18 | 09/17/18 | EPA 300.0/9056A | |

Partial or incomplete reproduction of this report is prohibited, unless approved by Envirotech, Inc.

Alamosa Drilling
 8150 N. Central Expway, Suite 750
 Dallas TX, 75206

Project Name: Chavez #1 Spill
 Project Number: 18076-0001
 Project Manager: Felipe Aragon

Reported:
 09/19/18 15:48

Trip Blank
P809029-04 (Water)

| Reporting | | | | | | | | | |
|--|--------|--------|-------|----------|---------|----------|----------|-----------|-------|
| Analyte | Result | Limit | Units | Dilution | Batch | Prepared | Analyzed | Method | Notes |
| Volatile Organics by EPA 8021 | | | | | | | | | |
| Benzene | ND | 1.00 | ug/L | 1 | 1838008 | 09/18/18 | 09/18/18 | EPA 8021B | |
| Toluene | ND | 1.00 | ug/L | 1 | 1838008 | 09/18/18 | 09/18/18 | EPA 8021B | |
| Ethylbenzene | ND | 1.00 | ug/L | 1 | 1838008 | 09/18/18 | 09/18/18 | EPA 8021B | |
| p,m-Xylene | ND | 2.00 | ug/L | 1 | 1838008 | 09/18/18 | 09/18/18 | EPA 8021B | |
| o-Xylene | ND | 1.00 | ug/L | 1 | 1838008 | 09/18/18 | 09/18/18 | EPA 8021B | |
| Total Xylenes | ND | 1.00 | ug/L | 1 | 1838008 | 09/18/18 | 09/18/18 | EPA 8021B | |
| Total BTEX | ND | 1.00 | ug/L | 1 | 1838008 | 09/18/18 | 09/18/18 | EPA 8021B | |
| <i>Surrogate: 4-Bromochlorobenzene-PID</i> | | 98.3 % | | 50-150 | 1838008 | 09/18/18 | 09/18/18 | EPA 8021B | |

Partial or incomplete reproduction of this report is prohibited, unless approved by Envirotech, Inc.

Alamosa Drilling
 8150 N. Central Expway, Suite 750
 Dallas TX, 75206

Project Name: Chavez #1 Spill
 Project Number: 18076-0001
 Project Manager: Felipe Aragon

Reported:
 09/19/18 15:48

Wet Chem/Gravimetric - Quality Control

Envirotech Analytical Laboratory

| Analyte | Result | Reporting Limit | Units | Spike Level | Source Result | %REC | %REC Limits | RPD | RPD Limit | Notes |
|---------|--------|--------------------|-------|----------------|------------------|------|----------------|-----|--------------|-------|
|---------|--------|--------------------|-------|----------------|------------------|------|----------------|-----|--------------|-------|

Batch 1837024 - Wet Chemistry Preparation

Blank (1837024-BLK1)

Prepared: 09/13/18 1 Analyzed: 09/17/18 1

| | | | |
|------------------------|----|------|------|
| Total Dissolved Solids | ND | 10.0 | mg/L |
|------------------------|----|------|------|

LCS (1837024-BS1)

Prepared: 09/13/18 1 Analyzed: 09/17/18 1

| | | | | | | |
|------------------------|-----|------|------|-----|-----|--------|
| Total Dissolved Solids | 110 | 10.0 | mg/L | 100 | 110 | 50-150 |
|------------------------|-----|------|------|-----|-----|--------|

Duplicate (1837024-DUP1)

Source: P809017-01

Prepared: 09/13/18 1 Analyzed: 09/17/18 1

| | | | | | | |
|------------------------|-----|------|------|-----|------|---|
| Total Dissolved Solids | 423 | 10.0 | mg/L | 430 | 1.64 | 5 |
|------------------------|-----|------|------|-----|------|---|

Partial or incomplete reproduction of this report is prohibited, unless approved by Envirotech, Inc.

Alamosa Drilling
 8150 N. Central Expway, Suite 750
 Dallas TX, 75206

 Project Name: Chavez #1 Spill
 Project Number: 18076-0001
 Project Manager: Felipe Aragon

Reported:
 09/19/18 15:48

Volatile Organics by EPA 8021 - Quality Control
Envirotech Analytical Laboratory

| Analyte | Result | Reporting Limit | Units | Spike Level | Source Result | %REC | %REC Limits | RPD | RPD Limit | Notes |
|---------|--------|-----------------|-------|-------------|---------------|------|-------------|-----|-----------|-------|
|---------|--------|-----------------|-------|-------------|---------------|------|-------------|-----|-----------|-------|

Batch 1838008 - Purge and Trap EPA 5030A
Blank (1838008-BLK1)

Prepared & Analyzed: 09/18/18 1

| | | | | | | | | | | |
|-------------------------------------|-----|------|------|-----|--|------|--------|--|--|--|
| Benzene | ND | 1.00 | ug/L | | | | | | | |
| Toluene | ND | 1.00 | " | | | | | | | |
| Ethylbenzene | ND | 1.00 | " | | | | | | | |
| p,m-Xylene | ND | 2.00 | " | | | | | | | |
| o-Xylene | ND | 1.00 | " | | | | | | | |
| Total Xylenes | ND | 1.00 | " | | | | | | | |
| Total BTEX | ND | 1.00 | " | | | | | | | |
| Surrogate: 4-Bromochlorobenzene-PID | 154 | | " | 160 | | 96.4 | 50-150 | | | |

LCS (1838008-BS1)

Prepared & Analyzed: 09/18/18 1

| | | | | | | | | | | |
|-------------------------------------|------|------|------|-----|--|------|--------|--|--|--|
| Benzene | 91.7 | 1.00 | ug/L | 100 | | 91.8 | 70-130 | | | |
| Toluene | 94.3 | 1.00 | " | 100 | | 94.3 | 70-130 | | | |
| Ethylbenzene | 96.0 | 1.00 | " | 100 | | 96.1 | 70-130 | | | |
| p,m-Xylene | 198 | 2.00 | " | 200 | | 98.8 | 70-130 | | | |
| o-Xylene | 95.1 | 1.00 | " | 100 | | 95.1 | 70-130 | | | |
| Total Xylenes | 293 | 1.00 | " | 300 | | 97.6 | 70-130 | | | |
| Surrogate: 4-Bromochlorobenzene-PID | 157 | | " | 160 | | 98.4 | 50-150 | | | |

Matrix Spike (1838008-MS1)

Source: P809029-03

Prepared & Analyzed: 09/18/18 1

| | | | | | | | | | | |
|-------------------------------------|------|------|------|------|----|------|----------|--|--|--|
| Benzene | 487 | 5.00 | ug/L | 500 | ND | 97.5 | 54.3-133 | | | |
| Toluene | 500 | 5.00 | " | 500 | ND | 100 | 61.4-130 | | | |
| Ethylbenzene | 509 | 5.00 | " | 500 | ND | 102 | 61.4-133 | | | |
| p,m-Xylene | 1040 | 10.0 | " | 1000 | ND | 104 | 63.3-131 | | | |
| o-Xylene | 504 | 5.00 | " | 500 | ND | 101 | 63.3-131 | | | |
| Total Xylenes | 1550 | 5.00 | " | 1500 | ND | 103 | 63.3-131 | | | |
| Surrogate: 4-Bromochlorobenzene-PID | 798 | | " | 800 | | 99.8 | 50-150 | | | |

Matrix Spike Dup (1838008-MSD1)

Source: P809029-03

Prepared & Analyzed: 09/18/18 1

| | | | | | | | | | | |
|-------------------------------------|------|------|------|------|----|------|----------|------|----|--|
| Benzene | 498 | 5.00 | ug/L | 500 | ND | 99.6 | 54.3-133 | 2.07 | 20 | |
| Toluene | 511 | 5.00 | " | 500 | ND | 102 | 61.4-130 | 2.15 | 20 | |
| Ethylbenzene | 520 | 5.00 | " | 500 | ND | 104 | 61.4-133 | 2.07 | 20 | |
| p,m-Xylene | 1070 | 10.0 | " | 1000 | ND | 107 | 63.3-131 | 2.18 | 20 | |
| o-Xylene | 516 | 5.00 | " | 500 | ND | 103 | 63.3-131 | 2.26 | 20 | |
| Total Xylenes | 1580 | 5.00 | " | 1500 | ND | 106 | 63.3-131 | 2.21 | 20 | |
| Surrogate: 4-Bromochlorobenzene-PID | 799 | | " | 800 | | 99.9 | 50-150 | | | |

Partial or incomplete reproduction of this report is prohibited, unless approved by Envirotech, Inc.

Alamosa Drilling
 8150 N. Central Expway, Suite 750
 Dallas TX, 75206

Project Name: Chavez #1 Spill
 Project Number: 18076-0001
 Project Manager: Felipe Aragon

Reported:
 09/19/18 15:48

Anions by 300.0/9056A - Quality Control

Envirotech Analytical Laboratory

| Analyte | Result | Reporting Limit | Units | Spike Level | Source Result | %REC | %REC Limits | RPD | RPD Limit | Notes |
|---------|--------|--------------------|-------|----------------|------------------|------|----------------|-----|--------------|-------|
|---------|--------|--------------------|-------|----------------|------------------|------|----------------|-----|--------------|-------|

Batch 1838004 - Anion Extraction EPA 300.0/9056A

Blank (1838004-BLK1)

Prepared & Analyzed: 09/17/18 1

| | | | | | | | | | | |
|----------|----|------|------|--|--|--|--|--|--|--|
| Chloride | ND | 2.00 | mg/L | | | | | | | |
| Sulfate | ND | 2.00 | " | | | | | | | |

LCS (1838004-BS1)

Prepared & Analyzed: 09/17/18 1

| | | | | | | | | | | |
|----------|------|------|------|------|--|-----|--------|--|--|--|
| Chloride | 25.6 | 2.00 | mg/L | 25.0 | | 102 | 90-110 | | | |
| Sulfate | 25.7 | 2.00 | " | 25.0 | | 103 | 90-110 | | | |

Matrix Spike (1838004-MS1)

Source: P809029-01

Prepared & Analyzed: 09/17/18 1

| | | | | | | | | | | |
|----------|------|------|------|------|------|------|--------|--|--|--|
| Chloride | 28.1 | 2.00 | mg/L | 25.0 | 2.78 | 101 | 80-120 | | | |
| Sulfate | 69.0 | 2.00 | " | 25.0 | 44.4 | 98.5 | 80-120 | | | |

Matrix Spike Dup (1838004-MSD1)

Source: P809029-01

Prepared & Analyzed: 09/17/18 1

| | | | | | | | | | | |
|----------|------|------|------|------|------|------|--------|-------|----|--|
| Chloride | 28.3 | 2.00 | mg/L | 25.0 | 2.78 | 102 | 80-120 | 0.766 | 20 | |
| Sulfate | 69.2 | 2.00 | " | 25.0 | 44.4 | 99.3 | 80-120 | 0.265 | 20 | |

Partial or incomplete reproduction of this report is prohibited, unless approved by Envirotech, Inc.

Alamosa Drilling
8150 N. Central Expway, Suite 750
Dallas TX, 75206

Project Name: Chavez #1 Spill
Project Number: 18076-0001
Project Manager: Felipe Aragon

Reported:
09/19/18 15:48

Notes and Definitions

DET Analyte DETECTED
ND Analyte NOT DETECTED at or above the reporting limit
NR Not Reported
RPD Relative Percent Difference
** Methods marked with ** are non-accredited methods.

Partial or incomplete reproduction of this report is prohibited, unless approved by Envirotech, Inc.

| | | | | | | | | | | | | | | | | | | | | | | |
|---|--------------|--------|---------------|------------------|------------|--|--|---------------------|-----------------|-------------------|-------------|-------------|----------------|-------------|--------|-------------|--|----------------------------|----|----|----|----|
| Client: <u>Alamosa Drilling</u> | | | | Report Attention | | | | Lab Use Only | | | | TAT | | EPA Program | | | | | | | | |
| Project: <u>Chavez #1 Spill</u> | | | | Report due by: | | | | Lab WO# | | Job Number | | 1D | 3D | RCRA | CWA | SDWA | | | | | | |
| Project Manager: <u>Jason</u> | | | | Attention: | | | | <u>P809029</u> | | <u>18070-0001</u> | | | | | | | | | | | | |
| Address: | | | | Address: | | | | Analysis and Method | | | | | | | | State | | | | | | |
| City, State, Zip | | | | City, State, Zip | | | | DRO/ORO by 8015 | GRO/DRO by 8015 | BTEX by 8021 | VOC by 8260 | Metals 6010 | Chloride 300.0 | TPH 418.1 | LOG CC | 910-1 table | | | NM | CO | UT | AZ |
| Phone: | | | | Phone: | | | | | | | | | | | | | | | | | | |
| Email: <u>Brian Carter GCS@btcc.com</u> | | | | Email: | | | | | | | | | | | | | | | | | | |
| Time Sampled | Date Sampled | Matrix | No Containers | Sample ID | Lab Number | | | | | | | | | | | | | Remarks | | | | |
| 1515 | 9/13/18 | A | 4 | Excavation | 1 | | | | | | | | | X | | | | 3-VOA w/ HCL 1-500 poly | | | | |
| 1530 | 9/13/18 | A | 4 | Water Well | 2 | | | | | | | | | X | | | | | | | | |
| 1545 | 9/13/18 | A | 4 | Nazko River | 3 | | | | | | | | | X | | | | | | | | |
| | | A | 2 | Top Blank | 4 | | | | | | | | | X | | | | | | | | |
| | | | | | | | | | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | | | | | | | | | |

Additional Instructions:

vis ice in cooler

I, (field sampler), attest to the validity and authenticity of this sample. I am aware that tampering with or intentionally mislabelling the sample location, date or time of collection is considered fraud and may be grounds for legal action. Sampled by: Brian Hall

Samples requiring thermal preservation must be received on ice the day they are sampled or received packed in ice at an avg temp above 0 but less than 6°C on subsequent days.

| | | | | | | |
|------------------------------|---------|-------|--------------------------|---------|------|--|
| Relinquished by: (Signature) | Date | Time | Received by: (Signature) | Date | Time | Lab Use Only Received on ice: <u>Y</u> / <u>N</u> T1 _____ T2 _____ T3 _____ AVG Temp °C <u>4</u> |
| <u>Brian Hall</u> | 9/13/18 | 18:26 | <u>Danone Carter</u> | 9/14/18 | 7:30 | |
| Relinquished by: (Signature) | Date | Time | Received by: (Signature) | Date | Time | |
| <u>Danone Carter</u> | 9/14/18 | 8:43 | <u>Ali W</u> | 9/14/18 | 8:44 | |

Sample Matrix: S - Soil, Sd - Solid, Sg - Sludge, A - Aqueous, O - Other

Container Type: g - glass, p - poly/plastic, ag - amber glass, v - VOA

Note: Samples are discarded 30 days after results are reported unless other arrangements are made. Hazardous samples will be returned to client or disposed of at the client expense. The report for the analysis of the above samples is applicable only to those samples received by the laboratory with this COC. The liability of the laboratory is limited to the amount paid for on the report.