

# Session Report

2/15/2019

## Information Panel

Name S018\_BGN010013\_14022019\_185032  
Start Time 2/13/2019 1:10:04 PM  
Stop Time 2/14/2019 12:02:35 PM  
Device Name BGN010013  
Model Type SoundPro DL  
Device Firmware Rev R.13H  
Comments See attachment for weather data

## Calibration History

<u>Date</u>	<u>Calibration Action</u>	<u>Level</u>	<u>Cal. Model Type</u>	<u>Serial Number</u>	<u>Cert. Due Date</u>
2/13/2019 1:09:43 PM	Calibration	114.0			
2/14/2019 12:03:11 PM	Verification	114.4			

## Summary Data Panel

<u>Description</u>	<u>Meter</u>	<u>Value</u>	<u>Description</u>	<u>Meter</u>	<u>Value</u>
Leq	1	55.1 dB			
Exchange Rate	1	3 dB	Weighting	1	A
Response	1	SLOW	Bandwidth	1	OFF
Leq	2	71.6 dB			
Exchange Rate	2	3 dB	Weighting	2	C
Response	2	FAST			

## Information Panel

Name S018\_BGN010013\_14022019\_185032  
Start Time 2/13/2019 1:10:04 PM  
Stop Time 2/14/2019 12:02:35 PM  
Device Name BGN010013  
Model Type SoundPro DL  
Device Firmware Rev R.13H  
Comments See attachment for weather data

## Logged Data Chart

S018\_BGN010013\_14022019\_185032: Logged Data Chart

