

Session Report


2/6/2020

Information Panel

Name S017_BGN010014_06022020_130405
Start Time 1/31/2020 2:18:58 PM
Stop Time 2/3/2020 2:44:12 PM
Device Name BGN010014
Model Type SoundPro DL
Device Firmware Rev R.13F
Comments

Logged Data Chart

S017_BGN010014_06022020_130405: Logged Data Chart



Summary Data Panel

Description	Meter	Value	Description	Meter	Value
Leq	1	52 dB			
Exchange Rate	1	3 dB	Weighting	1	A
Response	1	SLOW	Bandwidth	1	OFF
Leq	2	71.9 dB			
Exchange Rate	2	3 dB	Weighting	2	C
Response	2	FAST			

Calibration History

Date	Calibration Action	Level	Cal. Model Type	Serial Number	Cert. Due Date
1/31/2020 2:18:26 PM	Calibration	114.0			
2/3/2020 2:46:49 PM	Verification	114.4			