

# Session Report

8/13/2021

## Information Panel


Name S056\_BGN010014\_13082021\_095627  
Start Time 8/10/2021 8:27:58 AM  
Stop Time 8/13/2021 7:32:49 AM  
Device Name BGN010014  
Model Type SoundPro DL  
Device Firmware Rev R.13F  
Comments

## Summary Data Panel

Description	Meter	Value	Description	Meter	Value
Leq	1	60.9 dB			
Exchange Rate	1	3 dB	Weighting	1	A
Response	1	SLOW	Bandwidth	1	OFF
Leq	2	76 dB			
Exchange Rate	2	3 dB	Weighting	2	C
Response	2	SLOW			

## Logged Data Chart

S056\_BGN010014\_13082021\_095627: Logged Data Chart



## Calibration History

Date	Calibration Action	Level	Cal. Model Type	Serial Number	Cert. Due Date
8/10/2021 8:27:15 AM	Calibration	114.0			
8/13/2021 7:33:20 AM	Verification	113.3			