

Session Report

8/13/2021

Information Panel


Name S056_BGN010014_13082021_095627
Start Time 8/10/2021 8:27:58 AM
Stop Time 8/13/2021 7:32:49 AM
Device Name BGN010014
Model Type SoundPro DL
Device Firmware Rev R.13F
Comments

Summary Data Panel

Description	Meter	Value	Description	Meter	Value
Leq	1	60.9 dB			
Exchange Rate	1	3 dB	Weighting	1	A
Response	1	SLOW	Bandwidth	1	OFF
Leq	2	76 dB			
Exchange Rate	2	3 dB	Weighting	2	C
Response	2	SLOW			

Logged Data Chart

S056_BGN010014_13082021_095627: Logged Data Chart



Calibration History

Date	Calibration Action	Level	Cal. Model Type	Serial Number	Cert. Due Date
8/10/2021 8:27:15 AM	Calibration	114.0			
8/13/2021 7:33:20 AM	Verification	113.3			