

# Session Report

1/25/2022

## Information Panel


Name S072\_BGN010014\_25012022\_084410  
Start Time 1/24/2022 8:31:55 AM  
Stop Time 1/25/2022 8:36:48 AM  
Device Name BGN010014  
Model Type SoundPro DL  
Device Firmware Rev R.13F  
Comments

## Summary Data Panel

Description	Meter	Value	Description	Meter	Value
Leq	1	62.5 dB			
Exchange Rate	1	3 dB	Weighting	1	A
Response	1	SLOW	Bandwidth	1	OFF
Leq	2	75.4 dB			
Exchange Rate	2	3 dB	Weighting	2	C
Response	2	SLOW			

## Logged Data Chart

S072\_BGN010014\_25012022\_084410: Logged Data Chart



## Calibration History

Date	Calibration Action	Level	Cal. Model Type	Serial Number	Cert. Due Date
1/24/2022 8:31:06 AM	Calibration	114.0			
1/25/2022 8:38:03 AM	Verification	114.2			