

Own Resources Operating, LLC
36695 Hwy 385, PO Box 250
Wray, CO 80758
(970) 332-3585

Calibration Report

Meter SN	4007	Prior Calibration Date	5/14/2021
Station Name	Vonderwahl 03-02	Date	11/9/2022
Site	Schramm	Time	12:42 PM
COC/COA (if applicable)		Legals	NESW/4 Sec. 02- T1N-R47W

Meter Test							
Make	Barton	Meter Run Size		2.067			
Diff Range	10	20	50	100	150	200	
Static Range	50	100	200	250	500	1000	1500 2000
Chart Rotation	24 hr	8 day	31 day				
Diff. Zero	Found	0 Left		0			
Diff. Reading	Found	12 Left		12			
Meter In Calibration	Yes	No	Notes				
Static Recording PSI	Found	21 Left		21			
HG Condition							
Pen Arc	OK	Reset	Time Lag	OK	Reset		
Leaks Found	Yes	No	Notes				
Friction Foun	Yes	No	Notes				
Chart Scale	Linear	Sq. Root					
Zero Set	Gauge	Absolute					

Orifice Plate Inspection					
Inspected	Yes	No	Clean	Yes	No
Sharp	Yes	No	Damaged	Yes	No
Seal Ring	Good	Poor			
Orifice Plate Removed	Size	Time			
Orifice Plate Installed	Size	Time			
Plate Size	0.25				
Orifice Fitting Type	SR.	Simplex	Jr	Flange	
Orifice Fitting Condition	Good	Poor			

Thermometer			
Type	Record	Indicate	Dial Stick
Make	Barton	202A	
Range	0-150		
Temp Test	Found	50 Left	50
Adjust	Yes	No	

Comments

DP FOUND				DP LEFT (leave blank if no changes made)			
Up		Down		Up		Down	
WC	Meter	WC	Meter	WC	Meter	WC	Meter
0	0	100	100				
20	20	70	70				
40	40	50	50				
60	60	30	30				
80	80	10	10				
100	100	0	0				

Static Spring Calibration							
Found				Left (leave blank if no changes made)			
DW		Meter		DW		Meter	
0	0	250	250				
125	125	75	75				
250	250	0	0				

Brent Jones
20

Witness

Differential Adj.
12 3.464102
12 3.464102
0
0
0

Static Adjustment
21 4.582576
21 4.582576
0
0
0