



Calibration Information	
Gas	Time
Pure N <sub>2</sub>	1143
Fresh Air	1146
50 ppm CO, 10 ppm H <sub>2</sub> S	1148
15% CH <sub>4</sub> , 15% CO <sub>2</sub> , Bal N <sub>2</sub>	1151

Site Sketch

SVP03

SVP01

SVP05

SVP04

SVP02

Bump Test						
Time		CH <sub>4</sub> (%)	CO <sub>2</sub> (%)	O <sub>2</sub> (%)	H <sub>2</sub> S (ppm)	CO (ppm)
1153	Calibration Gas	15.0	15.0	0.0	0	0
	Readings	14.8	14.7	0.2	0	1

[illegible]

Notes and Observations	