

**TECHNOLOGY
LABORATORY, INC.**

1012 Centre Avenue
Fort Collins, Colorado 80526
www.techlabusa.com
info@techlabusa.com

November 1, 2023

Fritzler Resources
300 E Burlington Ave
Fort Morgan, CO 80701

Case Narrative for Project No.: Scott Fed

Laboratory ID: C5951


To Whom It May Concern:

Case Narrative:

Please note that the sample on the attached report was analyzed outside the 48-hour holding time for nitrate, nitrite, and ortho-phosphate because the samples were brought to the lab outside the holding time. However, the sample was analyzed as soon as possible once brought to the lab.

Thank you for choosing TLI for your analysis needs. Please feel free to contact our laboratory manager Katrina Alsum at 970-490-1414 if you have any questions or concerns regarding the data in this report.

Sincerely,



Katrina L. Alsum
Laboratory Director
Technology Laboratory, Inc.



TECHNOLOGY LABORATORY, INC.

CENTRE PROFESSIONAL PARK

1012 Centre Avenue
Fort Collins, Colorado 80526
(970) 490-1414

CERTIFICATE OF ANALYSIS

Fritzler Resources
300 E Burlington Ave
Fort Morgan, CO 80701

Sample ID: Scott Fed

Laboratory ID: C5951-01

Sampled: 10/24/23

Received: 10/27/23

Project No.: Scott Fed

Matrix: Water

<u>CAS Number</u>	<u>Parameter</u>	<u>Result</u>	<u>Units</u>	<u>MDL</u>	<u>Method</u>	<u>Date Analyzed</u>
71-43-2	Benzene	< 0.001	mg/L	0.001	EPA-8260B	11/02/23
108-88-3	Toluene	< 0.001	mg/L	0.001	EPA-8260B	11/02/23
100-41-4	Ethylbenzene	< 0.001	mg/L	0.001	EPA-8260B	11/02/23
1330-20-7	Total Xylenes	0.006	mg/L	0.001	EPA-8260B	11/02/23
N/A	GRO (TVPH)	3.49	mg/L	0.5	EPA-8260B	11/02/23

QA/QC SURROGATE RECOVERIES

<u>Compound</u>	<u>% Recovery</u>	<u>% Rec. Limits</u>
Dibromofluoromethane	102	68-120
Toluene-d8	102	81-128
Bromofluorobenzene	101	70-113

Katrina L. Alsum

Katrina L. Alsum
Laboratory Director
Technology Laboratory, Inc.

The results contained in this report
relate only to those items tested.



TECHNOLOGY LABORATORY, INC.

CENTRE PROFESSIONAL PARK

1012 Centre Avenue
Fort Collins, Colorado 80526
(970) 490-1414

CERTIFICATE OF ANALYSIS

Fritzler Resources
300 E Burlington Ave
Fort Morgan, CO 80701

Sample ID: Scott Fed

Laboratory ID: C5951-01

Sampled: 10/24/23

Received: 10/27/23

Project No.: Scott Fed

Matrix: Water

CAS Number	Parameter	Result	Units	MDL	Method	Date Analyzed
N/A	Nitrate+Nitrite as N (Calc)	< 0.05	mg/L	0.05	Calculation	10/30/23
7440-39-3	Total Barium	0.095	mg/L	0.08	EPA-6010B	11/03/23
7440-42-8	Total Boron	1.08	mg/L	0.1	EPA-6010B	11/03/23
7440-70-2	Total Calcium	1.42	mg/L	0.5	EPA-6010B	11/03/23
7439-89-6	Total Iron	0.065	mg/L	0.02	EPA-6010B	11/03/23
7439-95-4	Total Magnesium	< 2.0	mg/L	2	EPA-6010B	11/03/23
7439-96-5	Total Manganese	< 0.09	mg/L	0.09	EPA-6010B	11/03/23
7440-09-7	Total Potassium	3.38	mg/L	1	EPA-6010B	11/03/23
7440-23-5	Total Sodium	405	mg/L	3	EPA-6010B	11/03/23
7440-24-6	Total Strontium	< 0.07	mg/L	0.07	EPA-6010B	11/03/23
7782-49-2	Total Selenium	< 0.09	mg/L	0.09	EPA-6010B	11/03/23
N/A	DRO (TEPH)	65.2	mg/L	0.5	EPA-8015B	11/03/23
N/A	Specific Conductance	1222	µohms/cm @ 25°C		EPA-9050A	11/01/23
7726-95-6	Bromide	0.192	mg/L	0.05	EPA-9056A	10/27/23
16887-00-6	Chloride	177	mg/L	1	EPA-9056A	10/27/23
16984-48-8	Fluoride	6.31	mg/L	0.05	EPA-9056A	10/27/23
84145-82-4	Nitrate-N	0.161	mg/L	0.1	EPA-9056A	10/27/23
14797-65-0	Nitrite-N	< 0.05	mg/L	0.05	EPA-9056A	10/27/23
18785-72-3	Sulfate	15.2	mg/L	0.05	EPA-9056A	10/27/23
N/A	Total Dissolved Solids (TDS)	1220	mg/L	5	SM-2540 C	11/01/23
N/A	TSS	97	mg/L	5	SM-2540 D	11/01/23
98059-61-1	Phosphate Ortho	< 0.05	mg/L	0.05	SM-4110	10/27/23
N/A	pH	7.85	Units		SM-4500 HB	11/01/23
N/A	Nitrate+Nitrite as N	1.26	mg/L	0.3	SM-4500-NO3F	11/07/23

Katrina L. Alsum

Katrina L. Alsum
Laboratory Director
Technology Laboratory, Inc.

The results contained in this report
relate only to those items tested.



15657

④

CHAIN-OF-CUSTODY REPORT

[illegible]