

This review was performed with guidance from the National Functional Guidelines for Organic Superfund Methods Data Review (US EPA, 2020, US EPA) and/or the National Functional Guidelines for Inorganic Superfund Methods Data Review (US EPA, 2020, US EPA). These validation guidance documents specifically address analyses performed in accordance with the CLP analytical methods and are not completely applicable to the type of analyses and analytical protocols performed for the Standard Method (SM), SW-846, and/or US EPA methods utilized by the laboratory for these samples. Environmental Standards, Inc. (Environmental Standards) used professional judgment to determine the quality of the analytical results and compliance relative to the Standard Method (SM), SW-846, and/or US EPA utilized by the laboratory. This QA review was performed on the data associated with Sample Delivery Group (SDG):

### L1851298

The findings offered in this report are based on a review of the Chain-of-Custody Record and Case Narrative, sample preservation and condition upon laboratory receipt, holding times, surrogate recovery, field and laboratory blank results, laboratory and field duplicate precision, laboratory control sample / laboratory control sample duplicate recoveries and precision, matrix spike / matrix spike duplicate recoveries and precision, total and dissolved results comparisons, and/or percent solids (as applicable). All review items may not have been included in this SDG; therefore, only those items included in this SDG were addressed in the QA review.

A summary of the results of the data review process is provided below:

Sample	Sample Type	Method	Analyte	T/D	Result	Qual	Reason Code(s)	MDL	QL	Unit	Detect?
GACO0423SDW001	N	CALC	Total Nitrogen	N	1990	J	CR	79.4	100	ug/L	Y
GACO0423SDW001	N	E351.2	Kjeldahl Nitrogen, TKN	N	1990	J+	BL	131	250	ug/L	Y
GACO0423SDW001	N	SW6020	Copper,Dissolved	D	4.19	J	RL	0.700	5.00	ug/L	Y
GACO0423SDW001	N	SW6020	Nickel,Dissolved	D	0.710	J	RL	0.500	2.00	ug/L	Y

#### Data Qualifiers

U	The analyte was analyzed for, but was not detected above the level of the adjusted detection limit or quantitation limit, as appropriate, or was observed in a blank at a similar level.
R	The data are unusable. The sample results are rejected due to serious deficiencies in meeting QC criteria. The analyte may or may not be present in the sample.
J	The result is an estimated quantity. The associated numerical value is the approximate concentration of the analyte in the sample.
J+	The result is an estimated quantity, but the result may be biased high.
J-	The result is an estimated quantity, but the result may be biased low.
UJ	The analyte was analyzed for but was not detected. The reported quantitation limit is approximate and may be inaccurate or imprecise.

#### Reason Codes and Explanations

BF	Contamination present in a field blank (e.g ., Field Blank, Equipment Blank, etc .); evaluation criteria exceeded
BL	Contamination present in a laboratory blank (e.g ., Method Blank, Instrument Blank, etc .); evaluation criteria exceeded
BT	Contamination present in the Trip Blank; evaluation criteria exceeded
CC	Possible contamination due to carryover from a previous sample
CR	Calculated result in which one or more of the components has been qualified
CRQ	Calculated result flagged due to reporting protocol
CT	Cooler temperature criteria not met
CY	Chemical Yield recovery criteria not met
EC	Result exceeds the calibration range; potential bias indeterminate
FD	Field duplicate imprecision; potential bias indeterminate
GH	Headspace present in the gamma spectrometer sample analysis vessel; potential bias indeterminate
GS	Low sample density in the gamma spectrometer sample analysis vessel; potential bias indeterminate
HT	Holding time exceeded
HV	Headspace present in volatile vials
IN	Interference (e.g ., laboratory, chemical, chromatographic/instrumental, and/or matrix) present in the analysis

LC	Laboratory control sample/laboratory control sample duplicate recovery criteria not met
LCP	Laboratory control sample/laboratory control sample duplicate precision criteria not met; potential bias indeterminate
LD	Laboratory duplicate precision criteria not met; potential bias indeterminate
MDP	Laboratory deviated from the method for a method-defined parameter, based on regulatory requirements
MS	Matrix spike/matrix spike duplicate recovery criteria not met
MSP	Matrix spike/matrix spike duplicate precision criteria not met; potential bias indeterminate
PD	Post-digestion spike recovery criteria not met
OT	Other deficiencies, see report for additional details
PS	Low percent solids; potential bias indeterminate
RA	Replicate/multiple analyses criteria not met; potential bias indeterminate
RL	The analysis meets all qualitative identification criteria, but the measured concentration is between the method detection limit and the quantitation or reporting limit; potential bias indeterminate
RS	Reporting limit standard(s) outside of acceptance limits
SC	Relative percent difference between two columns exceeds criteria; potential bias indeterminate
SP	Sample preservation criteria not met
SR	Surrogate recovery criteria not met
ST	Sample container type incorrect
SU	Sample result is less than the two-sigma uncertainty
SUN	Absolute value of the negative sample result is greater than the two-sigma uncertainty
SW	Sample switch suspected
TD	Result for dissolved constituent significantly exceeded result for total constituent; potential bias indeterminate
TIC	Tentatively identified compound, quantified using an assumed calibration factor; potential bias indeterminate

Lab Sample ID	L1851298-01
Sys Sample Code	GACO0423SDW001
Sample Name	GACO0423SDW001
Sample Date	4/23/2025 11:55:00 AM
Sample Type	N
Matrix	SW
Parent Sample	
% Moisture	

Analytic Method	Chemical Name	CAS Rn	Fraction	Test Type	Result Unit	Final Result	Final Qual	Reason code	Final MDL	Final RL	Final QL	Final Detect	Final Report	DF	Basis
5310-B	TOC (Total Organic Carbon)	10-35-5	N	INITIAL	ug/L	14100			495	495	1000	Y	Y	1	NA
CALC	Total Nitrogen	TN	N	INITIAL	ug/L	1990	J	CR	79.4	79.4	100	Y	Y	1	NA
E130.1	Hardness (colorimetric) as CaCO3	10-92-4	N	INITIAL	ug/L	530000			53000	53000	150000	Y	Y	5	NA
E300.0	Bromide	24959-67-9	N	INITIAL	ug/L		U		680	680	1000	N	Y	1	NA
	Chloride	16887-00-6	N	DILUTION	ug/L	116000			5470	5470	10000	Y	Y	10	NA
	Fluoride	16984-48-8	N	INITIAL	ug/L	938			76.1	76.1	150	Y	Y	1	NA
	Nitrate as (N)	14797-55-8	N	INITIAL	ug/L		U		88.4	88.4	100	N	Y	1	NA
	Nitrite as (N)	14797-65-0	N	INITIAL	ug/L		U		79.4	79.4	100	N	Y	1	NA
	Sulfate	14808-79-8	N	DILUTION	ug/L	932000			6370	6370	50000	Y	Y	10	NA
E351.2	Kjeldahl Nitrogen, TKN	7727-37-	N	INITIAL	ug/L	1990	J+	BL	131	131	250	Y	Y	1	NA
E365.4	Phosphorus, Total	7723-14-0	N	INITIAL	ug/L		U		64.2	64.2	100	N	Y	1	NA
SM 2540C	Dissolved Solids	10-33-3	N	INITIAL	ug/L	1560000			25000	25000	25000	Y	Y	1	NA
SM2320B	Alkalinity	TotAlk	N	INITIAL	ug/L	94700			4750	4750	20000	Y	Y	1	NA
SM2540D	Suspended Solids	10-32-2	N	INITIAL	ug/L	10700			283	283	2500	Y	Y	1	NA
SM5540C	MBAS	MBAS	N	INITIAL	ug/L		U		19.0	19.0	100	N	Y	1	NA
SW6020	Cadmium, Dissolved	7440-43-9	D	REANALYSIS	ug/L		U		0.120	0.120	1.00	N	Y	1	NA
	Calcium	7440-70-2	T	INITIAL	ug/L	103000			92.5	92.5	1000	Y	Y	1	NA
	Copper, Dissolved	7440-50-8	D	REANALYSIS	ug/L	4.19	J	RL	0.700	0.700	5.00	Y	Y	1	NA
	Iron	7439-89-6	T	INITIAL	ug/L		U		22.6	22.6	100	N	Y	1	NA
	Lead, Dissolved	7439-92-1	D	REANALYSIS	ug/L		U		0.500	0.500	2.00	N	Y	1	NA
	Magnesium	7439-95-4	T	INITIAL	ug/L	69800			82.7	82.7	1000	Y	Y	1	NA
	Manganese	7439-96-5	T	INITIAL	ug/L	35.1			0.700	0.700	5.00	Y	Y	1	NA
	Manganese, Dissolved	7439-96-5	D	REANALYSIS	ug/L	37.4			0.700	0.700	5.00	Y	Y	1	NA
	Nickel, Dissolved	7440-02-0	D	REANALYSIS	ug/L	0.710	J	RL	0.500	0.500	2.00	Y	Y	1	NA
	Potassium	7440-09-7	T	INITIAL	ug/L	15500			96.5	96.5	2000	Y	Y	1	NA
	Sodium	7440-23-5	T	INITIAL	ug/L	263000			142	142	2000	Y	Y	1	NA
	Zinc, Dissolved	7440-66-6	D	REANALYSIS	ug/L		U		4.00	4.00	25.0	N	Y	1	NA
SW7199	Hexavalent Chromium	18540-29-9	N	INITIAL	ug/L		U		0.100	0.100	5.00	N	Y	1	NA