

Stage 2 Data Validation Memorandum
Chevron Bishop Loss of Containment Response Site
Galeton, Colorado
Solid Samples
Sample Delivery Group: L1853224
Report Date: May 22, 2025

This quality assurance (QA) review is based upon an examination of the data generated from the analyses of the six solid samples and associated quality control (QC) samples (including aqueous blanks) collected on April 29, 2025, at the Chevron Bishop Loss of Containment Response Site in Galeton, Colorado. These samples were collectively analyzed by Pace Analytical National Center for Testing and Innovation (Pace National) of Mount Juliet, Tennessee, for volatile organic compounds (VOCs) by SW-846 Method 8260D, semivolatile organic compounds (SVOCs) by SW-846 Method 8270E, total metals by SW-846 Method 6010D, nitrate-nitrite by SW-846 Method 9056A, ammonia nitrogen by US EPA Method 350.1, total nitrogen by calculation, total Kjeldahl nitrogen (TKN) by Standard Method (SM) 4500-NORG-D, and total organic carbon (TOC) by American Society of Agronomy (ASA) Walkley-Black method.

This review was performed in accordance with the Bishop Loss of Containment, Galeton, Colorado Environmental Sampling and Analysis Plan (CTEH; Version 1.4, May 7, 2025), the Bishop Loss of Containment Incident Draft Quality Assurance Project Plan (QAPP; Environmental Standards, Inc. [Environmental Standards]; Version 1.0, April 25, 2025), and the above-referenced analytical methods. This review was performed with guidance from the National Functional Guidelines for Organic Superfund Methods Data Review (US EPA, 2020) and the National Functional Guidelines for Inorganic Superfund Methods Data Review (US EPA, 2020). These validation guidance documents specifically address analyses performed in accordance with the CLP analytical methods and are not completely applicable to the type of analyses and analytical protocols performed for the SM, SW-846, US EPA, and ASA methods utilized by the laboratory for these samples. Environmental Standards used professional judgment to determine the quality of the analytical results and compliance relative to the SM, SW-846, US EPA, and ASA methods utilized by the laboratory.

Summary

The analytical results and associated laboratory QC samples were reviewed to determine the integrity of the reported analytical results and to ensure that the data met the established measurement quality objectives. This QA review includes all samples in Pace National Sample Delivery Group (SDG) L1853224.

The samples that have undergone Stage 2 data validation are listed below:

Sample Identification	Laboratory Sample Identification	Laboratory SDG	Matrix	Date Sample Collected	Parameter(s) Examined
GACO0429T000S004	L1853224-01	L1853224	Solid	4/29/25	VOC, SVOC, M, N+N, NH ₃ , TN, TKN, TOC
GACO0429T000S005	L1853224-02	L1853224	Solid	4/29/25	VOC, SVOC, M, N+N, NH ₃ , TN, TKN, TOC
GACO0429T000S006	L1853224-03	L1853224	Solid	4/29/25	VOC, SVOC, M, N+N, NH ₃ , TN, TKN, TOC
GACO0429T000S007	L1853224-04	L1853224	Solid	4/29/25	VOC, SVOC, M, N+N, NH ₃ , TN, TKN, TOC
GACO0429T000T002 (Trip Blank)	L1853224-05	L1853224	Aq	4/29/25	VOC
GACO0429T000T005 (Trip Blank)	L1853224-06	L1853224	Aq	4/29/25	VOC

Notes:

VOC - VOCs by SW-846 Method 8260D.
 SVOC - SVOCs by SW-846 Method 8270E.
 M - Total Metals by SW-846 Method 6010D.
 N+N - Nitrate-Nitrite by SW-846 Method 9056A.
 NH₃ - Ammonia Nitrogen by US EPA Method 350.1.
 TN - Total Nitrogen by Calculation.
 TKN - TKN by SM 4500-NORG-D.
 TOC - TOC by ASA Walkley-Black Method.
 Aq - Aqueous.

ITEMS REVIEWED

Chain-of-Custody (COC) Record and Case Narrative	Sample Preservation and Condition Upon Laboratory Receipt
Holding Times	Surrogate Recovery
Blank Results	Percent Solids
Laboratory Control Sample and Laboratory Control Sample Duplicate (LCS/LCSD) Results	Results Reported Between the Method Detection Limit (MDL)/Reporting Limit (RL) and Quantitation Limit (QL)

Comments

1. The laboratory personnel/courier did not record the date and time of receipt of the samples on the first line of the COC Record.
2. The laboratory personnel/courier did not sign or record the date and time relinquished on the second line of the COC Record.

Based on the items included in this QA review, the following qualifiers are offered.

Analyte	Samples	Validation Qualifier	Reason for Qualification
nitrate-nitrite	all samples	U	BL
total nitrogen	all samples	J	CR

- For results qualified "not-detected" due to blank contamination, the MDLs/RLs and QLs (if the reported results exceeded the QL) have been elevated to the original laboratory-reported result, the laboratory-reported result has been removed from the result field, and the detect flag has been changed to "N."
- When sample results were qualified both as estimated with a direction of bias ("J+" or "J-") and as estimated with unknown bias ("J") or the opposite bias, only the unknown bias qualifier has been included on the data tables.

Review performed by:	Valerie J. Butterhof, Quality Assurance Chemist
Report reviewed by:	Jared K. Acker, CEAC, Senior Quality Assurance Chemist
Report approved by:	Amanda J. Cover, CEAC, Associate Chemist/Project Manager
Date review completed:	5/22/2025

SECTION 2

ANALYTICAL RESULTS

DATA QUALIFIERS

- U** The analyte was analyzed for, but was not detected above the level of the adjusted detection limit or quantitation limit, as appropriate.
- R** The data are unusable. The sample results are rejected due to serious deficiencies in meeting QC criteria. The analyte may or may not be present in the sample.
- J** The result is an estimated quantity. The associated numerical value is the approximate concentration of the analyte in the sample.
- J+** The result is an estimated quantity, but the result may be biased high.
- J-** The result is an estimated quantity, but the result may be biased low.
- UJ** The analyte was analyzed for but was not detected. The reported quantitation limit is approximate and may be inaccurate or imprecise.
- NJ** The analyte has been “tentatively identified” or “presumptively” as present and the associated numerical value is the estimated concentration in the sample.



REASON CODES AND EXPLANATIONS

Reason Code¹	Description
<i>¹ For any Reason Code that does not indicate that the potential bias is indeterminate, the “+” or “-” reason code may be appended to the qualification reason code in order to indicate a direction of bias (e.g., MS+ would be used to indicate potential high bias due to a high matrix spike recovery)</i>	
+	The associated quality control item indicates a potential high bias in the sample result
-	The associated quality control item indicates a potential low bias in the sample result
AST	Compound not quantitated against an authentic standard; potential bias indeterminate
BF	Contamination present in a field blank (e.g., Field Blank, Equipment Blank, etc.); evaluation criteria exceeded
BL	Contamination present in a laboratory blank (e.g., Method Blank, Instrument Blank, etc.); evaluation criteria exceeded
BN	Elevated detection limit or estimated result due to negative instrument drift (e.g., negative instrument blank result with an absolute value > 2× the method detection limit)
BT	Contamination present in the Trip Blank; evaluation criteria exceeded
CC	Possible contamination due to carryover from a previous sample
CR	Calculated result in which one or more of the components has been qualified
CRQ	Calculated result flagged due to reporting protocol
CT	Cooler temperature criteria not met
CV	Continuing calibration verification evaluation criteria not met
CY	Chemical Yield recovery criteria not met
DI	Detector instability (radionuclide chemistry); potential bias indeterminate
EC	Result exceeds the calibration range; potential bias indeterminate
FD	Field duplicate imprecision; potential bias indeterminate
FP	Target compound identification criteria not met; potential false positive
GH	Headspace present in the gamma spectrometer sample analysis vessel; potential bias indeterminate
GS	Low sample density in the gamma spectrometer sample analysis vessel; potential bias indeterminate
HT	Holding time exceeded
HV	Headspace present in volatile vials
IC	Initial calibration evaluation criteria not met

Reason Code ¹	Description
IN	Interference (e.g., laboratory, chemical, chromatographic/instrumental, and/or matrix) present in the analysis
IR	Interference check standard evaluation criteria not met
IS	Internal standard evaluation criteria not met
LC	Laboratory control sample/laboratory control sample duplicate recovery criteria not met
LCP	Laboratory control sample/laboratory control sample duplicate precision criteria not met; potential bias indeterminate
LD	Laboratory duplicate precision criteria not met; potential bias indeterminate
LR	Linear range exceeded; potential bias indeterminate
MDP	Laboratory deviated from the method for a method-defined parameter, based on regulatory requirements
MS	Matrix spike/matrix spike duplicate recovery criteria not met
MSP	Matrix spike/matrix spike duplicate precision criteria not met; potential bias indeterminate
NQC	Absence of supporting quality control samples
PD	Post-digestion spike recovery criteria not met
OT	Other deficiencies, see validation report for additional details
PM	Performance evaluation mixture criteria not met
PS	Low percent solids; potential bias indeterminate
PT	Chromatographic pattern in sample does not match pattern of calibration standard
QCI	Quantitation/confirmation ion ratios in sample are inconsistent with reference spectra; potential bias indeterminate
RA	Replicate/multiple analyses criteria not met; potential bias indeterminate
RM	Reference material recovery criteria not met
RL	The analysis meets all qualitative identification criteria, but the measured concentration is between the method detection limit and the quantitation or reporting limit; potential bias indeterminate
RS	Reporting limit standard(s) outside of acceptance limits
SA	Method of standard additions criteria not met; potential bias indeterminate
SC	Relative percent difference between two columns exceeds criteria; potential bias indeterminate
SCC	Second column confirmation was not performed as required by the analysis method

Reason Code ¹	Description
SCT	Sample counting time error (radionuclide chemistry); potential bias indeterminate
SD	Serial dilution results did not meet evaluation criteria
SP	Sample preservation criteria not met
SR	Surrogate recovery criteria not met
SS	Second source calibration verification/initial calibration verification criteria not met
ST	Sample container type incorrect
SU	Sample result is less than the two-sigma uncertainty
SUN	Absolute value of the negative sample result is greater than the two-sigma uncertainty
SW	Sample switch suspected
TD	Result for dissolved constituent significantly exceeded result for total constituent; potential bias indeterminate
TIR	Tentatively identified compound; observed in an associated laboratory, equipment, field, or trip blank.
TN	Instrument tune criteria not met
Y	Potential bias due to the y-intercept in the calibration curve significantly affecting the analyte response

Lab Sample ID	L1853224-01
Sys Sample Code	GACO0429T000S004
Sample Name	GACO0429T000S004
Sample Date	4/29/2025 10:35:00 AM
Sample Type	N
Matrix	SO
Parent Sample	
% Moisture	12.20

Analytic Method	Chemical Name	CAS Rn	Fraction	Test Type	Result Unit	Final Result	Final Qual	Reason code	Final MDL	Final RL	Final QL	Final Detect	Final Report	DF	Basis
CALC	Total Nitrogen	TN	N	INITIAL	ug/Kg	1270000	J	CR	690	690	22800	Y	Y	1	DRY
E350.1	Ammonia Nitrogen	7664-41-7	N	INITIAL	ug/Kg		U		8190	8190	11400	N	Y	1	DRY
SM2540G	Total Solids	10-31-1	N	INITIAL	%	87.8						Y	Y	1	NA
SM4500-NORG-D	Kjeldahl Nitrogen, TKN	7727-37-9TKN	N	INITIAL	ug/Kg	1240000			173000	173000	228000	Y	Y	10	DRY
SW6010	Aluminum	7429-90-5	T	INITIAL	ug/Kg	2210000			6930	6930	22800	Y	Y	1	DRY
	Antimony	7440-36-0	T	INITIAL	ug/Kg		U		787	787	2280	N	Y	1	DRY
	Beryllium	7440-41-7	T	INITIAL	ug/Kg	373			54.3	54.3	228	Y	Y	1	DRY
	Calcium	7440-70-2	T	INITIAL	ug/Kg	2700000			21600	21600	114000	Y	Y	1	DRY
	Cobalt	7440-48-4	T	INITIAL	ug/Kg	2850			202	202	1140	Y	Y	1	DRY
	Iron	7439-89-6	T	INITIAL	ug/Kg	3160000			2550	2550	11400	Y	Y	1	DRY
	Magnesium	7439-95-4	T	INITIAL	ug/Kg	1200000			22700	22700	114000	Y	Y	1	DRY
	Manganese	7439-96-5	T	INITIAL	ug/Kg	199000			197	197	1140	Y	Y	1	DRY
	Potassium	7440-09-7	T	INITIAL	ug/Kg	1360000			23800	23800	114000	Y	Y	1	DRY
	Sodium	7440-23-5	T	INITIAL	ug/Kg	143000			46900	46900	114000	Y	Y	1	DRY
	Thallium	7440-28-0	T	INITIAL	ug/Kg		U		590	590	2280	N	Y	1	DRY
	Vanadium	7440-62-2	T	INITIAL	ug/Kg	7780			436	436	2280	Y	Y	1	DRY
SW8260	1,1,1,2-Tetrachloroethane	630-20-6	N	INITIAL	ug/Kg		U		1.21	1.21	3.20	N	Y	1	DRY
	1,1,1-Trichloroethane	71-55-6	N	INITIAL	ug/Kg		U		1.18	1.18	3.20	N	Y	1	DRY
	1,1,2,2-Tetrachloroethane	79-34-5	N	INITIAL	ug/Kg		U		0.889	0.889	3.20	N	Y	1	DRY
	1,1,2-Trichloroethane	79-00-5	N	INITIAL	ug/Kg		U		0.764	0.764	3.20	N	Y	1	DRY
	1,1,2-Trichlorotrifluoroethane	76-13-1	N	INITIAL	ug/Kg		U		0.965	0.965	3.20	N	Y	1	DRY
	1,1-Dichloroethane	75-34-3	N	INITIAL	ug/Kg		U		0.628	0.628	3.20	N	Y	1	DRY
	1,1-Dichloroethene	75-35-4	N	INITIAL	ug/Kg		U		0.775	0.775	3.20	N	Y	1	DRY
	1,1-Dichloropropene	563-58-6	N	INITIAL	ug/Kg		U		1.03	1.03	3.20	N	Y	1	DRY
	1,2,3-Trichlorobenzene	87-61-6	N	INITIAL	ug/Kg		U		9.38	9.38	16.0	N	Y	1	DRY
	1,2,3-Trichloropropane	96-18-4	N	INITIAL	ug/Kg		U		2.07	2.07	16.0	N	Y	1	DRY
	1,2,3-Trimethylbenzene	526-73-8	N	INITIAL	ug/Kg		U		2.02	2.02	6.40	N	Y	1	DRY
	1,2,4-Trichlorobenzene	120-82-1	N	INITIAL	ug/Kg		U		5.63	5.63	16.0	N	Y	1	DRY
	1,2-Dibromo-3-Chloropropane	96-12-8	N	INITIAL	ug/Kg		U		4.99	4.99	32.0	N	Y	1	DRY
	1,2-Dibromoethane	106-93-4	N	INITIAL	ug/Kg		U		0.829	0.829	3.20	N	Y	1	DRY
	1,2-Dichlorobenzene	95-50-1	N	INITIAL	ug/Kg		U		0.544	0.544	6.40	N	Y	1	DRY
	1,2-Dichloroethane	107-06-2	N	INITIAL	ug/Kg		U		0.830	0.830	3.20	N	Y	1	DRY

Lab Sample ID	L1853224-01
Sys Sample Code	GACO0429T000S004
Sample Name	GACO0429T000S004
Sample Date	4/29/2025 10:35:00 AM
Sample Type	N
Matrix	SO
Parent Sample	
% Moisture	12.20

Analytic Method	Chemical Name	CAS Rn	Fraction	Test Type	Result Unit	Final Result	Final Qual	Reason code	Final MDL	Final RL	Final QL	Final Detect	Final Report	DF	Basis
SW8260	1,2-Dichloropropane	78-87-5	N	INITIAL	ug/Kg		U		1.82	1.82	6.40	N	Y	1	DRY
	1,3-Dichlorobenzene	541-73-1	N	INITIAL	ug/Kg		U		0.768	0.768	6.40	N	Y	1	DRY
	1,3-Dichloropropane	142-28-9	N	INITIAL	ug/Kg		U		0.641	0.641	6.40	N	Y	1	DRY
	1,4-Dichlorobenzene	106-46-7	N	INITIAL	ug/Kg		U		0.895	0.895	6.40	N	Y	1	DRY
	2,2-Dichloropropane	594-20-7	N	INITIAL	ug/Kg		U		1.77	1.77	3.20	N	Y	1	DRY
	2-Butanone (MEK)	78-93-3	N	INITIAL	ug/Kg		U		81.2	81.2	128	N	Y	1	DRY
	2-Chlorotoluene	95-49-8	N	INITIAL	ug/Kg		U		1.11	1.11	3.20	N	Y	1	DRY
	4-Chlorotoluene	106-43-4	N	INITIAL	ug/Kg		U		0.576	0.576	6.40	N	Y	1	DRY
	4-Methyl-2-pentanone (MIBK)	108-10-1	N	INITIAL	ug/Kg		U		2.92	2.92	32.0	N	Y	1	DRY
	Acetone	67-64-1	N	INITIAL	ug/Kg		U		46.7	46.7	64.0	N	Y	1	DRY
	Acrylonitrile	107-13-1	N	INITIAL	ug/Kg		U		4.62	4.62	16.0	N	Y	1	DRY
	Bromobenzene	108-86-1	N	INITIAL	ug/Kg		U		1.15	1.15	16.0	N	Y	1	DRY
	Bromodichloromethane	75-27-4	N	INITIAL	ug/Kg		U		0.927	0.927	3.20	N	Y	1	DRY
	Bromoform	75-25-2	N	INITIAL	ug/Kg		U		1.50	1.50	32.0	N	Y	1	DRY
	Bromomethane	74-83-9	N	INITIAL	ug/Kg		U		2.52	2.52	16.0	N	Y	1	DRY
	Carbon tetrachloride	56-23-5	N	INITIAL	ug/Kg		U		1.15	1.15	6.40	N	Y	1	DRY
	Chlorobenzene	108-90-7	N	INITIAL	ug/Kg		U		0.269	0.269	3.20	N	Y	1	DRY
	Chlorodibromomethane	124-48-1	N	INITIAL	ug/Kg		U		0.783	0.783	3.20	N	Y	1	DRY
	Chloroethane	75-00-3	N	INITIAL	ug/Kg		U		2.17	2.17	6.40	N	Y	1	DRY
	Chloroform	67-66-3	N	INITIAL	ug/Kg		U		1.32	1.32	3.20	N	Y	1	DRY
	Chloromethane	74-87-3	N	INITIAL	ug/Kg		U		5.56	5.56	16.0	N	Y	1	DRY
	cis-1,2-Dichloroethene	156-59-2	N	INITIAL	ug/Kg		U		0.939	0.939	3.20	N	Y	1	DRY
	cis-1,3-Dichloropropene	10061-01-5	N	INITIAL	ug/Kg		U		0.968	0.968	3.20	N	Y	1	DRY
	Dibromomethane	74-95-3	N	INITIAL	ug/Kg		U		0.959	0.959	6.40	N	Y	1	DRY
	Dichlorodifluoromethane	75-71-8	N	INITIAL	ug/Kg		U		2.06	2.06	6.40	N	Y	1	DRY
	Di-isopropyl ether	108-20-3	N	INITIAL	ug/Kg		U		0.524	0.524	1.28	N	Y	1	DRY
	Hexachloro-1,3-butadiene	87-68-3	N	INITIAL	ug/Kg		U		7.68	7.68	32.0	N	Y	1	DRY
	Isopropylbenzene	98-82-8	N	INITIAL	ug/Kg		U		0.544	0.544	3.20	N	Y	1	DRY
	Methyl tert-butyl ether	1634-04-4	N	INITIAL	ug/Kg		U		0.448	0.448	1.28	N	Y	1	DRY
	Methylene Chloride	75-09-2	N	INITIAL	ug/Kg		U		8.49	8.49	32.0	N	Y	1	DRY
	n-Butylbenzene	104-51-8	N	INITIAL	ug/Kg		U		6.72	6.72	16.0	N	Y	1	DRY
	n-Propylbenzene	103-65-1	N	INITIAL	ug/Kg		U		1.22	1.22	6.40	N	Y	1	DRY
	p-Isopropyltoluene	99-87-6	N	INITIAL	ug/Kg		U		3.26	3.26	6.40	N	Y	1	DRY

Lab Sample ID	L1853224-01
Sys Sample Code	GACO0429T000S004
Sample Name	GACO0429T000S004
Sample Date	4/29/2025 10:35:00 AM
Sample Type	N
Matrix	SO
Parent Sample	
% Moisture	12.20

Analytic Method	Chemical Name	CAS Rn	Fraction	Test Type	Result Unit	Final Result	Final Qual	Reason code	Final MDL	Final RL	Final QL	Final Detect	Final Report	DF	Basis
SW8260	sec-Butylbenzene	135-98-8	N	INITIAL	ug/Kg		U		3.68	3.68	16.0	N	Y	1	DRY
	Styrene	100-42-5	N	INITIAL	ug/Kg		U		0.293	0.293	16.0	N	Y	1	DRY
	tert-Butylbenzene	98-06-6	N	INITIAL	ug/Kg		U		2.49	2.49	6.40	N	Y	1	DRY
	Tetrachloroethene	127-18-4	N	INITIAL	ug/Kg		U		1.15	1.15	3.20	N	Y	1	DRY
	trans-1,2-Dichloroethene	156-60-5	N	INITIAL	ug/Kg		U		1.33	1.33	6.40	N	Y	1	DRY
	trans-1,3-Dichloropropene	10061-02-6	N	INITIAL	ug/Kg		U		1.46	1.46	6.40	N	Y	1	DRY
	Trichloroethene	79-01-6	N	INITIAL	ug/Kg		U		0.747	0.747	1.28	N	Y	1	DRY
	Trichlorofluoromethane	75-69-4	N	INITIAL	ug/Kg		U		1.06	1.06	3.20	N	Y	1	DRY
	Vinyl chloride	75-01-4	N	INITIAL	ug/Kg		U		1.48	1.48	3.20	N	Y	1	DRY
SW8270	1,2,4-Trichlorobenzene	120-82-1	N	INITIAL	ug/Kg		U		23.7	23.7	759	N	Y	2	DRY
	1,2-Dichlorobenzene	95-50-1	N	INITIAL	ug/Kg		U		22.4	22.4	759	N	Y	2	DRY
	1,3-Dichlorobenzene	541-73-1	N	INITIAL	ug/Kg		U		23.0	23.0	759	N	Y	2	DRY
	1,4-Dichlorobenzene	106-46-7	N	INITIAL	ug/Kg		U		22.6	22.6	759	N	Y	2	DRY
	2,2-Oxybis(1-Chloropropane)	108-60-1	N	INITIAL	ug/Kg		U		32.8	32.8	759	N	Y	2	DRY
	2,4,6-Trichlorophenol	88-06-2	N	INITIAL	ug/Kg		U		24.4	24.4	759	N	Y	2	DRY
	2,4-Dichlorophenol	120-83-2	N	INITIAL	ug/Kg		U		22.1	22.1	759	N	Y	2	DRY
	2,4-Dimethylphenol	105-67-9	N	INITIAL	ug/Kg		U		19.8	19.8	759	N	Y	2	DRY
	2,4-Dinitrophenol	51-28-5	N	INITIAL	ug/Kg		U		178	178	759	N	Y	2	DRY
	2,4-Dinitrotoluene	121-14-2	N	INITIAL	ug/Kg		U		21.8	21.8	759	N	Y	2	DRY
	2,6-Dinitrotoluene	606-20-2	N	INITIAL	ug/Kg		U		24.8	24.8	759	N	Y	2	DRY
	2-Chloronaphthalene	91-58-7	N	INITIAL	ug/Kg		U		13.3	13.3	75.9	N	Y	2	DRY
	2-Chlorophenol	95-57-8	N	INITIAL	ug/Kg		U		25.1	25.1	759	N	Y	2	DRY
	2-Nitrophenol	88-75-5	N	INITIAL	ug/Kg		U		27.1	27.1	759	N	Y	2	DRY
	3,3-Dichlorobenzidine	91-94-1	N	INITIAL	ug/Kg		U		28.0	28.0	759	N	Y	2	DRY
	4,6-Dinitro-2-methylphenol	534-52-1	N	INITIAL	ug/Kg		U		172	172	759	N	Y	2	DRY
	4-Bromophenyl-phenylether	101-55-3	N	INITIAL	ug/Kg		U		26.7	26.7	759	N	Y	2	DRY
	4-Chloro-3-methylphenol	59-50-7	N	INITIAL	ug/Kg		U		24.6	24.6	759	N	Y	2	DRY
	4-Chlorophenyl-phenylether	7005-72-3	N	INITIAL	ug/Kg		U		26.4	26.4	759	N	Y	2	DRY
	4-Nitrophenol	100-02-7	N	INITIAL	ug/Kg		U		23.7	23.7	759	N	Y	2	DRY
	Acenaphthylene	208-96-8	N	INITIAL	ug/Kg		U		10.7	10.7	75.9	N	Y	2	DRY
	Benzidine	92-87-5	N	INITIAL	ug/Kg		U		142	142	3810	N	Y	2	DRY
	Benzo(g,h,i)perylene	191-24-2	N	INITIAL	ug/Kg		U		13.9	13.9	75.9	N	Y	2	DRY
	Benzylbutyl phthalate	85-68-7	N	INITIAL	ug/Kg		U		23.7	23.7	759	N	Y	2	DRY

Lab Sample ID	L1853224-01
Sys Sample Code	GACO0429T000S004
Sample Name	GACO0429T000S004
Sample Date	4/29/2025 10:35:00 AM
Sample Type	N
Matrix	SO
Parent Sample	
% Moisture	12.20

Analytic Method	Chemical Name	CAS Rn	Fraction	Test Type	Result Unit	Final Result	Final Qual	Reason code	Final MDL	Final RL	Final QL	Final Detect	Final Report	DF	Basis
SW8270	Bis(2-chlorethoxy)methane	111-91-1	N	INITIAL	ug/Kg		U		22.8	22.8	759	N	Y	2	DRY
	Bis(2-chloroethyl)ether	111-44-4	N	INITIAL	ug/Kg		U		25.1	25.1	759	N	Y	2	DRY
	Bis(2-ethylhexyl)phthalate	117-81-7	N	INITIAL	ug/Kg		U		96.2	96.2	759	N	Y	2	DRY
	Diethyl phthalate	84-66-2	N	INITIAL	ug/Kg		U		25.1	25.1	759	N	Y	2	DRY
	Dimethyl phthalate	131-11-3	N	INITIAL	ug/Kg		U		161	161	759	N	Y	2	DRY
	Di-n-butyl phthalate	84-74-2	N	INITIAL	ug/Kg		U		26.0	26.0	759	N	Y	2	DRY
	Di-n-octyl phthalate	117-84-0	N	INITIAL	ug/Kg		U		51.3	51.3	759	N	Y	2	DRY
	Hexachloro-1,3-butadiene	87-68-3	N	INITIAL	ug/Kg		U		25.5	25.5	759	N	Y	2	DRY
	Hexachlorobenzene	118-74-1	N	INITIAL	ug/Kg		U		26.9	26.9	759	N	Y	2	DRY
	Hexachlorocyclopentadiene	77-47-4	N	INITIAL	ug/Kg		U		39.9	39.9	759	N	Y	2	DRY
	Hexachloroethane	67-72-1	N	INITIAL	ug/Kg		U		29.9	29.9	759	N	Y	2	DRY
	Isophorone	78-59-1	N	INITIAL	ug/Kg		U		23.2	23.2	759	N	Y	2	DRY
	Nitrobenzene	98-95-3	N	INITIAL	ug/Kg		U		26.4	26.4	759	N	Y	2	DRY
	n-Nitrosodimethylamine	62-75-9	N	INITIAL	ug/Kg		U		113	113	759	N	Y	2	DRY
	n-Nitrosodi-n-propylamine	621-64-7	N	INITIAL	ug/Kg		U		25.3	25.3	759	N	Y	2	DRY
	n-Nitrosodiphenylamine	86-30-6	N	INITIAL	ug/Kg		U		57.4	57.4	759	N	Y	2	DRY
	Pentachlorophenol	87-86-5	N	INITIAL	ug/Kg		U		20.4	20.4	759	N	Y	2	DRY
	Phenanthrene	85-01-8	N	INITIAL	ug/Kg		U		15.0	15.0	75.9	N	Y	2	DRY
	Phenol	108-95-2	N	INITIAL	ug/Kg		U		30.5	30.5	759	N	Y	2	DRY
SW9056	Nitrate-Nitrite	NO2-NO3	N	INITIAL	ug/Kg		U	BL	21800	21800	22800	N	Y	1	DRY
WBLACK	TOC By Walkley Black	10-35-5	N	INITIAL	ug/Kg	17500000			128000	128000	500000	Y	Y	5	NA

Lab Sample ID	L1853224-02
Sys Sample Code	GACO0429T000S005
Sample Name	GACO0429T000S005
Sample Date	4/29/2025 11:28:00 AM
Sample Type	N
Matrix	SO
Parent Sample	
% Moisture	9.37

Analytic Method	Chemical Name	CAS Rn	Fraction	Test Type	Result Unit	Final Result	Final Qual	Reason code	Final MDL	Final RL	Final QL	Final Detect	Final Report	DF	Basis
CALC	Total Nitrogen	TN	N	INITIAL	ug/Kg	1340000	J	CR	669	669	22100	Y	Y	1	DRY
E350.1	Ammonia Nitrogen	7664-41-7	N	INITIAL	ug/Kg		U		7930	7930	11000	N	Y	1	DRY
SM2540G	Total Solids	10-31-1	N	INITIAL	%	90.6						Y	Y	1	NA
SM4500-NORG-D	Kjeldahl Nitrogen, TKN	7727-37-9TKN	N	INITIAL	ug/Kg	1320000			168000	168000	221000	Y	Y	10	DRY
SW6010	Aluminum	7429-90-5	T	INITIAL	ug/Kg	2200000			6710	6710	22100	Y	Y	1	DRY
	Antimony	7440-36-0	T	INITIAL	ug/Kg		U		762	762	2210	N	Y	1	DRY
	Beryllium	7440-41-7	T	INITIAL	ug/Kg	381			52.6	52.6	221	Y	Y	1	DRY
	Calcium	7440-70-2	T	INITIAL	ug/Kg	13800000			21000	21000	110000	Y	Y	1	DRY
	Cobalt	7440-48-4	T	INITIAL	ug/Kg	2890			195	195	1100	Y	Y	1	DRY
	Iron	7439-89-6	T	INITIAL	ug/Kg	3060000			2470	2470	11000	Y	Y	1	DRY
	Magnesium	7439-95-4	T	INITIAL	ug/Kg	1740000			22000	22000	110000	Y	Y	1	DRY
	Manganese	7439-96-5	T	INITIAL	ug/Kg	176000			191	191	1100	Y	Y	1	DRY
	Potassium	7440-09-7	T	INITIAL	ug/Kg	1290000			23100	23100	110000	Y	Y	1	DRY
	Sodium	7440-23-5	T	INITIAL	ug/Kg	220000			45500	45500	110000	Y	Y	1	DRY
	Thallium	7440-28-0	T	INITIAL	ug/Kg		U		572	572	2210	N	Y	1	DRY
	Vanadium	7440-62-2	T	INITIAL	ug/Kg	8770			423	423	2210	Y	Y	1	DRY
SW8260	1,1,1,2-Tetrachloroethane	630-20-6	N	INITIAL	ug/Kg		U		1.14	1.14	3.02	N	Y	1	DRY
	1,1,1-Trichloroethane	71-55-6	N	INITIAL	ug/Kg		U		1.11	1.11	3.02	N	Y	1	DRY
	1,1,2,2-Tetrachloroethane	79-34-5	N	INITIAL	ug/Kg		U		0.839	0.839	3.02	N	Y	1	DRY
	1,1,2-Trichloroethane	79-00-5	N	INITIAL	ug/Kg		U		0.720	0.720	3.02	N	Y	1	DRY
	1,1,2-Trichlorotrifluoroethane	76-13-1	N	INITIAL	ug/Kg		U		0.910	0.910	3.02	N	Y	1	DRY
	1,1-Dichloroethane	75-34-3	N	INITIAL	ug/Kg		U		0.593	0.593	3.02	N	Y	1	DRY
	1,1-Dichloroethene	75-35-4	N	INITIAL	ug/Kg		U		0.731	0.731	3.02	N	Y	1	DRY
	1,1-Dichloropropene	563-58-6	N	INITIAL	ug/Kg		U		0.976	0.976	3.02	N	Y	1	DRY
	1,2,3-Trichlorobenzene	87-61-6	N	INITIAL	ug/Kg		U		8.85	8.85	15.1	N	Y	1	DRY
	1,2,3-Trichloropropane	96-18-4	N	INITIAL	ug/Kg		U		1.96	1.96	15.1	N	Y	1	DRY
	1,2,3-Trimethylbenzene	526-73-8	N	INITIAL	ug/Kg		U		1.91	1.91	6.03	N	Y	1	DRY
	1,2,4-Trichlorobenzene	120-82-1	N	INITIAL	ug/Kg		U		5.31	5.31	15.1	N	Y	1	DRY
	1,2-Dibromo-3-Chloropropane	96-12-8	N	INITIAL	ug/Kg		U		4.71	4.71	30.2	N	Y	1	DRY
	1,2-Dibromoethane	106-93-4	N	INITIAL	ug/Kg		U		0.782	0.782	3.02	N	Y	1	DRY
	1,2-Dichlorobenzene	95-50-1	N	INITIAL	ug/Kg		U		0.513	0.513	6.03	N	Y	1	DRY
	1,2-Dichloroethane	107-06-2	N	INITIAL	ug/Kg		U		0.783	0.783	3.02	N	Y	1	DRY

Lab Sample ID	L1853224-02
Sys Sample Code	GACO0429T000S005
Sample Name	GACO0429T000S005
Sample Date	4/29/2025 11:28:00 AM
Sample Type	N
Matrix	SO
Parent Sample	
% Moisture	9.37

Analytic Method	Chemical Name	CAS Rn	Fraction	Test Type	Result Unit	Final Result	Final Qual	Reason code	Final MDL	Final RL	Final QL	Final Detect	Final Report	DF	Basis
SW8260	1,2-Dichloropropane	78-87-5	N	INITIAL	ug/Kg		U		1.71	1.71	6.03	N	Y	1	DRY
	1,3-Dichlorobenzene	541-73-1	N	INITIAL	ug/Kg		U		0.724	0.724	6.03	N	Y	1	DRY
	1,3-Dichloropropane	142-28-9	N	INITIAL	ug/Kg		U		0.605	0.605	6.03	N	Y	1	DRY
	1,4-Dichlorobenzene	106-46-7	N	INITIAL	ug/Kg		U		0.845	0.845	6.03	N	Y	1	DRY
	2,2-Dichloropropane	594-20-7	N	INITIAL	ug/Kg		U		1.67	1.67	3.02	N	Y	1	DRY
	2-Butanone (MEK)	78-93-3	N	INITIAL	ug/Kg		U		76.6	76.6	121	N	Y	1	DRY
	2-Chlorotoluene	95-49-8	N	INITIAL	ug/Kg		U		1.04	1.04	3.02	N	Y	1	DRY
	4-Chlorotoluene	106-43-4	N	INITIAL	ug/Kg		U		0.543	0.543	6.03	N	Y	1	DRY
	4-Methyl-2-pentanone (MIBK)	108-10-1	N	INITIAL	ug/Kg		U		2.75	2.75	30.2	N	Y	1	DRY
	Acetone	67-64-1	N	INITIAL	ug/Kg		U		44.0	44.0	60.3	N	Y	1	DRY
	Acrylonitrile	107-13-1	N	INITIAL	ug/Kg		U		4.36	4.36	15.1	N	Y	1	DRY
	Bromobenzene	108-86-1	N	INITIAL	ug/Kg		U		1.09	1.09	15.1	N	Y	1	DRY
	Bromodichloromethane	75-27-4	N	INITIAL	ug/Kg		U		0.875	0.875	3.02	N	Y	1	DRY
	Bromoform	75-25-2	N	INITIAL	ug/Kg		U		1.41	1.41	30.2	N	Y	1	DRY
	Bromomethane	74-83-9	N	INITIAL	ug/Kg		U		2.38	2.38	15.1	N	Y	1	DRY
	Carbon tetrachloride	56-23-5	N	INITIAL	ug/Kg		U		1.08	1.08	6.03	N	Y	1	DRY
	Chlorobenzene	108-90-7	N	INITIAL	ug/Kg		U		0.253	0.253	3.02	N	Y	1	DRY
	Chlorodibromomethane	124-48-1	N	INITIAL	ug/Kg		U		0.739	0.739	3.02	N	Y	1	DRY
	Chloroethane	75-00-3	N	INITIAL	ug/Kg		U		2.05	2.05	6.03	N	Y	1	DRY
	Chloroform	67-66-3	N	INITIAL	ug/Kg		U		1.24	1.24	3.02	N	Y	1	DRY
	Chloromethane	74-87-3	N	INITIAL	ug/Kg		U		5.25	5.25	15.1	N	Y	1	DRY
	cis-1,2-Dichloroethene	156-59-2	N	INITIAL	ug/Kg		U		0.886	0.886	3.02	N	Y	1	DRY
	cis-1,3-Dichloropropene	10061-01-5	N	INITIAL	ug/Kg		U		0.914	0.914	3.02	N	Y	1	DRY
	Dibromomethane	74-95-3	N	INITIAL	ug/Kg		U		0.905	0.905	6.03	N	Y	1	DRY
	Dichlorodifluoromethane	75-71-8	N	INITIAL	ug/Kg		U		1.94	1.94	6.03	N	Y	1	DRY
	Di-isopropyl ether	108-20-3	N	INITIAL	ug/Kg		U		0.495	0.495	1.21	N	Y	1	DRY
	Hexachloro-1,3-butadiene	87-68-3	N	INITIAL	ug/Kg		U		7.24	7.24	30.2	N	Y	1	DRY
	Isopropylbenzene	98-82-8	N	INITIAL	ug/Kg		U		0.513	0.513	3.02	N	Y	1	DRY
	Methyl tert-butyl ether	1634-04-4	N	INITIAL	ug/Kg		U		0.422	0.422	1.21	N	Y	1	DRY
	Methylene Chloride	75-09-2	N	INITIAL	ug/Kg		U		8.01	8.01	30.2	N	Y	1	DRY
	n-Butylbenzene	104-51-8	N	INITIAL	ug/Kg		U		6.34	6.34	15.1	N	Y	1	DRY
	n-Propylbenzene	103-65-1	N	INITIAL	ug/Kg		U		1.15	1.15	6.03	N	Y	1	DRY
	p-Isopropyltoluene	99-87-6	N	INITIAL	ug/Kg		U		3.08	3.08	6.03	N	Y	1	DRY

Lab Sample ID	L1853224-02
Sys Sample Code	GACO0429T000S005
Sample Name	GACO0429T000S005
Sample Date	4/29/2025 11:28:00 AM
Sample Type	N
Matrix	SO
Parent Sample	
% Moisture	9.37

Analytic Method	Chemical Name	CAS Rn	Fraction	Test Type	Result Unit	Final Result	Final Qual	Reason code	Final MDL	Final RL	Final QL	Final Detect	Final Report	DF	Basis
SW8260	sec-Butylbenzene	135-98-8	N	INITIAL	ug/Kg		U		3.48	3.48	15.1	N	Y	1	DRY
	Styrene	100-42-5	N	INITIAL	ug/Kg		U		0.276	0.276	15.1	N	Y	1	DRY
	tert-Butylbenzene	98-06-6	N	INITIAL	ug/Kg		U		2.35	2.35	6.03	N	Y	1	DRY
	Tetrachloroethene	127-18-4	N	INITIAL	ug/Kg		U		1.08	1.08	3.02	N	Y	1	DRY
	trans-1,2-Dichloroethene	156-60-5	N	INITIAL	ug/Kg		U		1.26	1.26	6.03	N	Y	1	DRY
	trans-1,3-Dichloropropene	10061-02-6	N	INITIAL	ug/Kg		U		1.38	1.38	6.03	N	Y	1	DRY
	Trichloroethene	79-01-6	N	INITIAL	ug/Kg		U		0.705	0.705	1.21	N	Y	1	DRY
	Trichlorofluoromethane	75-69-4	N	INITIAL	ug/Kg		U		0.998	0.998	3.02	N	Y	1	DRY
	Vinyl chloride	75-01-4	N	INITIAL	ug/Kg		U		1.40	1.40	3.02	N	Y	1	DRY
SW8270	1,2,4-Trichlorobenzene	120-82-1	N	INITIAL	ug/Kg		U		11.5	11.5	367	N	Y	1	DRY
	1,2-Dichlorobenzene	95-50-1	N	INITIAL	ug/Kg		U		10.9	10.9	367	N	Y	1	DRY
	1,3-Dichlorobenzene	541-73-1	N	INITIAL	ug/Kg		U		11.1	11.1	367	N	Y	1	DRY
	1,4-Dichlorobenzene	106-46-7	N	INITIAL	ug/Kg		U		10.9	10.9	367	N	Y	1	DRY
	2,2-Oxybis(1-Chloropropane)	108-60-1	N	INITIAL	ug/Kg		U		15.9	15.9	367	N	Y	1	DRY
	2,4,6-Trichlorophenol	88-06-2	N	INITIAL	ug/Kg		U		11.8	11.8	367	N	Y	1	DRY
	2,4-Dichlorophenol	120-83-2	N	INITIAL	ug/Kg		U		10.7	10.7	367	N	Y	1	DRY
	2,4-Dimethylphenol	105-67-9	N	INITIAL	ug/Kg		U		9.60	9.60	367	N	Y	1	DRY
	2,4-Dinitrophenol	51-28-5	N	INITIAL	ug/Kg		U		86.0	86.0	367	N	Y	1	DRY
	2,4-Dinitrotoluene	121-14-2	N	INITIAL	ug/Kg		U		10.5	10.5	367	N	Y	1	DRY
	2,6-Dinitrotoluene	606-20-2	N	INITIAL	ug/Kg		U		12.0	12.0	367	N	Y	1	DRY
	2-Chloronaphthalene	91-58-7	N	INITIAL	ug/Kg		U		6.45	6.45	36.7	N	Y	1	DRY
	2-Chlorophenol	95-57-8	N	INITIAL	ug/Kg		U		12.1	12.1	367	N	Y	1	DRY
	2-Nitrophenol	88-75-5	N	INITIAL	ug/Kg		U		13.1	13.1	367	N	Y	1	DRY
	3,3-Dichlorobenzidine	91-94-1	N	INITIAL	ug/Kg		U		13.6	13.6	367	N	Y	1	DRY
	4,6-Dinitro-2-methylphenol	534-52-1	N	INITIAL	ug/Kg		U		83.3	83.3	367	N	Y	1	DRY
	4-Bromophenyl-phenylether	101-55-3	N	INITIAL	ug/Kg		U		12.9	12.9	367	N	Y	1	DRY
	4-Chloro-3-methylphenol	59-50-7	N	INITIAL	ug/Kg		U		11.9	11.9	367	N	Y	1	DRY
	4-Chlorophenyl-phenylether	7005-72-3	N	INITIAL	ug/Kg		U		12.8	12.8	367	N	Y	1	DRY
	4-Nitrophenol	100-02-7	N	INITIAL	ug/Kg		U		11.5	11.5	367	N	Y	1	DRY
	Acenaphthylene	208-96-8	N	INITIAL	ug/Kg		U		5.18	5.18	36.7	N	Y	1	DRY
	Benzidine	92-87-5	N	INITIAL	ug/Kg		U		69.1	69.1	1840	N	Y	1	DRY
	Benzo(g,h,i)perylene	191-24-2	N	INITIAL	ug/Kg		U		6.72	6.72	36.7	N	Y	1	DRY
	Benzylbutyl phthalate	85-68-7	N	INITIAL	ug/Kg		U		11.5	11.5	367	N	Y	1	DRY

Lab Sample ID	L1853224-02
Sys Sample Code	GACO0429T000S005
Sample Name	GACO0429T000S005
Sample Date	4/29/2025 11:28:00 AM
Sample Type	N
Matrix	SO
Parent Sample	
% Moisture	9.37

Analytic Method	Chemical Name	CAS Rn	Fraction	Test Type	Result Unit	Final Result	Final Qual	Reason code	Final MDL	Final RL	Final QL	Final Detect	Final Report	DF	Basis
SW8270	Bis(2-chlorethoxy)methane	111-91-1	N	INITIAL	ug/Kg		U		11.0	11.0	367	N	Y	1	DRY
	Bis(2-chloroethyl)ether	111-44-4	N	INITIAL	ug/Kg		U		12.1	12.1	367	N	Y	1	DRY
	Bis(2-ethylhexyl)phthalate	117-81-7	N	INITIAL	ug/Kg		U		46.6	46.6	367	N	Y	1	DRY
	Diethyl phthalate	84-66-2	N	INITIAL	ug/Kg		U		12.1	12.1	367	N	Y	1	DRY
	Dimethyl phthalate	131-11-3	N	INITIAL	ug/Kg		U		77.9	77.9	367	N	Y	1	DRY
	Di-n-butyl phthalate	84-74-2	N	INITIAL	ug/Kg		U		12.6	12.6	367	N	Y	1	DRY
	Di-n-octyl phthalate	117-84-0	N	INITIAL	ug/Kg		U		24.8	24.8	367	N	Y	1	DRY
	Hexachloro-1,3-butadiene	87-68-3	N	INITIAL	ug/Kg		U		12.4	12.4	367	N	Y	1	DRY
	Hexachlorobenzene	118-74-1	N	INITIAL	ug/Kg		U		13.0	13.0	367	N	Y	1	DRY
	Hexachlorocyclopentadiene	77-47-4	N	INITIAL	ug/Kg		U		19.3	19.3	367	N	Y	1	DRY
	Hexachloroethane	67-72-1	N	INITIAL	ug/Kg		U		14.5	14.5	367	N	Y	1	DRY
	Isophorone	78-59-1	N	INITIAL	ug/Kg		U		11.3	11.3	367	N	Y	1	DRY
	Nitrobenzene	98-95-3	N	INITIAL	ug/Kg		U		12.8	12.8	367	N	Y	1	DRY
	n-Nitrosodimethylamine	62-75-9	N	INITIAL	ug/Kg		U		54.5	54.5	367	N	Y	1	DRY
	n-Nitrosodi-n-propylamine	621-64-7	N	INITIAL	ug/Kg		U		12.2	12.2	367	N	Y	1	DRY
	n-Nitrosodiphenylamine	86-30-6	N	INITIAL	ug/Kg		U		27.8	27.8	367	N	Y	1	DRY
	Pentachlorophenol	87-86-5	N	INITIAL	ug/Kg		U		9.89	9.89	367	N	Y	1	DRY
	Phenanthrene	85-01-8	N	INITIAL	ug/Kg		U		7.29	7.29	36.7	N	Y	1	DRY
	Phenol	108-95-2	N	INITIAL	ug/Kg		U		14.8	14.8	367	N	Y	1	DRY
SW9056	Nitrate-Nitrite	NO2-NO3	N	INITIAL	ug/Kg		U	BL	21000	21000	22100	N	Y	1	DRY
WBLACK	TOC By Walkley Black	10-35-5	N	INITIAL	ug/Kg	18400000			128000	128000	500000	Y	Y	5	NA

Lab Sample ID	L1853224-03
Sys Sample Code	GACO0429T000S006
Sample Name	GACO0429T000S006
Sample Date	4/29/2025 11:06:00 AM
Sample Type	N
Matrix	SO
Parent Sample	
% Moisture	15.30

Analytic Method	Chemical Name	CAS Rn	Fraction	Test Type	Result Unit	Final Result	Final Qual	Reason code	Final MDL	Final RL	Final QL	Final Detect	Final Report	DF	Basis
CALC	Total Nitrogen	TN	N	INITIAL	ug/Kg	1230000	J	CR	715	715	23600	Y	Y	1	DRY
E350.1	Ammonia Nitrogen	7664-41-7	N	INITIAL	ug/Kg		U		8490	8490	11800	N	Y	1	DRY
SM2540G	Total Solids	10-31-1	N	INITIAL	%	84.7						Y	Y	1	NA
SM4500-NORG-D	Kjeldahl Nitrogen, TKN	7727-37-9TKN	N	INITIAL	ug/Kg	1220000			179000	179000	236000	Y	Y	10	DRY
SW6010	Aluminum	7429-90-5	T	INITIAL	ug/Kg	1840000			7180	7180	23600	Y	Y	1	DRY
	Antimony	7440-36-0	T	INITIAL	ug/Kg		U		816	816	2360	N	Y	1	DRY
	Beryllium	7440-41-7	T	INITIAL	ug/Kg	470			56.3	56.3	236	Y	Y	1	DRY
	Calcium	7440-70-2	T	INITIAL	ug/Kg	5420000			22400	22400	118000	Y	Y	1	DRY
	Cobalt	7440-48-4	T	INITIAL	ug/Kg	2650			209	209	1180	Y	Y	1	DRY
	Iron	7439-89-6	T	INITIAL	ug/Kg	2320000			2640	2640	11800	Y	Y	1	DRY
	Magnesium	7439-95-4	T	INITIAL	ug/Kg	1900000			23500	23500	118000	Y	Y	1	DRY
	Manganese	7439-96-5	T	INITIAL	ug/Kg	183000			204	204	1180	Y	Y	1	DRY
	Potassium	7440-09-7	T	INITIAL	ug/Kg	1100000			24700	24700	118000	Y	Y	1	DRY
	Sodium	7440-23-5	T	INITIAL	ug/Kg	158000			48600	48600	118000	Y	Y	1	DRY
	Thallium	7440-28-0	T	INITIAL	ug/Kg		U		612	612	2360	N	Y	1	DRY
	Vanadium	7440-62-2	T	INITIAL	ug/Kg	7860			452	452	2360	Y	Y	1	DRY
SW8260	1,1,1,2-Tetrachloroethane	630-20-6	N	INITIAL	ug/Kg		U		1.29	1.29	3.40	N	Y	1	DRY
	1,1,1-Trichloroethane	71-55-6	N	INITIAL	ug/Kg		U		1.26	1.26	3.40	N	Y	1	DRY
	1,1,2,2-Tetrachloroethane	79-34-5	N	INITIAL	ug/Kg		U		0.946	0.946	3.40	N	Y	1	DRY
	1,1,2-Trichloroethane	79-00-5	N	INITIAL	ug/Kg		U		0.813	0.813	3.40	N	Y	1	DRY
	1,1,2-Trichlorotrifluoroethane	76-13-1	N	INITIAL	ug/Kg		U		1.03	1.03	3.40	N	Y	1	DRY
	1,1-Dichloroethane	75-34-3	N	INITIAL	ug/Kg		U		0.668	0.668	3.40	N	Y	1	DRY
	1,1-Dichloroethene	75-35-4	N	INITIAL	ug/Kg		U		0.825	0.825	3.40	N	Y	1	DRY
	1,1-Dichloropropene	563-58-6	N	INITIAL	ug/Kg		U		1.10	1.10	3.40	N	Y	1	DRY
	1,2,3-Trichlorobenzene	87-61-6	N	INITIAL	ug/Kg		U		9.98	9.98	17.0	N	Y	1	DRY
	1,2,3-Trichloropropane	96-18-4	N	INITIAL	ug/Kg		U		2.21	2.21	17.0	N	Y	1	DRY
	1,2,3-Trimethylbenzene	526-73-8	N	INITIAL	ug/Kg		U		2.15	2.15	6.81	N	Y	1	DRY
	1,2,4-Trichlorobenzene	120-82-1	N	INITIAL	ug/Kg		U		5.99	5.99	17.0	N	Y	1	DRY
	1,2-Dibromo-3-Chloropropane	96-12-8	N	INITIAL	ug/Kg		U		5.31	5.31	34.0	N	Y	1	DRY
	1,2-Dibromoethane	106-93-4	N	INITIAL	ug/Kg		U		0.882	0.882	3.40	N	Y	1	DRY
	1,2-Dichlorobenzene	95-50-1	N	INITIAL	ug/Kg		U		0.579	0.579	6.81	N	Y	1	DRY
	1,2-Dichloroethane	107-06-2	N	INITIAL	ug/Kg		U		0.884	0.884	3.40	N	Y	1	DRY

Lab Sample ID	L1853224-03
Sys Sample Code	GACO0429T000S006
Sample Name	GACO0429T000S006
Sample Date	4/29/2025 11:06:00 AM
Sample Type	N
Matrix	SO
Parent Sample	
% Moisture	15.30

Analytic Method	Chemical Name	CAS Rn	Fraction	Test Type	Result Unit	Final Result	Final Qual	Reason code	Final MDL	Final RL	Final QL	Final Detect	Final Report	DF	Basis
SW8260	1,2-Dichloropropane	78-87-5	N	INITIAL	ug/Kg		U		1.93	1.93	6.81	N	Y	1	DRY
	1,3-Dichlorobenzene	541-73-1	N	INITIAL	ug/Kg		U		0.817	0.817	6.81	N	Y	1	DRY
	1,3-Dichloropropane	142-28-9	N	INITIAL	ug/Kg		U		0.682	0.682	6.81	N	Y	1	DRY
	1,4-Dichlorobenzene	106-46-7	N	INITIAL	ug/Kg		U		0.953	0.953	6.81	N	Y	1	DRY
	2,2-Dichloropropane	594-20-7	N	INITIAL	ug/Kg		U		1.88	1.88	3.40	N	Y	1	DRY
	2-Butanone (MEK)	78-93-3	N	INITIAL	ug/Kg		U		86.4	86.4	136	N	Y	1	DRY
	2-Chlorotoluene	95-49-8	N	INITIAL	ug/Kg		U		1.18	1.18	3.40	N	Y	1	DRY
	4-Chlorotoluene	106-43-4	N	INITIAL	ug/Kg		U		0.613	0.613	6.81	N	Y	1	DRY
	4-Methyl-2-pentanone (MIBK)	108-10-1	N	INITIAL	ug/Kg		U		3.10	3.10	34.0	N	Y	1	DRY
	Acetone	67-64-1	N	INITIAL	ug/Kg		U		49.7	49.7	68.1	N	Y	1	DRY
	Acrylonitrile	107-13-1	N	INITIAL	ug/Kg		U		4.91	4.91	17.0	N	Y	1	DRY
	Bromobenzene	108-86-1	N	INITIAL	ug/Kg		U		1.23	1.23	17.0	N	Y	1	DRY
	Bromodichloromethane	75-27-4	N	INITIAL	ug/Kg		U		0.987	0.987	3.40	N	Y	1	DRY
	Bromoform	75-25-2	N	INITIAL	ug/Kg		U		1.59	1.59	34.0	N	Y	1	DRY
	Bromomethane	74-83-9	N	INITIAL	ug/Kg		U		2.68	2.68	17.0	N	Y	1	DRY
	Carbon tetrachloride	56-23-5	N	INITIAL	ug/Kg		U		1.22	1.22	6.81	N	Y	1	DRY
	Chlorobenzene	108-90-7	N	INITIAL	ug/Kg		U		0.286	0.286	3.40	N	Y	1	DRY
	Chlorodibromomethane	124-48-1	N	INITIAL	ug/Kg		U		0.833	0.833	3.40	N	Y	1	DRY
	Chloroethane	75-00-3	N	INITIAL	ug/Kg		U		2.31	2.31	6.81	N	Y	1	DRY
	Chloroform	67-66-3	N	INITIAL	ug/Kg		U		1.40	1.40	3.40	N	Y	1	DRY
	Chloromethane	74-87-3	N	INITIAL	ug/Kg		U		5.92	5.92	17.0	N	Y	1	DRY
	cis-1,2-Dichloroethene	156-59-2	N	INITIAL	ug/Kg		U		0.999	0.999	3.40	N	Y	1	DRY
	cis-1,3-Dichloropropene	10061-01-5	N	INITIAL	ug/Kg		U		1.03	1.03	3.40	N	Y	1	DRY
	Dibromomethane	74-95-3	N	INITIAL	ug/Kg		U		1.02	1.02	6.81	N	Y	1	DRY
	Dichlorodifluoromethane	75-71-8	N	INITIAL	ug/Kg		U		2.19	2.19	6.81	N	Y	1	DRY
	Di-isopropyl ether	108-20-3	N	INITIAL	ug/Kg		U		0.558	0.558	1.36	N	Y	1	DRY
	Hexachloro-1,3-butadiene	87-68-3	N	INITIAL	ug/Kg		U		8.17	8.17	34.0	N	Y	1	DRY
	Isopropylbenzene	98-82-8	N	INITIAL	ug/Kg		U		0.579	0.579	3.40	N	Y	1	DRY
	Methyl tert-butyl ether	1634-04-4	N	INITIAL	ug/Kg		U		0.476	0.476	1.36	N	Y	1	DRY
	Methylene Chloride	75-09-2	N	INITIAL	ug/Kg		U		9.04	9.04	34.0	N	Y	1	DRY
	n-Butylbenzene	104-51-8	N	INITIAL	ug/Kg		U		7.15	7.15	17.0	N	Y	1	DRY
	n-Propylbenzene	103-65-1	N	INITIAL	ug/Kg		U		1.29	1.29	6.81	N	Y	1	DRY
	p-Isopropyltoluene	99-87-6	N	INITIAL	ug/Kg		U		3.47	3.47	6.81	N	Y	1	DRY

Lab Sample ID	L1853224-03
Sys Sample Code	GACO0429T000S006
Sample Name	GACO0429T000S006
Sample Date	4/29/2025 11:06:00 AM
Sample Type	N
Matrix	SO
Parent Sample	
% Moisture	15.30

Analytic Method	Chemical Name	CAS Rn	Fraction	Test Type	Result Unit	Final Result	Final Qual	Reason code	Final MDL	Final RL	Final QL	Final Detect	Final Report	DF	Basis
SW8260	sec-Butylbenzene	135-98-8	N	INITIAL	ug/Kg		U		3.92	3.92	17.0	N	Y	1	DRY
	Styrene	100-42-5	N	INITIAL	ug/Kg		U		0.312	0.312	17.0	N	Y	1	DRY
	tert-Butylbenzene	98-06-6	N	INITIAL	ug/Kg		U		2.65	2.65	6.81	N	Y	1	DRY
	Tetrachloroethene	127-18-4	N	INITIAL	ug/Kg		U		1.22	1.22	3.40	N	Y	1	DRY
	trans-1,2-Dichloroethene	156-60-5	N	INITIAL	ug/Kg		U		1.42	1.42	6.81	N	Y	1	DRY
	trans-1,3-Dichloropropene	10061-02-6	N	INITIAL	ug/Kg		U		1.55	1.55	6.81	N	Y	1	DRY
	Trichloroethene	79-01-6	N	INITIAL	ug/Kg		U		0.795	0.795	1.36	N	Y	1	DRY
	Trichlorofluoromethane	75-69-4	N	INITIAL	ug/Kg		U		1.13	1.13	3.40	N	Y	1	DRY
	Vinyl chloride	75-01-4	N	INITIAL	ug/Kg		U		1.58	1.58	3.40	N	Y	1	DRY
SW8270	1,2,4-Trichlorobenzene	120-82-1	N	INITIAL	ug/Kg		U		12.3	12.3	393	N	Y	1	DRY
	1,2-Dichlorobenzene	95-50-1	N	INITIAL	ug/Kg		U		11.7	11.7	393	N	Y	1	DRY
	1,3-Dichlorobenzene	541-73-1	N	INITIAL	ug/Kg		U		11.9	11.9	393	N	Y	1	DRY
	1,4-Dichlorobenzene	106-46-7	N	INITIAL	ug/Kg		U		11.7	11.7	393	N	Y	1	DRY
	2,2-Oxybis(1-Chloropropane)	108-60-1	N	INITIAL	ug/Kg		U		17.0	17.0	393	N	Y	1	DRY
	2,4,6-Trichlorophenol	88-06-2	N	INITIAL	ug/Kg		U		12.6	12.6	393	N	Y	1	DRY
	2,4-Dichlorophenol	120-83-2	N	INITIAL	ug/Kg		U		11.5	11.5	393	N	Y	1	DRY
	2,4-Dimethylphenol	105-67-9	N	INITIAL	ug/Kg		U		10.3	10.3	393	N	Y	1	DRY
	2,4-Dinitrophenol	51-28-5	N	INITIAL	ug/Kg		U		92.0	92.0	393	N	Y	1	DRY
	2,4-Dinitrotoluene	121-14-2	N	INITIAL	ug/Kg		U		11.3	11.3	393	N	Y	1	DRY
	2,6-Dinitrotoluene	606-20-2	N	INITIAL	ug/Kg		U		12.9	12.9	393	N	Y	1	DRY
	2-Chloronaphthalene	91-58-7	N	INITIAL	ug/Kg		U		6.91	6.91	39.3	N	Y	1	DRY
	2-Chlorophenol	95-57-8	N	INITIAL	ug/Kg		U		13.0	13.0	393	N	Y	1	DRY
	2-Nitrophenol	88-75-5	N	INITIAL	ug/Kg		U		14.0	14.0	393	N	Y	1	DRY
	3,3-Dichlorobenzidine	91-94-1	N	INITIAL	ug/Kg		U		14.5	14.5	393	N	Y	1	DRY
	4,6-Dinitro-2-methylphenol	534-52-1	N	INITIAL	ug/Kg		U		89.1	89.1	393	N	Y	1	DRY
	4-Bromophenyl-phenylether	101-55-3	N	INITIAL	ug/Kg		U		13.8	13.8	393	N	Y	1	DRY
	4-Chloro-3-methylphenol	59-50-7	N	INITIAL	ug/Kg		U		12.8	12.8	393	N	Y	1	DRY
	4-Chlorophenyl-phenylether	7005-72-3	N	INITIAL	ug/Kg		U		13.7	13.7	393	N	Y	1	DRY
	4-Nitrophenol	100-02-7	N	INITIAL	ug/Kg		U		12.3	12.3	393	N	Y	1	DRY
	Acenaphthylene	208-96-8	N	INITIAL	ug/Kg		U		5.54	5.54	39.3	N	Y	1	DRY
	Benzidine	92-87-5	N	INITIAL	ug/Kg		U		73.9	73.9	1970	N	Y	1	DRY
	Benzo(g,h,i)perylene	191-24-2	N	INITIAL	ug/Kg		U		7.19	7.19	39.3	N	Y	1	DRY
	Benzylbutyl phthalate	85-68-7	N	INITIAL	ug/Kg		U		12.3	12.3	393	N	Y	1	DRY

Lab Sample ID	L1853224-03
Sys Sample Code	GACO0429T000S006
Sample Name	GACO0429T000S006
Sample Date	4/29/2025 11:06:00 AM
Sample Type	N
Matrix	SO
Parent Sample	
% Moisture	15.30

Analytic Method	Chemical Name	CAS Rn	Fraction	Test Type	Result Unit	Final Result	Final Qual	Reason code	Final MDL	Final RL	Final QL	Final Detect	Final Report	DF	Basis
SW8270	Bis(2-chlorethoxy)methane	111-91-1	N	INITIAL	ug/Kg		U		11.8	11.8	393	N	Y	1	DRY
	Bis(2-chloroethyl)ether	111-44-4	N	INITIAL	ug/Kg		U		13.0	13.0	393	N	Y	1	DRY
	Bis(2-ethylhexyl)phthalate	117-81-7	N	INITIAL	ug/Kg		U		49.8	49.8	393	N	Y	1	DRY
	Diethyl phthalate	84-66-2	N	INITIAL	ug/Kg		U		13.0	13.0	393	N	Y	1	DRY
	Dimethyl phthalate	131-11-3	N	INITIAL	ug/Kg		U		83.3	83.3	393	N	Y	1	DRY
	Di-n-butyl phthalate	84-74-2	N	INITIAL	ug/Kg		U		13.5	13.5	393	N	Y	1	DRY
	Di-n-octyl phthalate	117-84-0	N	INITIAL	ug/Kg		U		26.6	26.6	393	N	Y	1	DRY
	Hexachloro-1,3-butadiene	87-68-3	N	INITIAL	ug/Kg		U		13.2	13.2	393	N	Y	1	DRY
	Hexachlorobenzene	118-74-1	N	INITIAL	ug/Kg		U		13.9	13.9	393	N	Y	1	DRY
	Hexachlorocyclopentadiene	77-47-4	N	INITIAL	ug/Kg		U		20.7	20.7	393	N	Y	1	DRY
	Hexachloroethane	67-72-1	N	INITIAL	ug/Kg		U		15.5	15.5	393	N	Y	1	DRY
	Isophorone	78-59-1	N	INITIAL	ug/Kg		U		12.0	12.0	393	N	Y	1	DRY
	Nitrobenzene	98-95-3	N	INITIAL	ug/Kg		U		13.7	13.7	393	N	Y	1	DRY
	n-Nitrosodimethylamine	62-75-9	N	INITIAL	ug/Kg		U		58.3	58.3	393	N	Y	1	DRY
	n-Nitrosodi-n-propylamine	621-64-7	N	INITIAL	ug/Kg		U		13.1	13.1	393	N	Y	1	DRY
	n-Nitrosodiphenylamine	86-30-6	N	INITIAL	ug/Kg		U		29.8	29.8	393	N	Y	1	DRY
	Pentachlorophenol	87-86-5	N	INITIAL	ug/Kg		U		10.6	10.6	393	N	Y	1	DRY
	Phenanthrene	85-01-8	N	INITIAL	ug/Kg		U		7.80	7.80	39.3	N	Y	1	DRY
	Phenol	108-95-2	N	INITIAL	ug/Kg		U		15.8	15.8	393	N	Y	1	DRY
SW9056	Nitrate-Nitrite	NO2-NO3	N	INITIAL	ug/Kg		U	BL	9180	9180	23600	N	Y	1	DRY
WBLACK	TOC By Walkley Black	10-35-5	N	INITIAL	ug/Kg	17700000			128000	128000	500000	Y	Y	5	NA

Lab Sample ID	L1853224-04
Sys Sample Code	GACO0429T000S007
Sample Name	GACO0429T000S007
Sample Date	4/29/2025 10:42:00 AM
Sample Type	N
Matrix	SO
Parent Sample	
% Moisture	25.80

Analytic Method	Chemical Name	CAS Rn	Fraction	Test Type	Result Unit	Final Result	Final Qual	Reason code	Final MDL	Final RL	Final QL	Final Detect	Final Report	DF	Basis
CALC	Total Nitrogen	TN	N	INITIAL	ug/Kg	1720000	J	CR	816	816	26900	Y	Y	1	DRY
E350.1	Ammonia Nitrogen	7664-41-7	N	INITIAL	ug/Kg		U		9680	9680	13500	N	Y	1	DRY
SM2540G	Total Solids	10-31-1	N	INITIAL	%	74.2						Y	Y	1	NA
SM4500-NORG-D	Kjeldahl Nitrogen, TKN	7727-37-9TKN	N	INITIAL	ug/Kg	1720000			205000	205000	269000	Y	Y	10	DRY
SW6010	Aluminum	7429-90-5	T	INITIAL	ug/Kg	2210000			8190	8190	26900	Y	Y	1	DRY
	Antimony	7440-36-0	T	INITIAL	ug/Kg		U		931	931	2690	N	Y	1	DRY
	Beryllium	7440-41-7	T	INITIAL	ug/Kg	451			64.2	64.2	269	Y	Y	1	DRY
	Calcium	7440-70-2	T	INITIAL	ug/Kg	4640000			25600	25600	135000	Y	Y	1	DRY
	Cobalt	7440-48-4	T	INITIAL	ug/Kg	3180			238	238	1350	Y	Y	1	DRY
	Iron	7439-89-6	T	INITIAL	ug/Kg	2620000			3020	3020	13500	Y	Y	1	DRY
	Magnesium	7439-95-4	T	INITIAL	ug/Kg	1350000			26800	26800	135000	Y	Y	1	DRY
	Manganese	7439-96-5	T	INITIAL	ug/Kg	227000			233	233	1350	Y	Y	1	DRY
	Potassium	7440-09-7	T	INITIAL	ug/Kg	1750000			28100	28100	135000	Y	Y	1	DRY
	Sodium	7440-23-5	T	INITIAL	ug/Kg	461000			55500	55500	135000	Y	Y	1	DRY
	Thallium	7440-28-0	T	INITIAL	ug/Kg		U		698	698	2690	N	Y	1	DRY
	Vanadium	7440-62-2	T	INITIAL	ug/Kg	8190			516	516	2690	Y	Y	1	DRY
SW8260	1,1,1,2-Tetrachloroethane	630-20-6	N	INITIAL	ug/Kg		U		1.61	1.61	4.24	N	Y	1	DRY
	1,1,1-Trichloroethane	71-55-6	N	INITIAL	ug/Kg		U		1.56	1.56	4.24	N	Y	1	DRY
	1,1,2,2-Tetrachloroethane	79-34-5	N	INITIAL	ug/Kg		U		1.18	1.18	4.24	N	Y	1	DRY
	1,1,2-Trichloroethane	79-00-5	N	INITIAL	ug/Kg		U		1.01	1.01	4.24	N	Y	1	DRY
	1,1,2-Trichlorotrifluoroethane	76-13-1	N	INITIAL	ug/Kg		U		1.28	1.28	4.24	N	Y	1	DRY
	1,1-Dichloroethane	75-34-3	N	INITIAL	ug/Kg		U		0.832	0.832	4.24	N	Y	1	DRY
	1,1-Dichloroethene	75-35-4	N	INITIAL	ug/Kg		U		1.03	1.03	4.24	N	Y	1	DRY
	1,1-Dichloropropene	563-58-6	N	INITIAL	ug/Kg		U		1.37	1.37	4.24	N	Y	1	DRY
	1,2,3-Trichlorobenzene	87-61-6	N	INITIAL	ug/Kg		U		12.4	12.4	21.2	N	Y	1	DRY
	1,2,3-Trichloropropane	96-18-4	N	INITIAL	ug/Kg		U		2.74	2.74	21.2	N	Y	1	DRY
	1,2,3-Trimethylbenzene	526-73-8	N	INITIAL	ug/Kg		U		2.68	2.68	8.47	N	Y	1	DRY
	1,2,4-Trichlorobenzene	120-82-1	N	INITIAL	ug/Kg		U		7.45	7.45	21.2	N	Y	1	DRY
	1,2-Dibromo-3-Chloropropane	96-12-8	N	INITIAL	ug/Kg		U		6.61	6.61	42.4	N	Y	1	DRY
	1,2-Dibromoethane	106-93-4	N	INITIAL	ug/Kg		U		1.10	1.10	4.24	N	Y	1	DRY
	1,2-Dichlorobenzene	95-50-1	N	INITIAL	ug/Kg		U		0.720	0.720	8.47	N	Y	1	DRY
	1,2-Dichloroethane	107-06-2	N	INITIAL	ug/Kg		U		1.10	1.10	4.24	N	Y	1	DRY

Lab Sample ID	L1853224-04
Sys Sample Code	GACO0429T000S007
Sample Name	GACO0429T000S007
Sample Date	4/29/2025 10:42:00 AM
Sample Type	N
Matrix	SO
Parent Sample	
% Moisture	25.80

Analytic Method	Chemical Name	CAS Rn	Fraction	Test Type	Result Unit	Final Result	Final Qual	Reason code	Final MDL	Final RL	Final QL	Final Detect	Final Report	DF	Basis
SW8260	1,2-Dichloropropane	78-87-5	N	INITIAL	ug/Kg		U		2.41	2.41	8.47	N	Y	1	DRY
	1,3-Dichlorobenzene	541-73-1	N	INITIAL	ug/Kg		U		1.02	1.02	8.47	N	Y	1	DRY
	1,3-Dichloropropane	142-28-9	N	INITIAL	ug/Kg		U		0.849	0.849	8.47	N	Y	1	DRY
	1,4-Dichlorobenzene	106-46-7	N	INITIAL	ug/Kg		U		1.19	1.19	8.47	N	Y	1	DRY
	2,2-Dichloropropane	594-20-7	N	INITIAL	ug/Kg		U		2.34	2.34	4.24	N	Y	1	DRY
	2-Butanone (MEK)	78-93-3	N	INITIAL	ug/Kg		U		108	108	169	N	Y	1	DRY
	2-Chlorotoluene	95-49-8	N	INITIAL	ug/Kg		U		1.47	1.47	4.24	N	Y	1	DRY
	4-Chlorotoluene	106-43-4	N	INITIAL	ug/Kg		U		0.762	0.762	8.47	N	Y	1	DRY
	4-Methyl-2-pentanone (MIBK)	108-10-1	N	INITIAL	ug/Kg		U		3.86	3.86	42.4	N	Y	1	DRY
	Acetone	67-64-1	N	INITIAL	ug/Kg		U		61.8	61.8	84.7	N	Y	1	DRY
	Acrylonitrile	107-13-1	N	INITIAL	ug/Kg		U		6.12	6.12	21.2	N	Y	1	DRY
	Bromobenzene	108-86-1	N	INITIAL	ug/Kg		U		1.52	1.52	21.2	N	Y	1	DRY
	Bromodichloromethane	75-27-4	N	INITIAL	ug/Kg		U		1.23	1.23	4.24	N	Y	1	DRY
	Bromoform	75-25-2	N	INITIAL	ug/Kg		U		1.98	1.98	42.4	N	Y	1	DRY
	Bromomethane	74-83-9	N	INITIAL	ug/Kg		U		3.34	3.34	21.2	N	Y	1	DRY
	Carbon tetrachloride	56-23-5	N	INITIAL	ug/Kg		U		1.52	1.52	8.47	N	Y	1	DRY
	Chlorobenzene	108-90-7	N	INITIAL	ug/Kg		U		0.356	0.356	4.24	N	Y	1	DRY
	Chlorodibromomethane	124-48-1	N	INITIAL	ug/Kg		U		1.04	1.04	4.24	N	Y	1	DRY
	Chloroethane	75-00-3	N	INITIAL	ug/Kg		U		2.88	2.88	8.47	N	Y	1	DRY
	Chloroform	67-66-3	N	INITIAL	ug/Kg		U		1.74	1.74	4.24	N	Y	1	DRY
	Chloromethane	74-87-3	N	INITIAL	ug/Kg		U		7.37	7.37	21.2	N	Y	1	DRY
	cis-1,2-Dichloroethene	156-59-2	N	INITIAL	ug/Kg		U		1.24	1.24	4.24	N	Y	1	DRY
	cis-1,3-Dichloropropene	10061-01-5	N	INITIAL	ug/Kg		U		1.28	1.28	4.24	N	Y	1	DRY
	Dibromomethane	74-95-3	N	INITIAL	ug/Kg		U		1.27	1.27	8.47	N	Y	1	DRY
	Dichlorodifluoromethane	75-71-8	N	INITIAL	ug/Kg		U		2.73	2.73	8.47	N	Y	1	DRY
	Di-isopropyl ether	108-20-3	N	INITIAL	ug/Kg		U		0.695	0.695	1.69	N	Y	1	DRY
	Hexachloro-1,3-butadiene	87-68-3	N	INITIAL	ug/Kg		U		10.2	10.2	42.4	N	Y	1	DRY
	Isopropylbenzene	98-82-8	N	INITIAL	ug/Kg		U		0.720	0.720	4.24	N	Y	1	DRY
	Methyl tert-butyl ether	1634-04-4	N	INITIAL	ug/Kg		U		0.593	0.593	1.69	N	Y	1	DRY
	Methylene Chloride	75-09-2	N	INITIAL	ug/Kg		U		11.2	11.2	42.4	N	Y	1	DRY
	n-Butylbenzene	104-51-8	N	INITIAL	ug/Kg		U		8.89	8.89	21.2	N	Y	1	DRY
	n-Propylbenzene	103-65-1	N	INITIAL	ug/Kg		U		1.61	1.61	8.47	N	Y	1	DRY
	p-Isopropyltoluene	99-87-6	N	INITIAL	ug/Kg		U		4.32	4.32	8.47	N	Y	1	DRY

Lab Sample ID	L1853224-04
Sys Sample Code	GACO0429T000S007
Sample Name	GACO0429T000S007
Sample Date	4/29/2025 10:42:00 AM
Sample Type	N
Matrix	SO
Parent Sample	
% Moisture	25.80

Analytic Method	Chemical Name	CAS Rn	Fraction	Test Type	Result Unit	Final Result	Final Qual	Reason code	Final MDL	Final RL	Final QL	Final Detect	Final Report	DF	Basis
SW8260	sec-Butylbenzene	135-98-8	N	INITIAL	ug/Kg		U		4.88	4.88	21.2	N	Y	1	DRY
	Styrene	100-42-5	N	INITIAL	ug/Kg		U		0.388	0.388	21.2	N	Y	1	DRY
	tert-Butylbenzene	98-06-6	N	INITIAL	ug/Kg		U		3.30	3.30	8.47	N	Y	1	DRY
	Tetrachloroethene	127-18-4	N	INITIAL	ug/Kg		U		1.52	1.52	4.24	N	Y	1	DRY
	trans-1,2-Dichloroethene	156-60-5	N	INITIAL	ug/Kg		U		1.76	1.76	8.47	N	Y	1	DRY
	trans-1,3-Dichloropropene	10061-02-6	N	INITIAL	ug/Kg		U		1.93	1.93	8.47	N	Y	1	DRY
	Trichloroethene	79-01-6	N	INITIAL	ug/Kg		U		0.989	0.989	1.69	N	Y	1	DRY
	Trichlorofluoromethane	75-69-4	N	INITIAL	ug/Kg		U		1.40	1.40	4.24	N	Y	1	DRY
	Vinyl chloride	75-01-4	N	INITIAL	ug/Kg		U		1.97	1.97	4.24	N	Y	1	DRY
SW8270	1,2,4-Trichlorobenzene	120-82-1	N	INITIAL	ug/Kg		U		14.0	14.0	448	N	Y	1	DRY
	1,2-Dichlorobenzene	95-50-1	N	INITIAL	ug/Kg		U		13.3	13.3	448	N	Y	1	DRY
	1,3-Dichlorobenzene	541-73-1	N	INITIAL	ug/Kg		U		13.6	13.6	448	N	Y	1	DRY
	1,4-Dichlorobenzene	106-46-7	N	INITIAL	ug/Kg		U		13.3	13.3	448	N	Y	1	DRY
	2,2-Oxybis(1-Chloropropane)	108-60-1	N	INITIAL	ug/Kg		U		19.4	19.4	448	N	Y	1	DRY
	2,4,6-Trichlorophenol	88-06-2	N	INITIAL	ug/Kg		U		14.4	14.4	448	N	Y	1	DRY
	2,4-Dichlorophenol	120-83-2	N	INITIAL	ug/Kg		U		13.1	13.1	448	N	Y	1	DRY
	2,4-Dimethylphenol	105-67-9	N	INITIAL	ug/Kg		U		11.7	11.7	448	N	Y	1	DRY
	2,4-Dinitrophenol	51-28-5	N	INITIAL	ug/Kg		U		105	105	448	N	Y	1	DRY
	2,4-Dinitrotoluene	121-14-2	N	INITIAL	ug/Kg		U		12.9	12.9	448	N	Y	1	DRY
	2,6-Dinitrotoluene	606-20-2	N	INITIAL	ug/Kg		U		14.7	14.7	448	N	Y	1	DRY
	2-Chloronaphthalene	91-58-7	N	INITIAL	ug/Kg		U		7.88	7.88	44.8	N	Y	1	DRY
	2-Chlorophenol	95-57-8	N	INITIAL	ug/Kg		U		14.8	14.8	448	N	Y	1	DRY
	2-Nitrophenol	88-75-5	N	INITIAL	ug/Kg		U		16.0	16.0	448	N	Y	1	DRY
	3,3-Dichlorobenzidine	91-94-1	N	INITIAL	ug/Kg		U		16.6	16.6	448	N	Y	1	DRY
	4,6-Dinitro-2-methylphenol	534-52-1	N	INITIAL	ug/Kg		U		102	102	448	N	Y	1	DRY
	4-Bromophenyl-phenylether	101-55-3	N	INITIAL	ug/Kg		U		15.8	15.8	448	N	Y	1	DRY
	4-Chloro-3-methylphenol	59-50-7	N	INITIAL	ug/Kg		U		14.5	14.5	448	N	Y	1	DRY
	4-Chlorophenyl-phenylether	7005-72-3	N	INITIAL	ug/Kg		U		15.6	15.6	448	N	Y	1	DRY
	4-Nitrophenol	100-02-7	N	INITIAL	ug/Kg		U		14.0	14.0	448	N	Y	1	DRY
	Acenaphthylene	208-96-8	N	INITIAL	ug/Kg		U		6.32	6.32	44.8	N	Y	1	DRY
	Benzidine	92-87-5	N	INITIAL	ug/Kg		U		84.3	84.3	2250	N	Y	1	DRY
	Benzo(g,h,i)perylene	191-24-2	N	INITIAL	ug/Kg	53.2			8.20	8.20	44.8	Y	Y	1	DRY
	Benzylbutyl phthalate	85-68-7	N	INITIAL	ug/Kg		U		14.0	14.0	448	N	Y	1	DRY

Lab Sample ID	L1853224-04
Sys Sample Code	GACO0429T000S007
Sample Name	GACO0429T000S007
Sample Date	4/29/2025 10:42:00 AM
Sample Type	N
Matrix	SO
Parent Sample	
% Moisture	25.80

Analytic Method	Chemical Name	CAS Rn	Fraction	Test Type	Result Unit	Final Result	Final Qual	Reason code	Final MDL	Final RL	Final QL	Final Detect	Final Report	DF	Basis
SW8270	Bis(2-chlorethoxy)methane	111-91-1	N	INITIAL	ug/Kg		U		13.5	13.5	448	N	Y	1	DRY
	Bis(2-chloroethyl)ether	111-44-4	N	INITIAL	ug/Kg		U		14.8	14.8	448	N	Y	1	DRY
	Bis(2-ethylhexyl)phthalate	117-81-7	N	INITIAL	ug/Kg		U		56.8	56.8	448	N	Y	1	DRY
	Diethyl phthalate	84-66-2	N	INITIAL	ug/Kg		U		14.8	14.8	448	N	Y	1	DRY
	Dimethyl phthalate	131-11-3	N	INITIAL	ug/Kg		U		95.1	95.1	448	N	Y	1	DRY
	Di-n-butyl phthalate	84-74-2	N	INITIAL	ug/Kg		U		15.4	15.4	448	N	Y	1	DRY
	Di-n-octyl phthalate	117-84-0	N	INITIAL	ug/Kg		U		30.3	30.3	448	N	Y	1	DRY
	Hexachloro-1,3-butadiene	87-68-3	N	INITIAL	ug/Kg		U		15.1	15.1	448	N	Y	1	DRY
	Hexachlorobenzene	118-74-1	N	INITIAL	ug/Kg		U		15.9	15.9	448	N	Y	1	DRY
	Hexachlorocyclopentadiene	77-47-4	N	INITIAL	ug/Kg		U		23.6	23.6	448	N	Y	1	DRY
	Hexachloroethane	67-72-1	N	INITIAL	ug/Kg		U		17.6	17.6	448	N	Y	1	DRY
	Isophorone	78-59-1	N	INITIAL	ug/Kg		U		13.7	13.7	448	N	Y	1	DRY
	Nitrobenzene	98-95-3	N	INITIAL	ug/Kg		U		15.6	15.6	448	N	Y	1	DRY
	n-Nitrosodimethylamine	62-75-9	N	INITIAL	ug/Kg		U		66.5	66.5	448	N	Y	1	DRY
	n-Nitrosodi-n-propylamine	621-64-7	N	INITIAL	ug/Kg		U		14.9	14.9	448	N	Y	1	DRY
	n-Nitrosodiphenylamine	86-30-6	N	INITIAL	ug/Kg		U		33.9	33.9	448	N	Y	1	DRY
	Pentachlorophenol	87-86-5	N	INITIAL	ug/Kg		U		12.1	12.1	448	N	Y	1	DRY
	Phenanthrene	85-01-8	N	INITIAL	ug/Kg	529			8.90	8.90	44.8	Y	Y	1	DRY
	Phenol	108-95-2	N	INITIAL	ug/Kg		U		18.0	18.0	448	N	Y	1	DRY
SW9056	Nitrate-Nitrite	NO2-NO3	N	INITIAL	ug/Kg		U	BL	2710	2710	26900	N	Y	1	DRY
WBLACK	TOC By Walkley Black	10-35-5	N	INITIAL	ug/Kg	26100000			128000	128000	500000	Y	Y	5	NA

Lab Sample ID	L1853224-05
Sys Sample Code	GACO0429T000T002
Sample Name	GACO0429T000T002
Sample Date	4/29/2025 7:30:00 AM
Sample Type	TB
Matrix	WQ
Parent Sample	
% Moisture	

Analytic Method	Chemical Name	CAS Rn	Fraction	Test Type	Result Unit	Final Result	Final Qual	Reason code	Final MDL	Final RL	Final QL	Final Detect	Final Report	DF	Basis
SW8260	1,1,1,2-Tetrachloroethane	630-20-6	N	INITIAL	ug/L		U		0.147	0.147	1.00	N	Y	1	NA
	1,1,1-Trichloroethane	71-55-6	N	INITIAL	ug/L		U		0.149	0.149	1.00	N	Y	1	NA
	1,1,2,2-Tetrachloroethane	79-34-5	N	INITIAL	ug/L		U		0.133	0.133	1.00	N	Y	1	NA
	1,1,2-Trichloroethane	79-00-5	N	INITIAL	ug/L		U		0.158	0.158	1.00	N	Y	1	NA
	1,1,2-Trichlorotrifluoroethane	76-13-1	N	INITIAL	ug/L		U		0.180	0.180	1.00	N	Y	1	NA
	1,1-Dichloroethane	75-34-3	N	INITIAL	ug/L		U		0.100	0.100	1.00	N	Y	1	NA
	1,1-Dichloroethene	75-35-4	N	INITIAL	ug/L		U		0.188	0.188	1.00	N	Y	1	NA
	1,1-Dichloropropene	563-58-6	N	INITIAL	ug/L		U		0.142	0.142	1.00	N	Y	1	NA
	1,2,3-Trichlorobenzene	87-61-6	N	INITIAL	ug/L		U		0.230	0.230	1.00	N	Y	1	NA
	1,2,3-Trichloropropane	96-18-4	N	INITIAL	ug/L		U		0.237	0.237	2.50	N	Y	1	NA
	1,2,3-Trimethylbenzene	526-73-8	N	INITIAL	ug/L		U		0.104	0.104	1.00	N	Y	1	NA
	1,2,4-Trichlorobenzene	120-82-1	N	INITIAL	ug/L		U		0.481	0.481	1.00	N	Y	1	NA
	1,2,4-Trimethylbenzene	95-63-6	N	INITIAL	ug/L		U		0.322	0.322	1.00	N	Y	1	NA
	1,2-Dibromo-3-Chloropropane	96-12-8	N	INITIAL	ug/L		U		0.276	0.276	5.00	N	Y	1	NA
	1,2-Dibromoethane	106-93-4	N	INITIAL	ug/L		U		0.126	0.126	1.00	N	Y	1	NA
	1,2-Dichlorobenzene	95-50-1	N	INITIAL	ug/L		U		0.107	0.107	1.00	N	Y	1	NA
	1,2-Dichloroethane	107-06-2	N	INITIAL	ug/L		U		0.0819	0.0819	1.00	N	Y	1	NA
	1,2-Dichloropropane	78-87-5	N	INITIAL	ug/L		U		0.149	0.149	1.00	N	Y	1	NA
	1,3,5-Trimethylbenzene	108-67-8	N	INITIAL	ug/L		U		0.104	0.104	1.00	N	Y	1	NA
	1,3-Dichlorobenzene	541-73-1	N	INITIAL	ug/L		U		0.110	0.110	1.00	N	Y	1	NA
	1,3-Dichloropropane	142-28-9	N	INITIAL	ug/L		U		0.110	0.110	1.00	N	Y	1	NA
	1,4-Dichlorobenzene	106-46-7	N	INITIAL	ug/L		U		0.120	0.120	1.00	N	Y	1	NA
	2,2-Dichloropropane	594-20-7	N	INITIAL	ug/L		U		0.161	0.161	1.00	N	Y	1	NA
	2-Butanone (MEK)	78-93-3	N	INITIAL	ug/L		U		1.19	1.19	10.0	N	Y	1	NA
	2-Chlorotoluene	95-49-8	N	INITIAL	ug/L		U		0.106	0.106	1.00	N	Y	1	NA
	4-Chlorotoluene	106-43-4	N	INITIAL	ug/L		U		0.114	0.114	1.00	N	Y	1	NA
	4-Methyl-2-pentanone (MIBK)	108-10-1	N	INITIAL	ug/L		U		0.478	0.478	10.0	N	Y	1	NA
	Acetone	67-64-1	N	INITIAL	ug/L		U		11.3	11.3	50.0	N	Y	1	NA
	Acrolein	107-02-8	N	INITIAL	ug/L		U		2.54	2.54	50.0	N	Y	1	NA
	Acrylonitrile	107-13-1	N	INITIAL	ug/L		U		0.671	0.671	10.0	N	Y	1	NA
	Benzene	71-43-2	N	INITIAL	ug/L		U		0.0941	0.0941	1.00	N	Y	1	NA
	Bromobenzene	108-86-1	N	INITIAL	ug/L		U		0.118	0.118	1.00	N	Y	1	NA
	Bromodichloromethane	75-27-4	N	INITIAL	ug/L		U		0.136	0.136	1.00	N	Y	1	NA

Lab Sample ID	L1853224-05
Sys Sample Code	GACO0429T000T002
Sample Name	GACO0429T000T002
Sample Date	4/29/2025 7:30:00 AM
Sample Type	TB
Matrix	WQ
Parent Sample	
% Moisture	

Analytic Method	Chemical Name	CAS Rn	Fraction	Test Type	Result Unit	Final Result	Final Qual	Reason code	Final MDL	Final RL	Final QL	Final Detect	Final Report	DF	Basis
SW8260	Bromoform	75-25-2	N	INITIAL	ug/L		U		0.129	0.129	1.00	N	Y	1	NA
	Bromomethane	74-83-9	N	INITIAL	ug/L		U		0.605	0.605	5.00	N	Y	1	NA
	Carbon tetrachloride	56-23-5	N	INITIAL	ug/L		U		0.128	0.128	1.00	N	Y	1	NA
	Chlorobenzene	108-90-7	N	INITIAL	ug/L		U		0.116	0.116	1.00	N	Y	1	NA
	Chlorodibromomethane	124-48-1	N	INITIAL	ug/L		U		0.140	0.140	1.00	N	Y	1	NA
	Chloroethane	75-00-3	N	INITIAL	ug/L		U		0.192	0.192	5.00	N	Y	1	NA
	Chloroform	67-66-3	N	INITIAL	ug/L		U		0.111	0.111	5.00	N	Y	1	NA
	Chloromethane	74-87-3	N	INITIAL	ug/L		U		0.960	0.960	2.50	N	Y	1	NA
	cis-1,2-Dichloroethene	156-59-2	N	INITIAL	ug/L		U		0.126	0.126	1.00	N	Y	1	NA
	cis-1,3-Dichloropropene	10061-01-5	N	INITIAL	ug/L		U		0.111	0.111	1.00	N	Y	1	NA
	Dibromomethane	74-95-3	N	INITIAL	ug/L		U		0.122	0.122	1.00	N	Y	1	NA
	Dichlorodifluoromethane	75-71-8	N	INITIAL	ug/L		U		0.374	0.374	5.00	N	Y	1	NA
	Di-isopropyl ether	108-20-3	N	INITIAL	ug/L		U		0.105	0.105	1.00	N	Y	1	NA
	Ethylbenzene	100-41-4	N	INITIAL	ug/L		U		0.137	0.137	1.00	N	Y	1	NA
	Hexachloro-1,3-butadiene	87-68-3	N	INITIAL	ug/L		U		0.337	0.337	1.00	N	Y	1	NA
	Isopropylbenzene	98-82-8	N	INITIAL	ug/L		U		0.105	0.105	1.00	N	Y	1	NA
	Methyl tert-butyl ether	1634-04-4	N	INITIAL	ug/L		U		0.101	0.101	1.00	N	Y	1	NA
	Methylene Chloride	75-09-2	N	INITIAL	ug/L		U		0.430	0.430	5.00	N	Y	1	NA
	Naphthalene	91-20-3	N	INITIAL	ug/L		U		1.00	1.00	5.00	N	Y	1	NA
	n-Butylbenzene	104-51-8	N	INITIAL	ug/L		U		0.157	0.157	1.00	N	Y	1	NA
	n-Propylbenzene	103-65-1	N	INITIAL	ug/L		U		0.0993	0.0993	1.00	N	Y	1	NA
	p-Isopropyltoluene	99-87-6	N	INITIAL	ug/L		U		0.120	0.120	1.00	N	Y	1	NA
	sec-Butylbenzene	135-98-8	N	INITIAL	ug/L		U		0.125	0.125	1.00	N	Y	1	NA
	Styrene	100-42-5	N	INITIAL	ug/L		U		0.118	0.118	1.00	N	Y	1	NA
	tert-Butylbenzene	98-06-6	N	INITIAL	ug/L		U		0.127	0.127	1.00	N	Y	1	NA
	Tetrachloroethene	127-18-4	N	INITIAL	ug/L		U		0.300	0.300	1.00	N	Y	1	NA
	Toluene	108-88-3	N	INITIAL	ug/L		U		0.278	0.278	1.00	N	Y	1	NA
	trans-1,2-Dichloroethene	156-60-5	N	INITIAL	ug/L		U		0.149	0.149	1.00	N	Y	1	NA
	trans-1,3-Dichloropropene	10061-02-6	N	INITIAL	ug/L		U		0.118	0.118	1.00	N	Y	1	NA
	Trichloroethene	79-01-6	N	INITIAL	ug/L		U		0.190	0.190	1.00	N	Y	1	NA
	Trichlorofluoromethane	75-69-4	N	INITIAL	ug/L		U		0.160	0.160	5.00	N	Y	1	NA
	Vinyl chloride	75-01-4	N	INITIAL	ug/L		U		0.234	0.234	1.00	N	Y	1	NA
	Xylenes, Total	1330-20-7	N	INITIAL	ug/L		U		0.174	0.174	3.00	N	Y	1	NA

Lab Sample ID	L1853224-06
Sys Sample Code	GACO0429T000T005
Sample Name	GACO0429T000T005
Sample Date	4/29/2025 7:00:00 AM
Sample Type	TB
Matrix	WQ
Parent Sample	
% Moisture	

Analytic Method	Chemical Name	CAS Rn	Fraction	Test Type	Result Unit	Final Result	Final Qual	Reason code	Final MDL	Final RL	Final QL	Final Detect	Final Report	DF	Basis
SW8260	1,1,1,2-Tetrachloroethane	630-20-6	N	INITIAL	ug/L		U		0.147	0.147	1.00	N	Y	1	NA
	1,1,1-Trichloroethane	71-55-6	N	INITIAL	ug/L		U		0.149	0.149	1.00	N	Y	1	NA
	1,1,2,2-Tetrachloroethane	79-34-5	N	INITIAL	ug/L		U		0.133	0.133	1.00	N	Y	1	NA
	1,1,2-Trichloroethane	79-00-5	N	INITIAL	ug/L		U		0.158	0.158	1.00	N	Y	1	NA
	1,1,2-Trichlorotrifluoroethane	76-13-1	N	INITIAL	ug/L		U		0.180	0.180	1.00	N	Y	1	NA
	1,1-Dichloroethane	75-34-3	N	INITIAL	ug/L		U		0.100	0.100	1.00	N	Y	1	NA
	1,1-Dichloroethene	75-35-4	N	INITIAL	ug/L		U		0.188	0.188	1.00	N	Y	1	NA
	1,1-Dichloropropene	563-58-6	N	INITIAL	ug/L		U		0.142	0.142	1.00	N	Y	1	NA
	1,2,3-Trichlorobenzene	87-61-6	N	INITIAL	ug/L		U		0.230	0.230	1.00	N	Y	1	NA
	1,2,3-Trichloropropane	96-18-4	N	INITIAL	ug/L		U		0.237	0.237	2.50	N	Y	1	NA
	1,2,3-Trimethylbenzene	526-73-8	N	INITIAL	ug/L		U		0.104	0.104	1.00	N	Y	1	NA
	1,2,4-Trichlorobenzene	120-82-1	N	INITIAL	ug/L		U		0.481	0.481	1.00	N	Y	1	NA
	1,2,4-Trimethylbenzene	95-63-6	N	INITIAL	ug/L		U		0.322	0.322	1.00	N	Y	1	NA
	1,2-Dibromo-3-Chloropropane	96-12-8	N	INITIAL	ug/L		U		0.276	0.276	5.00	N	Y	1	NA
	1,2-Dibromoethane	106-93-4	N	INITIAL	ug/L		U		0.126	0.126	1.00	N	Y	1	NA
	1,2-Dichlorobenzene	95-50-1	N	INITIAL	ug/L		U		0.107	0.107	1.00	N	Y	1	NA
	1,2-Dichloroethane	107-06-2	N	INITIAL	ug/L		U		0.0819	0.0819	1.00	N	Y	1	NA
	1,2-Dichloropropane	78-87-5	N	INITIAL	ug/L		U		0.149	0.149	1.00	N	Y	1	NA
	1,3,5-Trimethylbenzene	108-67-8	N	INITIAL	ug/L		U		0.104	0.104	1.00	N	Y	1	NA
	1,3-Dichlorobenzene	541-73-1	N	INITIAL	ug/L		U		0.110	0.110	1.00	N	Y	1	NA
	1,3-Dichloropropane	142-28-9	N	INITIAL	ug/L		U		0.110	0.110	1.00	N	Y	1	NA
	1,4-Dichlorobenzene	106-46-7	N	INITIAL	ug/L		U		0.120	0.120	1.00	N	Y	1	NA
	2,2-Dichloropropane	594-20-7	N	INITIAL	ug/L		U		0.161	0.161	1.00	N	Y	1	NA
	2-Butanone (MEK)	78-93-3	N	INITIAL	ug/L		U		1.19	1.19	10.0	N	Y	1	NA
	2-Chlorotoluene	95-49-8	N	INITIAL	ug/L		U		0.106	0.106	1.00	N	Y	1	NA
	4-Chlorotoluene	106-43-4	N	INITIAL	ug/L		U		0.114	0.114	1.00	N	Y	1	NA
	4-Methyl-2-pentanone (MIBK)	108-10-1	N	INITIAL	ug/L		U		0.478	0.478	10.0	N	Y	1	NA
	Acetone	67-64-1	N	INITIAL	ug/L		U		11.3	11.3	50.0	N	Y	1	NA
	Acrolein	107-02-8	N	INITIAL	ug/L		U		2.54	2.54	50.0	N	Y	1	NA
	Acrylonitrile	107-13-1	N	INITIAL	ug/L		U		0.671	0.671	10.0	N	Y	1	NA
	Benzene	71-43-2	N	INITIAL	ug/L		U		0.0941	0.0941	1.00	N	Y	1	NA
	Bromobenzene	108-86-1	N	INITIAL	ug/L		U		0.118	0.118	1.00	N	Y	1	NA
	Bromodichloromethane	75-27-4	N	INITIAL	ug/L		U		0.136	0.136	1.00	N	Y	1	NA

Lab Sample ID	L1853224-06
Sys Sample Code	GACO0429T000T005
Sample Name	GACO0429T000T005
Sample Date	4/29/2025 7:00:00 AM
Sample Type	TB
Matrix	WQ
Parent Sample	
% Moisture	

Analytic Method	Chemical Name	CAS Rn	Fraction	Test Type	Result Unit	Final Result	Final Qual	Reason code	Final MDL	Final RL	Final QL	Final Detect	Final Report	DF	Basis
SW8260	Bromoform	75-25-2	N	INITIAL	ug/L		U		0.129	0.129	1.00	N	Y	1	NA
	Bromomethane	74-83-9	N	INITIAL	ug/L		U		0.605	0.605	5.00	N	Y	1	NA
	Carbon tetrachloride	56-23-5	N	INITIAL	ug/L		U		0.128	0.128	1.00	N	Y	1	NA
	Chlorobenzene	108-90-7	N	INITIAL	ug/L		U		0.116	0.116	1.00	N	Y	1	NA
	Chlorodibromomethane	124-48-1	N	INITIAL	ug/L		U		0.140	0.140	1.00	N	Y	1	NA
	Chloroethane	75-00-3	N	INITIAL	ug/L		U		0.192	0.192	5.00	N	Y	1	NA
	Chloroform	67-66-3	N	INITIAL	ug/L		U		0.111	0.111	5.00	N	Y	1	NA
	Chloromethane	74-87-3	N	INITIAL	ug/L		U		0.960	0.960	2.50	N	Y	1	NA
	cis-1,2-Dichloroethene	156-59-2	N	INITIAL	ug/L		U		0.126	0.126	1.00	N	Y	1	NA
	cis-1,3-Dichloropropene	10061-01-5	N	INITIAL	ug/L		U		0.111	0.111	1.00	N	Y	1	NA
	Dibromomethane	74-95-3	N	INITIAL	ug/L		U		0.122	0.122	1.00	N	Y	1	NA
	Dichlorodifluoromethane	75-71-8	N	INITIAL	ug/L		U		0.374	0.374	5.00	N	Y	1	NA
	Di-isopropyl ether	108-20-3	N	INITIAL	ug/L		U		0.105	0.105	1.00	N	Y	1	NA
	Ethylbenzene	100-41-4	N	INITIAL	ug/L		U		0.137	0.137	1.00	N	Y	1	NA
	Hexachloro-1,3-butadiene	87-68-3	N	INITIAL	ug/L		U		0.337	0.337	1.00	N	Y	1	NA
	Isopropylbenzene	98-82-8	N	INITIAL	ug/L		U		0.105	0.105	1.00	N	Y	1	NA
	Methyl tert-butyl ether	1634-04-4	N	INITIAL	ug/L		U		0.101	0.101	1.00	N	Y	1	NA
	Methylene Chloride	75-09-2	N	INITIAL	ug/L		U		0.430	0.430	5.00	N	Y	1	NA
	Naphthalene	91-20-3	N	INITIAL	ug/L		U		1.00	1.00	5.00	N	Y	1	NA
	n-Butylbenzene	104-51-8	N	INITIAL	ug/L		U		0.157	0.157	1.00	N	Y	1	NA
	n-Propylbenzene	103-65-1	N	INITIAL	ug/L		U		0.0993	0.0993	1.00	N	Y	1	NA
	p-Isopropyltoluene	99-87-6	N	INITIAL	ug/L		U		0.120	0.120	1.00	N	Y	1	NA
	sec-Butylbenzene	135-98-8	N	INITIAL	ug/L		U		0.125	0.125	1.00	N	Y	1	NA
	Styrene	100-42-5	N	INITIAL	ug/L		U		0.118	0.118	1.00	N	Y	1	NA
	tert-Butylbenzene	98-06-6	N	INITIAL	ug/L		U		0.127	0.127	1.00	N	Y	1	NA
	Tetrachloroethene	127-18-4	N	INITIAL	ug/L		U		0.300	0.300	1.00	N	Y	1	NA
	Toluene	108-88-3	N	INITIAL	ug/L		U		0.278	0.278	1.00	N	Y	1	NA
	trans-1,2-Dichloroethene	156-60-5	N	INITIAL	ug/L		U		0.149	0.149	1.00	N	Y	1	NA
	trans-1,3-Dichloropropene	10061-02-6	N	INITIAL	ug/L		U		0.118	0.118	1.00	N	Y	1	NA
	Trichloroethene	79-01-6	N	INITIAL	ug/L		U		0.190	0.190	1.00	N	Y	1	NA
	Trichlorofluoromethane	75-69-4	N	INITIAL	ug/L		U		0.160	0.160	5.00	N	Y	1	NA
	Vinyl chloride	75-01-4	N	INITIAL	ug/L		U		0.234	0.234	1.00	N	Y	1	NA
	Xylenes, Total	1330-20-7	N	INITIAL	ug/L		U		0.174	0.174	3.00	N	Y	1	NA